

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032423\
 Data File : BN024553.D
 Acq On : 24 Mar 2023 19:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 00:27:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 25 00:22:43 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.435	0.482	-10.8	101	0.00
3	n-Nitrosodimethylamine	0.638	0.764	-19.7	112	0.00
4 S	2-Fluorophenol	0.886	1.026	-15.8	110	0.00
5 S	Phenol-d6	1.138	1.273	-11.9	109	0.00
6	bis(2-Chloroethyl)ether	1.129	1.272	-12.7	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.267	0.368	-37.8#	137	-0.01
9	Naphthalene	1.017	1.195	-17.5	112	0.00
10	Hexachlorobutadiene	0.193	0.237	-22.8	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.492	0.571	-16.1	112	0.00
12	2-Methylnaphthalene	0.687	0.779	-13.4	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.204	0.194	4.9	92	0.00
15 S	2-Fluorobiphenyl	1.511	1.827	-20.9	108	0.00
16	Acenaphthylene	1.897	2.042	-7.6	99	0.00
17	Acenaphthene	1.208	1.350	-11.8	101	0.00
18	Fluorene	1.443	1.691	-17.2	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.028	47.2#	59	0.00
21	4-Bromophenyl-phenylether	0.238	0.259	-8.8	101	0.00
22	Hexachlorobenzene	0.292	0.322	-10.3	101	0.00
23	Atrazine	0.151	0.155	-2.6	104	0.00
24	Pentachlorophenol	0.136	0.095	30.1#	73	0.00
25	Phenanthrene	1.185	1.310	-10.5	101	0.00
26	Anthracene	1.094	1.130	-3.3	98	0.00
27 SURR	Fluoranthene-d10	0.848	0.970	-14.4	106	0.00
28	Fluoranthene	1.297	1.416	-9.2	101	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
30	Pyrene	1.621	2.258	-39.3#	102	0.00
31 S	Terphenyl-d14	0.674	0.878	-30.3#	95	0.00
32	Benzo(a)anthracene	1.336	1.522	-13.9	88	0.00
33	Chrysene	1.425	1.748	-22.7	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.629	0.840	-33.5#	107	0.00
35 I	Perylene-d12	1.000	1.000	0.0	74	0.00
36	Indeno(1,2,3-cd)pyrene	1.498	1.599	-6.7	83	0.00
37	Benzo(b)fluoranthene	1.556	1.791	-15.1	90	0.00
38	Benzo(k)fluoranthene	1.605	1.706	-6.3	83	0.00
39 C	Benzo(a)pyrene	1.427	1.540	-7.9	85	0.00
40	Dibenzo(a,h)anthracene	1.161	1.193	-2.8	81	0.00
41	Benzo(g,h,i)perylene	1.354	1.549	-14.4	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4