

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
 Data File : BN019271.D
 Acq On : 01 Apr 2022 02:43
 Operator : CG/JU
 Sample : N2172-05
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-6R

Quant Time: Apr 01 04:47:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 12:41:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	3876	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	10076	0.400	ng	# 0.01
13) Acenaphthene-d10	14.474	164	6603	0.400	ng	0.00
19) Phenanthrene-d10	17.226	188	12501	0.400	ng	0.01
29) Chrysene-d12	21.405	240	9479	0.400	ng	# 0.00
35) Perylene-d12	23.767	264	7933	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	1430	0.169	ng	0.00
5) Phenol-d6	7.024	99	514	0.057	ng	0.03
8) Nitrobenzene-d5	9.025	82	1575	0.218	ng	0.03
11) 2-Methylnaphthalene-d10	12.242	152	5161	0.323	ng	0.01
14) 2,4,6-Tribromophenol	15.974	330	738	0.239	ng	0.01
15) 2-Fluorobiphenyl	13.105	172	7302	0.276	ng	0.00
27) Fluoranthene-d10	19.249	212	16790	0.447	ng	0.00
31) Terphenyl-d14	19.847	244	10498	0.498	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.315	149	1186	0.060	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
Data File : BN019271.D
Acq On : 01 Apr 2022 02:43
Operator : CG/JU
Sample : N2172-05
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-6R

Quant Time: Apr 01 04:47:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 31 12:41:03 2022
Response via : Initial Calibration

