

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
 Data File : BN019300.D
 Acq On : 01 Apr 2022 23:55
 Operator : CG/JU
 Sample : PB143805BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143805BL

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 04/04/2022
 Supervised By :Jagrut Upadhyay 04/04/2022

Quant Time: Apr 02 01:47:20 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 12:41:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	4917	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	12928	0.400	ng	# 0.01
13) Acenaphthene-d10	14.474	164	8302	0.400	ng	0.00
19) Phenanthrene-d10	17.226	188	14866	0.400	ng	0.01
29) Chrysene-d12	21.404	240	11199	0.400	ng	# 0.00
35) Perylene-d12	23.764	264	9200	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	3939	0.366	ng	0.00
5) Phenol-d6	7.009	99	2941m	0.258	ng	0.01
8) Nitrobenzene-d5	9.025	82	1937	0.209	ng	0.03
11) 2-Methylnaphthalene-d10	12.246	152	6345	0.309	ng	0.02
14) 2,4,6-Tribromophenol	15.985	330	588	0.151	ng	0.02
15) 2-Fluorobiphenyl	13.116	172	7865	0.236	ng	0.01
27) Fluoranthene-d10	19.253	212	15380	0.344	ng	0.00
31) Terphenyl-d14	19.843	244	8402	0.337	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
Data File : BN019300.D
Acq On : 01 Apr 2022 23:55
Operator : CG/JU
Sample : PB143805BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143805BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 04/04/2022
Supervised By :Jagrut Upadhyay 04/04/2022

Quant Time: Apr 02 01:47:20 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 31 12:41:03 2022
Response via : Initial Calibration

