

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
 Data File : BN019261.D
 Acq On : 31 Mar 2022 20:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 04:46:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 12:41:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.391	2.3	110	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	98	0.00
4 S	2-Fluorophenol	0.400	0.393	1.8	108	0.00
5 S	Phenol-d6	0.400	0.365	8.8	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.00
8 S	Nitrobenzene-d5	0.400	0.350	12.5	107	0.00
9	Naphthalene	0.400	0.400	0.0	113	0.00
10	Hexachlorobutadiene	0.400	0.411	-2.7	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	114	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	115	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.333	16.8	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	114	0.00
16	Acenaphthylene	0.400	0.369	7.8	116	0.00
17	Acenaphthene	0.400	0.390	2.5	115	0.00
18	Fluorene	0.400	0.394	1.5	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.325	18.8	101	0.01
21	4-Bromophenyl-phenylether	0.400	0.387	3.3	114	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	114	0.00
23	Atrazine	0.400	0.349	12.8	109	0.00
24	Pentachlorophenol	0.400	0.321	19.8	101	0.00
25	Phenanthrene	0.400	0.398	0.5	115	0.00
26	Anthracene	0.400	0.366	8.5	112	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	114	0.00
28	Fluoranthene	0.400	0.389	2.8	114	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	117	0.00
30	Pyrene	0.400	0.411	-2.7	114	0.00
31 S	Terphenyl-d14	0.400	0.408	-2.0	117	0.00
32	Benzo(a)anthracene	0.400	0.350	12.5	116	0.00
33	Chrysene	0.400	0.403	-0.8	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	110	0.00
35 I	Perylene-d12	0.400	0.400	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	119	0.00
37	Benzo(b)fluoranthene	0.400	0.399	0.3	125	0.00
38	Benzo(k)fluoranthene	0.400	0.363	9.3	115	0.00
39 C	Benzo(a)pyrene	0.400	0.404	-1.0	126	0.00
40	Dibenzo(a,h)anthracene	0.400	0.355	11.3	117	0.00
41	Benzo(g,h,i)perylene	0.400	0.367	8.3	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4