

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
 Data File : BN019278.D
 Acq On : 01 Apr 2022 06:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 07:53:45 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 12:41:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.371	7.3	97	0.00
3	n-Nitrosodimethylamine	0.400	0.340	15.0	88	0.00
4 S	2-Fluorophenol	0.400	0.403	-0.8	103	0.00
5 S	Phenol-d6	0.400	0.370	7.5	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.346	13.5	99	0.00
9	Naphthalene	0.400	0.399	0.3	106	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	108	0.00
12	2-Methylnaphthalene	0.400	0.402	-0.5	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.347	13.3	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	107	0.00
16	Acenaphthylene	0.400	0.371	7.3	108	0.00
17	Acenaphthene	0.400	0.395	1.3	108	0.00
18	Fluorene	0.400	0.393	1.8	106	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.336	16.0	98	0.01
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	105	0.00
22	Hexachlorobenzene	0.400	0.400	0.0	104	0.00
23	Atrazine	0.400	0.357	10.8	104	0.00
24	Pentachlorophenol	0.400	0.377	5.8	110	0.00
25	Phenanthrene	0.400	0.394	1.5	106	0.00
26	Anthracene	0.400	0.368	8.0	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.398	0.5	108	0.00
28	Fluoranthene	0.400	0.381	4.8	104	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	110	0.00
30	Pyrene	0.400	0.410	-2.5	107	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	110	0.00
32	Benzo(a)anthracene	0.400	0.344	14.0	107	0.00
33	Chrysene	0.400	0.383	4.3	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	110	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.321	19.8	100	0.00
37	Benzo(b)fluoranthene	0.400	0.362	9.5	107	0.00
38	Benzo(k)fluoranthene	0.400	0.378	5.5	112	0.00
39 C	Benzo(a)pyrene	0.400	0.393	1.8	115	0.00
40	Dibenzo(a,h)anthracene	0.400	0.319	20.3	99	0.00
41	Benzo(g,h,i)perylene	0.400	0.325	18.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4