

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN033122\
 Data File : BN019280.D
 Acq On : 01 Apr 2022 09:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 16:40:17 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN033122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 31 12:41:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.355	11.3	98	0.00
3	n-Nitrosodimethylamine	0.400	0.329	17.8	89	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	107	0.00
5 S	Phenol-d6	0.400	0.371	7.3	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.386	3.5	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.345	13.8	104	0.00
9	Naphthalene	0.400	0.399	0.3	111	-0.01
10	Hexachlorobutadiene	0.400	0.401	-0.3	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.412	-3.0	114	0.00
12	2-Methylnaphthalene	0.400	0.401	-0.3	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.400	0.334	16.5	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.389	2.8	112	0.00
16	Acenaphthylene	0.400	0.362	9.5	112	0.00
17	Acenaphthene	0.400	0.387	3.3	112	0.00
18	Fluorene	0.400	0.392	2.0	112	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	112	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.292	27.0#	85	0.01
21	4-Bromophenyl-phenylether	0.400	0.382	4.5	111	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	110	0.00
23	Atrazine	0.400	0.342	14.5	105	0.00
24	Pentachlorophenol	0.400	0.352	12.0	109	0.00
25	Phenanthrene	0.400	0.394	1.5	112	0.00
26	Anthracene	0.400	0.364	9.0	110	0.00
27 SURR	Fluoranthene-d10	0.400	0.391	2.3	112	0.00
28	Fluoranthene	0.400	0.387	3.3	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
30	Pyrene	0.400	0.405	-1.3	111	0.00
31 S	Terphenyl-d14	0.400	0.403	-0.8	115	0.00
32	Benzo(a)anthracene	0.400	0.350	12.5	115	0.00
33	Chrysene	0.400	0.392	2.0	111	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.346	13.5	106	0.00
35 I	Perylene-d12	0.400	0.400	0.0	114	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.339	15.3	109	0.00
37	Benzo(b)fluoranthene	0.400	0.371	7.3	113	0.00
38	Benzo(k)fluoranthene	0.400	0.369	7.8	113	0.00
39 C	Benzo(a)pyrene	0.400	0.385	3.8	117	0.00
40	Dibenzo(a,h)anthracene	0.400	0.325	18.8	105	0.00
41	Benzo(g,h,i)perylene	0.400	0.329	17.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4