

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040722\
 Data File : BN019382.D
 Acq On : 07 Apr 2022 18:29
 Operator : CG/JU
 Sample : PB143945BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB143945BL

Quant Time: Apr 08 06:17:30 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 04 17:40:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	4312	0.400	ng	0.00
7) Naphthalene-d8	10.637	136	14531	0.400	ng	# 0.01
13) Acenaphthene-d10	14.463	164	9621	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	19433	0.400	ng	# 0.01
29) Chrysene-d12	21.395	240	16576	0.400	ng	# 0.00
35) Perylene-d12	23.749	264	12677	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	4356	0.377	ng	0.00
5) Phenol-d6	7.002	99	3836	0.276	ng	0.01
8) Nitrobenzene-d5	9.003	82	2908	0.220	ng	0.02
11) 2-Methylnaphthalene-d10	12.231	152	7838	0.307	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	1200	0.205	ng	0.01
15) 2-Fluorobiphenyl	13.105	172	10016	0.250	ng	0.01
27) Fluoranthene-d10	19.240	212	20655	0.330	ng	0.00
31) Terphenyl-d14	19.834	244	11925	0.331	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040722\
Data File : BN019382.D
Acq On : 07 Apr 2022 18:29
Operator : CG/JU
Sample : PB143945BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB143945BL

Quant Time: Apr 08 06:17:30 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 04 17:40:00 2022
Response via : Initial Calibration

