


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040722\
 Data File : BN019373.D
 Acq On : 07 Apr 2022 11:47
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040422.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 04 17:40:00 2022






AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.3	167722	PASS
68	69	0.00	2	1.9	3214	PASS
69	198	0.00	100	39.3	167878	PASS
70	69	0.00	2	0.6	994	PASS
127	198	10	80	46.9	200149	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	427136	PASS
199	198	5	9	6.8	29197	PASS
275	198	10	60	26.1	111475	PASS
365	198	1	100	3.5	14755	PASS
441	198	0.01	100	10.5	44987	PASS
442	442	50	100	100.0	269909	PASS
443	442	15	24	21.9	59029	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040722\
 Data File : BN019373.D
 Acq On : 07 Apr 2022 11:47
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Apr 08 06:54:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 23 16:56:32 2022
 Response via : Initial Calibration



TIC: BN019373.D\data.ms

(70) Pentachlorophenol (C)

16.863min (-0.009) 21252.86 ng




response 447323

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.48
264.00	61.60	58.92
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040722\
Data File : BN019373.D
Acq On : 07 Apr 2022 11:47
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 08 06:54:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019373.D\data.ms

(77) Benzidine

19.451min (-0.015) 0.00 ng

response 2630916

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.05
183.00	13.20	12.97
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/7/2022	BNA_N	<u>BN019373.D</u>
Compound Name	Response	Retention Time
DDT	123118	20.692
DDD	241559	20.251
DDE	1619	19.745
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
243178	366296	66.39

Instrument :
BNA_N
ClientSampleId :
DFTPP