

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040821\
 Data File : BN014230.D
 Acq On : 07 Apr 2021 18:04
 Operator : CG/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

mohammad
 4/8/2021 1:37:18 PM

Quant Time: Apr 07 18:35:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 07 18:22:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.733	152	10317	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	41064	0.40	ng	# 0.00
13) Acenaphthene-d10	14.359	164	26382	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	59464	0.40	ng	0.00
29) Chrysene-d12	21.290	240	59648	0.40	ng	#-0.01
36) Perylene-d12	23.538	264	61188	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.333	112	22292	0.91	ng	0.00
5) Phenol-d6	6.895	99	30009	0.94	ng	0.00
8) Nitrobenzene-d5	8.870	82	22994	0.81	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	55936	0.84	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	10777	1.10	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	77410	0.81	ng	0.00
27) Fluoranthene-d10	19.136	212	140467	0.82	ng	0.00
31) Terphenyl-d14	19.742	244	111714	0.85	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.295	88	9231	0.75	ng	# 81
3) n-Nitrosodimethylamine	3.633	42	4920	0.49	ng	# 69
6) bis(2-Chloroethyl)ether	7.161	93	21110	0.76	ng	94
9) Naphthalene	10.556	128	83344	0.82	ng	99
10) Hexachlorobutadiene	10.847	225	16342	0.86	ng	# 99
12) 2-Methylnaphthalene	12.173	142	60025	0.84	ng	99
16) Acenaphthylene	14.067	152	95156	0.91	ng	100
17) Acenaphthene	14.422	154	60291	0.81	ng	99
18) Fluorene	15.412	166	78228	0.83	ng	100
20) 4,6-Dinitro-2-methylph...	15.475	198	4648	0.55	ng	# 79
21) 4-Bromophenyl-phenylether	16.300	248	26023	0.82	ng	96
22) Hexachlorobenzene	16.410	284	28454	0.84	ng	97
23) Atrazine	16.568	200	25683	1.10	ng	99
24) Pentachlorophenol	16.751	266	11222	1.29	ng	96
25) Phenanthrene	17.140	178	126749	0.78	ng	100
26) Anthracene	17.237	178	123028	0.89	ng	100
28) Fluoranthene	19.166	202	148396	0.80	ng	99
30) Pyrene	19.528	202	153749	0.85	ng	99
32) Benzo(a)anthracene	21.279	228	151495	0.89	ng	100
33) Chrysene	21.332	228	148092	0.80	ng	99
34) Bis(2-ethylhexyl)phtha...	21.226	149	86073	1.47	ng	99
35) Indeno(1,2,3-cd)pyrene	25.800	276	184445	0.92	ng	99
37) Benzo(b)fluoranthene	22.864	252	160898	0.75	ng	97
38) Benzo(k)fluoranthene	22.908	252	156242m	0.73	ng	
39) Benzo(a)pyrene	23.439	252	145419	0.76	ng	# 91
40) Dibenzo(a,h)anthracene	25.821	278	154523	0.83	ng	97
41) Benzo(g,h,i)perylene	26.488	276	156257	0.83	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040821\
Data File : BN014230.D
Acq On : 07 Apr 2021 18:04
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED

mohammad
4/8/2021 1:37:18 PM


Quant Time: Apr 07 18:35:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 07 18:22:05 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.733 min Scan# 615
Delta R.T. 0.001 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Tgt Ion:152 Resp: 10317
Ion Ratio Lower Upper
152 100
150 150.6 126.2 189.2
115 64.4 49.4 74.0

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#2
1,4-Dioxane
Concen: 0.75 ng
RT: 3.295 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Tgt Ion: 88 Resp: 9231
Ion Ratio Lower Upper
88 100
58 67.3 70.0 105.0#
43 29.3 28.3 42.5









#3
n-Nitrosodimethylamine
Concen: 0.49 ng
RT: 3.633 min Scan# 106
Delta R.T. -0.016 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion: 42 Resp: 4920
Ion Ratio Lower Upper
42 100
74 139.8 85.0 127.6#
44 4.4 9.4 14.0#

Instrument :
BNA_N
ClientSampled :
SSTDIC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM



#4
2-Fluorophenol
Concen: 0.91 ng
RT: 5.333 min Scan# 317
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion: 112 Resp: 22292
Ion Ratio Lower Upper
112 100
64 54.1 48.4 72.6
63 29.6 26.1 39.1









#5
Phenol-d6
Concen: 0.94 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:	99	Resp:	30009
Ion	Ratio	Lower	Upper
99	100		
42	18.6	17.9	26.9
71	34.0	27.0	40.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.76 ng
RT: 7.161 min Scan# 544
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04


Tgt Ion:	93	Resp:	21110
Ion	Ratio	Lower	Upper
93	100		
63	73.7	65.3	97.9
95	33.1	26.3	39.5





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:	136	Resp:	41064
Ion	Ratio	Lower	Upper
136	100		
137	12.3	8.9	13.3
54	10.1	8.4	12.6
68	11.0	5.8	8.8#

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#8
Nitrobenzene-d5
Concen: 0.81 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:	82	Resp:	22994
Ion	Ratio	Lower	Upper
82	100		
128	36.7	36.1	54.1
54	46.2	39.1	58.7





#9
Naphthalene
Concen: 0.82 ng
RT: 10.556 min Scan# 864
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8






Tgt Ion:128 Resp: 83344


Ion	Ratio	Lower	Upper
128	100		
129	11.8	8.9	13.3
127	13.3	10.6	16.0

Manual Integrations
APPROVED

mohammad
4/8/2021 1:37:18 PM






#10
Hexachlorobutadiene
Concen: 0.86 ng
RT: 10.847 min Scan# 887
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Tgt Ion:225 Resp: 16342

Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.9	50.6	75.8









#11
2-Methylnaphthalene-d10
Concen: 0.84 ng
RT: 12.098 min Scan# 1072
Delta R.T. -0.004 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:152 Resp: 55936
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM




#12
2-Methylnaphthalene
Concen: 0.84 ng
RT: 12.173 min Scan# 1089
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:142 Resp: 60025
Ion Ratio Lower Upper
142 100
141 88.0 71.4 107.2
115 36.3 29.1 43.7







#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:164 Resp: 26382
Ion Ratio Lower Upper
164 100
162 100.2 80.4 120.6
160 44.9 36.0 54.0




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#14
2,4,6-Tribromophenol
Concen: 1.10 ng
RT: 15.843 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Tgt Ion:330 Resp: 10777
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.8 30.6 45.8









#15
2-Fluorobiphenyl
Concen: 0.81 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:172 Resp: 77410
Ion Ratio Lower Upper
172 100
171 35.7 27.8 41.6
170 24.6 18.7 28.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM



#16
Acenaphthylene
Concen: 0.91 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04


Tgt Ion:152 Resp: 95156
Ion Ratio Lower Upper
152 100
151 19.7 15.5 23.3
153 13.2 10.5 15.7





#17
Acenaphthene
Concen: 0.81 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:154 Resp: 60291
Ion Ratio Lower Upper
154 100
153 110.1 88.8 133.2
152 52.9 43.8 65.6

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#18
Fluorene
Concen: 0.83 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Tgt Ion:166 Resp: 78228
Ion Ratio Lower Upper
166 100
165 97.9 78.6 118.0
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:	188	Resp:	59464
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	8.3	6.2	9.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.55 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04


Tgt Ion:	198	Resp:	4648
Ion Ratio	Lower	Upper	
198	100		
51	60.6	46.3	69.5
105	72.9	35.5	53.3





#21
4-Bromophenyl-phenylether
Concen: 0.82 ng
RT: 16.300 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:248 Resp: 26023
Ion Ratio Lower Upper
248 100
250 97.9 76.1 114.1
141 53.2 45.5 68.3

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#22
Hexachlorobenzene
Concen: 0.84 ng
RT: 16.410 min Scan# 1473
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:284 Resp: 28454
Ion Ratio Lower Upper
284 100
142 33.6 28.9 43.3
249 29.8 25.0 37.4





#23
Atrazine
Concen: 1.10 ng
RT: 16.568 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:200 Resp: 25683
Ion Ratio Lower Upper
200 100
173 26.6 21.9 32.9
215 48.2 38.1 57.1

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#24
Pentachlorophenol
Concen: 1.29 ng
RT: 16.751 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:266 Resp: 11222
Ion Ratio Lower Upper
266 100
264 63.9 48.8 73.2
268 64.7 48.8 73.2









#25
Phenanthrene
Concen: 0.78 ng
RT: 17.140 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04





Tgt Ion:178 Resp: 126749
Ion Ratio Lower Upper
178 100
176 19.1 15.0 22.6
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.89 ng
RT: 17.237 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:178 Resp: 123028
Ion Ratio Lower Upper
178 100
176 18.5 14.6 21.8
179 15.4 12.3 18.5




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad
4/8/2021 1:37:18 PM




#27
Fluoranthene-d10
Concen: 0.82 ng
RT: 19.136 min Scan# 1832
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04





Tgt Ion:212 Resp: 140467
Ion Ratio Lower Upper
212 100
106 12.0 10.1 15.1
104 6.8 5.6 8.4



#28
Fluoranthene
Concen: 0.80 ng
RT: 19.166 min Scan# 1838
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:202 Resp: 148396
Ion Ratio Lower Upper
202 100
101 10.4 8.6 13.0
203 17.3 13.8 20.6




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad
4/8/2021 1:37:18 PM




#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.290 min Scan# 2128
Delta R.T. -0.010 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04





Tgt Ion:240 Resp: 59648
Ion Ratio Lower Upper
240 100
120 16.4 10.8 16.2#
236 27.4 21.4 32.2



#30
Pyrene
Concen: 0.85 ng
RT: 19.528 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:202 Resp: 153749
Ion Ratio Lower Upper
202 100
200 20.8 16.3 24.5
203 17.6 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED

mohammad
4/8/2021 1:37:18 PM



#31
 Terphenyl-d14
 Concen: 0.85 ng
 RT: 19.742 min Scan# 1954
 Delta R.T. 0.000 min
 Lab File: BN014230.D
 Acq: 07 Apr 2021 18:04




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8



Tgt Ion:244 Resp: 111714
 Ion Ratio Lower Upper
 244 100
 212 7.6 5.9 8.9
 122 11.3 9.2 13.8




Manual Integrations
 APPROVED


mohammad
 4/8/2021 1:37:18 PM



#32
 Benzo(a)anthracene
 Concen: 0.89 ng
 RT: 21.279 min Scan# 2127
 Delta R.T. 0.000 min
 Lab File: BN014230.D
 Acq: 07 Apr 2021 18:04


Tgt Ion:228 Resp: 151495
 Ion Ratio Lower Upper
 228 100
 226 26.8 21.1 31.7
 229 19.8 15.8 23.8





#33
Chrysene
Concen: 0.80 ng
RT: 21.332 min Scan# 2132
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Tgt Ion:228 Resp: 148092
Ion Ratio Lower Upper
228 100
226 29.4 23.0 34.6
229 19.9 15.8 23.8

Manual Integrations
APPROVED



mohammad
4/8/2021 1:37:18 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 1.47 ng
RT: 21.226 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04



Tgt Ion:149 Resp: 86073
Ion Ratio Lower Upper
149 100
167 27.0 21.4 32.0
279 4.2 3.1 4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.92 ng
RT: 25.800 min Scan# 3225
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion: 276 Resp: 184445
Ion Ratio Lower Upper
276 100
138 23.7 18.6 27.8
277 25.0 19.8 29.8

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8




Manual Integrations
APPROVED


mohammad
4/8/2021 1:37:18 PM




#36
Perylene-d12
Concen: 0.40 ng
RT: 23.538 min Scan# 2564
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion: 264 Resp: 61188
Ion Ratio Lower Upper
264 100
260 24.5 19.7 29.5
265 51.9 19.2 28.8#









#37
Benzo(b)fluoranthene
Concen: 0.75 ng
RT: 22.864 min Scan# 2367
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04





Tgt Ion:252 Resp: 160898
Ion Ratio Lower Upper
252 100
253 24.1 18.5 27.7
125 11.1 10.3 15.5



#38
Benzo(k)fluoranthene
Concen: 0.73 ng m
RT: 22.908 min Scan# 2380
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:252 Resp: 156242
Ion Ratio Lower Upper
252 100
253 24.4 18.1 27.1
125 13.3 7.8 11.6#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Manual Integrations
APPROVED




mohammad
4/8/2021 1:37:18 PM




#39
Benzo(a)pyrene
Concen: 0.76 ng
RT: 23.439 min Scan# 2535
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04





Tgt Ion:252 Resp: 145419
Ion Ratio Lower Upper
252 100
253 25.1 19.4 29.2
125 12.6 16.6 24.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.83 ng
RT: 25.821 min Scan# 3231
Delta R.T. 0.000 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04




Tgt Ion:278 Resp: 154523
Ion Ratio Lower Upper
278 100
139 16.9 12.6 19.0
279 25.4 19.4 29.0



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Manual Integrations
APPROVED

mohammad
4/8/2021 1:37:18 PM



#41
Benzo(g,h,i)perylene
Concen: 0.83 ng
RT: 26.488 min Scan# 3426
Delta R.T. -0.003 min
Lab File: BN014230.D
Acq: 07 Apr 2021 18:04

Tgt Ion:	276	Resp:	156257
Ion	Ratio	Lower	Upper
276	100		
277	24.4	18.9	28.3
138	21.0	16.4	24.6



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

mohammad
4/8/2021 1:37:18 PM