

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
 Data File : BN014305.D
 Acq On : 16 Apr 2021 12:22
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 16 14:11:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 14:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	10683	0.40	ng	0.00
7) Naphthalene-d8	10.479	136	38801	0.40	ng	0.00
13) Acenaphthene-d10	14.333	164	20031	0.40	ng	0.00
19) Phenanthrene-d10	17.079	188	42565	0.40	ng	# 0.01
29) Chrysene-d12	21.268	240	43109	0.40	ng	# 0.00
36) Perylene-d12	23.500	264	40047	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	2396	0.09	ng	0.00
5) Phenol-d6	6.871	99	2957	0.09	ng	0.00
8) Nitrobenzene-d5	8.845	82	2250	0.09	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	5766	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	669	0.09	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	7637	0.10	ng	0.00
27) Fluoranthene-d10	19.111	212	10812	0.09	ng	0.00
31) Terphenyl-d14	19.717	244	9643	0.10	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.271	88	1641	0.12	ng	90
3) n-Nitrosodimethylamine	3.617	42	547	0.09	ng	# 1
6) bis(2-Chloroethyl)ether	7.128	93	2706	0.10	ng	100
9) Naphthalene	10.518	128	9496	0.10	ng	98
10) Hexachlorobutadiene	10.822	225	1955	0.10	ng	# 99
12) 2-Methylnaphthalene	12.142	142	6153	0.10	ng	99
16) Acenaphthylene	14.042	152	6957	0.09	ng	99
17) Acenaphthene	14.384	154	5408	0.10	ng	99
18) Fluorene	15.374	166	6596	0.09	ng	100
20) 4,6-Dinitro-2-methylph...	15.463	198	346	0.07	ng	# 25
21) 4-Bromophenyl-phenylether	16.276	248	2126	0.10	ng	98
22) Hexachlorobenzene	16.385	284	2746	0.10	ng	99
23) Atrazine	16.544	200	1217	0.07	ng	# 94
24) Pentachlorophenol	16.726	266	589	0.07	ng	99
25) Phenanthrene	17.116	178	11493	0.10	ng	99
26) Anthracene	17.201	178	8596	0.09	ng	99
28) Fluoranthene	19.141	202	12104	0.09	ng	100
30) Pyrene	19.503	202	12375	0.09	ng	100
32) Benzo(a)anthracene	21.258	228	12035	0.10	ng	98
33) Chrysene	21.300	228	16139	0.12	ng	98
34) Bis(2-ethylhexyl)phtha...	21.194	149	3763	0.09	ng	# 99
35) Indeno(1,2,3-cd)pyrene	25.742	276	20717	0.13	ng	98
37) Benzo(b)fluoranthene	22.829	252	21134	0.15	ng	# 93
38) Benzo(k)fluoranthene	22.874	252	20302	0.14	ng	# 94
39) Benzo(a)pyrene	23.401	252	12615	0.10	ng	# 79
40) Dibenzo(a,h)anthracene	25.759	278	16358	0.12	ng	97
41) Benzo(g,h,i)perylene	26.423	276	18111	0.13	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
Data File : BN014305.D
Acq On : 16 Apr 2021 12:22
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Apr 16 14:11:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 14:08:36 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.700 min Scan# 611
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	152	Resp:	10683
Ion	Ratio	Lower	Upper
152	100		
150	154.5	123.0	184.4
115	58.4	47.6	71.4



1,4-Dioxane
Concen: 0.12 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.008 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	88	Resp:	1641
Ion	Ratio	Lower	Upper
88	100		
58	51.4	47.3	70.9
43	20.7	20.6	31.0









n-Nitrosodimethylamine
 Concen: 0.09 ng
 RT: 3.617 min Scan# 104
 Delta R.T. 0.008 min
 Lab File: BN014305.D
 Acq: 16 Apr 2021 12:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion: 42 Resp: 547
 Ion Ratio Lower Upper
 42 100
 74 0.0 104.6 157.0#
 44 8.8 2.4 3.6#



2-Fluorophenol
 Concen: 0.09 ng
 RT: 5.300 min Scan# 313
 Delta R.T. -0.000 min
 Lab File: BN014305.D
 Acq: 16 Apr 2021 12:22

Tgt Ion: 112 Resp: 2396
 Ion Ratio Lower Upper
 112 100
 64 55.2 43.7 65.5
 63 32.2 23.8 35.8









Phenol-d6
 Concen: 0.09 ng
 RT: 6.871 min Scan# 508
 Delta R.T. -0.000 min
 Lab File: BN014305.D
 Acq: 16 Apr 2021 12:22




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	99	Resp:	2957
Ion	Ratio	Lower	Upper
99	100		
42	20.3	15.7	23.5
71	33.7	27.6	41.4




bis(2-Chloroethyl)ether
 Concen: 0.10 ng
 RT: 7.128 min Scan# 540
 Delta R.T. -0.000 min
 Lab File: BN014305.D
 Acq: 16 Apr 2021 12:22

Tgt Ion:	93	Resp:	2706
Ion	Ratio	Lower	Upper
93	100		
63	72.8	58.2	87.4
95	33.4	26.7	40.1





Abundance Scan 858 (10.480 min): BN014307.D\data.ms (-857-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.479 min Scan# 858
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	38801
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.6
54	4.2	3.5	5.3
68	4.0	3.4	5.2



Abundance Scan 858 (10.479 min): BN014305.D\data.ms (-842-)





Abundance Scan 729 (8.845 min): BN014307.D\data.ms (-725-)




Nitrobenzene-d5
Concen: 0.09 ng
RT: 8.845 min Scan# 729
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	82	Resp:	2250
Ion	Ratio	Lower	Upper
82	100		
128	43.1	35.3	52.9
54	44.3	33.4	50.2




Abundance Scan 729 (8.845 min): BN014305.D\data.ms (-713-)






Abundance Scan 861 (10.518 min): BN014307.D\data.ms (-856) (-)

Naphthalene
Concen: 0.10 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	128	Resp:	9496
Ion	Ratio	Lower	Upper
128	100		
129	12.3	9.0	13.4
127	14.2	10.7	16.1






Abundance Scan 885 (10.822 min): BN014307.D\data.ms (-886) (-)

Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.822 min Scan# 885
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:	225	Resp:	1955
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	50.9	76.3






Abundance Scan 1065 (12.066 min): BN014307.D\data.ms (-1032) (-)

2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.066 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:152 Resp: 5766
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4






Abundance Scan 1082 (12.142 min): BN014307.D\data.ms (-1070) (-)

2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.142 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:142 Resp: 6153
Ion Ratio Lower Upper
142 100
141 88.2 70.0 105.0
115 37.4 28.8 43.2






Abundance Scan 1308 (14.334 min): BN014307.D\data.ms (-1308) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.333 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	164	Resp:	20031
Ion Ratio	Lower	Upper	
164	100		
162	99.7	78.6	117.8
160	44.9	36.2	54.4






Abundance Scan 1425 (15.818 min): BN014307.D\data.ms (-1425) (-)

2,4,6-Tribromophenol
Concen: 0.09 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:	330	Resp:	669
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	35.7	28.8	43.2






Abundance Scan 1199 (12.951 min): BN014307.D\data.ms (-1191) (-)

2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.950 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	172	Resp:	7637
Ion	Ratio	Lower	Upper
172	100		
171	34.6	27.9	41.9
170	23.8	18.8	28.2






Abundance Scan 1285 (14.042 min): BN014307.D\data.ms (-1269) (-)

Acenaphthylene
Concen: 0.09 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:	152	Resp:	6957
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.9	23.9
153	12.8	10.6	15.8






Abundance Scan 1312 (14.384 min): BN014307.D\data.ms (-1304) (-)

Acenaphthene
Concen: 0.10 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	154	Resp:	5408
Ion Ratio	Lower	Upper	
154	100		
153	112.3	89.4	134.2
152	56.6	44.7	67.1






Abundance Scan 1390 (15.374 min): BN014307.D\data.ms (-1374) (-)

Fluorene
Concen: 0.09 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:	166	Resp:	6596
Ion Ratio	Lower	Upper	
166	100		
165	97.8	78.2	117.2
167	13.8	10.9	16.3






Abundance Scan 1527 (17.068 min): BN014307.D\data.ms (-1511) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.079 min Scan# 1528
Delta R.T. 0.011 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1







Tgt Ion:	188	Resp:	42565
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	5.9	10.2	15.4#



Abundance Scan 1396 (15.450 min): BN014307.D\data.ms (-1322) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.07 ng
RT: 15.463 min Scan# 1397
Delta R.T. 0.012 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22


Tgt Ion:	198	Resp:	346
Ion Ratio	Lower	Upper	
198	100		
51	227.2	82.6	123.8#
105	75.4	56.1	84.1






Abundance Scan 1462 (16.277 min): BN014307.D\data.ms (-1425) (-)

4-Bromophenyl-phenylether
Concen: 0.10 ng
RT: 16.276 min Scan# 1462
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	248	Resp:	2126
Ion	Ratio	Lower	Upper
248	100		
250	102.8	81.9	122.9
141	40.6	29.3	43.9






Abundance Scan 1471 (16.386 min): BN014307.D\data.ms (-1426) (-)

Hexachlorobenzene
Concen: 0.10 ng
RT: 16.385 min Scan# 1471
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Tgt Ion:	284	Resp:	2746
Ion	Ratio	Lower	Upper
284	100		
142	34.0	27.0	40.6
249	30.6	23.7	35.5





Abundance Scan 1484 (16.544 min): BN014307.D\data.ms (-1428) (-)

Atrazine
Concen: 0.07 ng
RT: 16.544 min Scan# 1484
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	200	Resp:	1217
Ion Ratio	Lower	Upper	
200	100		
173	28.1	18.5	27.7#
215	52.8	40.0	60.0





Abundance Scan 1484 (16.544 min): BN014305.D\data.ms (-1443) (-)




Abundance Scan 1499 (16.727 min): BN014307.D\data.ms (-1424) (-)


Pentachlorophenol
Concen: 0.07 ng
RT: 16.726 min Scan# 1499
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	266	Resp:	589
Ion Ratio	Lower	Upper	
266	100		
264	62.0	49.8	74.8
268	64.0	50.7	76.1



Abundance Scan 1499 (16.726 min): BN014305.D\data.ms (-1491) (-)









#25 (-)
Phenanthrene
Concen: 0.10 ng
RT: 17.116 min Scan# 1531
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	178	Resp:	11493
Ion Ratio	Lower	Upper	
178	100		
176	19.1	15.3	22.9
179	16.0	12.2	18.4




#26 (-)
Anthracene
Concen: 0.09 ng
RT: 17.201 min Scan# 1538
Delta R.T. -0.001 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	178	Resp:	8596
Ion Ratio	Lower	Upper	
178	100		
176	17.9	14.8	22.2
179	15.7	12.2	18.4





Abundance Scan 1825 (19.109 min): BN014307.D\data.ms (-1825) (-)




Fluoranthene-d10
Concen: 0.09 ng
RT: 19.111 min Scan# 1827
Delta R.T. 0.002 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	212	Resp:	10812
Ion Ratio	Lower	Upper	
212	100		
106	12.2	9.7	14.5
104	6.9	5.5	8.3



Abundance Scan 1827 (19.111 min): BN014305.D\data.ms (-1786) (-)





Abundance Scan 1831 (19.139 min): BN014307.D\data.ms (-1823) (-)




Fluoranthene
Concen: 0.09 ng
RT: 19.141 min Scan# 1833
Delta R.T. 0.002 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22


Tgt Ion:	202	Resp:	12104
Ion Ratio	Lower	Upper	
202	100		
101	10.1	8.3	12.5
203	17.1	13.8	20.6



Abundance Scan 1833 (19.141 min): BN014305.D\data.ms (-1792) (-)






Abundance Scan 2125 (21.271 min): BN014307.D\data.ms (-2125) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.268 min Scan# 2126
Delta R.T. -0.003 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:240 Resp: 43109
Ion Ratio Lower Upper
240 100
120 8.8 5.4 8.0#
236 26.3 20.4 30.6






Abundance Scan 1904 (19.501 min): BN014307.D\data.ms (-1904) (-)




Pyrene
Concen: 0.09 ng
RT: 19.503 min Scan# 1906
Delta R.T. 0.002 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:202 Resp: 12375
Ion Ratio Lower Upper
202 100
200 20.6 16.6 24.8
203 17.2 14.0 21.0





Abundance Scan 1947 (19.715 min): BN014307.D\data.ms (-1920) (-)




Terphenyl-d14
Concen: 0.10 ng
RT: 19.717 min Scan# 1949
Delta R.T. 0.002 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.4	6.2	9.4
122	11.5	9.0	13.6



Abundance Scan 1949 (19.717 min): BN014305.D\data.ms (-1918) (-)





Abundance Scan 2123 (21.250 min): BN014307.D\data.ms (-2120) (-)




Benzo(a)anthracene
Concen: 0.10 ng
RT: 21.258 min Scan# 2125
Delta R.T. 0.008 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22


Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.5	21.7	32.5
229	20.8	15.7	23.5



Abundance Scan 2125 (21.258 min): BN014305.D\data.ms (-2105) (-)






Abundance Scan 2128 (21.303 min): BN014307.D\data.ms (-2128) (-)




Chrysene
Concen: 0.12 ng
RT: 21.300 min Scan# 2129
Delta R.T. -0.003 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	228	Resp:	16139
Ion	Ratio	Lower	Upper
228	100		
226	31.1	23.8	35.6
229	19.5	15.6	23.4






Abundance Scan 2118 (21.197 min): BN014307.D\data.ms (-2118) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.09 ng
RT: 21.194 min Scan# 2119
Delta R.T. -0.003 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	149	Resp:	3763
Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.0	31.4
279	5.1	3.3	4.9#




Abundance Scan 3205 (25.738 min): BN014307.D\data.ms (-3185) (-)





Indeno(1,2,3-cd)pyrene
Concen: 0.13 ng
RT: 25.742 min Scan# 3208
Delta R.T. 0.004 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.5	18.3	27.5
277	25.0	19.9	29.9





Abundance Scan 2551 (23.500 min): BN014307.D\data.ms (-2536) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.500 min Scan# 2553
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	20.0	30.0
265	36.1	37.8	56.8#





Abundance Scan 2355 (22.829 min): BN014307.D\data.ms (-2355) (-)




Benzo(b)fluoranthene
Concen: 0.15 ng
RT: 22.829 min Scan# 2357
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	252	Resp:	21134
Ion	Ratio	Lower	Upper
252	100		
253	26.6	19.0	28.6
125	15.8	9.3	13.9#


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Abundance Scan 2357 (22.829 min): BN014305.D\data.ms (-2298) (-)





Abundance Scan 2368 (22.873 min): BN014307.D\data.ms (-2368) (-)




Benzo(k)fluoranthene
Concen: 0.14 ng
RT: 22.874 min Scan# 2370
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Tgt Ion:	252	Resp:	20302
Ion	Ratio	Lower	Upper
252	100		
253	26.9	19.1	28.7
125	14.0	9.1	13.7#




Abundance Scan 2370 (22.874 min): BN014305.D\data.ms (-2282) (-)






Abundance Scan 2522 (23.400 min): BN014307.D\data.ms (-2539) (-)

Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.401 min Scan# 2524
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
252	100		
253	32.0	20.2	30.2#
125	29.5	12.4	18.6#






Abundance Scan 3209 (25.752 min): BN014307.D\data.ms (-3140) (-)


Dibenzo(a,h)anthracene
Concen: 0.12 ng
RT: 25.759 min Scan# 3213
Delta R.T. 0.007 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22



Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.1	12.8	19.2
279	27.3	20.0	30.0




Abundance Scan 3405 (26.423 min): BN014307.D\data.ms (-3374) (-)



Benzo(g,h,i)perylene
Concen: 0.13 ng
RT: 26.423 min Scan# 3407
Delta R.T. 0.000 min
Lab File: BN014305.D
Acq: 16 Apr 2021 12:22

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Abundance Scan 3407 (26.423 min): BN014305.D\data.ms



Tgt Ion:	276	Resp:	18111
Ion	Ratio	Lower	Upper
276	100		
277	25.0	18.9	28.3
138	20.9	16.6	24.8

Abundance Scan 3407 (26.423 min): BN014305.D\data.ms (-3348) (-)

