

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
 Data File : BN014309.D
 Acq On : 16 Apr 2021 14:45
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Apr 16 15:15:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 14:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	8116	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	28630	0.40	ng	#-0.01
13) Acenaphthene-d10	14.333	164	14761	0.40	ng	0.00
19) Phenanthrene-d10	17.079	188	29968	0.40	ng	# 0.01
29) Chrysene-d12	21.268	240	33657	0.40	ng	# 0.00
36) Perylene-d12	23.500	264	27079	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	28327	1.46	ng	0.00
5) Phenol-d6	6.871	99	35927	1.49	ng	0.00
8) Nitrobenzene-d5	8.845	82	28569	1.57	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	71172	1.59	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	8579	1.49	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	92600	1.71	ng	0.00
27) Fluoranthene-d10	19.111	212	135218	1.56	ng	0.00
31) Terphenyl-d14	19.717	244	107859	1.45	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	15488	1.50	ng	91
3) n-Nitrosodimethylamine	3.585	42	9693	2.00	ng	# 96
6) bis(2-Chloroethyl)ether	7.128	93	33011	1.69	ng	99
9) Naphthalene	10.517	128	116110	1.63	ng	99
10) Hexachlorobutadiene	10.822	225	22682	1.59	ng	# 99
12) 2-Methylnaphthalene	12.142	142	76491	1.60	ng	100
16) Acenaphthylene	14.042	152	92113	1.62	ng	99
17) Acenaphthene	14.384	154	67469	1.63	ng	98
18) Fluorene	15.374	166	83906	1.59	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	5486	1.54	ng	# 55
21) 4-Bromophenyl-phenylether	16.276	248	25784	1.68	ng	98
22) Hexachlorobenzene	16.385	284	32706	1.67	ng	99
23) Atrazine	16.544	200	15860	1.38	ng	99
24) Pentachlorophenol	16.726	266	8849	1.48	ng	99
25) Phenanthrene	17.116	178	135674	1.64	ng	100
26) Anthracene	17.201	178	109784	1.60	ng	100
28) Fluoranthene	19.141	202	154255	1.62	ng	100
30) Pyrene	19.503	202	150026	1.42	ng	100
32) Benzo(a)anthracene	21.247	228	139265	1.45	ng	99
33) Chrysene	21.300	228	163368	1.57	ng	99
34) Bis(2-ethylhexyl)phtha...	21.194	149	36949	1.07	ng	100
35) Indeno(1,2,3-cd)pyrene	25.742	276	190614	1.56	ng	99
37) Benzo(b)fluoranthene	22.829	252	169393	1.76	ng	96
38) Benzo(k)fluoranthene	22.874	252	166732	1.73	ng	96
39) Benzo(a)pyrene	23.401	252	138146	1.70	ng	# 93
40) Dibenzo(a,h)anthracene	25.756	278	160048	1.71	ng	98
41) Benzo(g,h,i)perylene	26.427	276	162668	1.72	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
Data File : BN014309.D
Acq On : 16 Apr 2021 14:45
Operator : CG/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Apr 16 15:15:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 14:08:36 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.700 min Scan# 611
 Delta R.T. 0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:	152	Resp:	8116
Ion	Ratio	Lower	Upper
152	100		
150	154.0	123.0	184.4
115	58.3	47.6	71.4



1,4-Dioxane
 Concen: 1.50 ng
 RT: 3.263 min Scan# 60
 Delta R.T. -0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45

Tgt Ion:	88	Resp:	15488
Ion	Ratio	Lower	Upper
88	100		
58	68.0	47.3	70.9
43	27.6	20.6	31.0









n-Nitrosodimethylamine
Concen: 2.00 ng
RT: 3.585 min Scan# 100
Delta R.T. -0.024 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 9693
Ion Ratio Lower Upper
42 100
74 135.1 104.6 157.0
44 6.9 2.4 3.6#



2-Fluorophenol
Concen: 1.46 ng
RT: 5.300 min Scan# 313
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion: 112 Resp: 28327
Ion Ratio Lower Upper
112 100
64 54.9 43.7 65.5
63 30.0 23.8 35.8









Phenol-d6
 Concen: 1.49 ng
 RT: 6.871 min Scan# 508
 Delta R.T. -0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	99	Resp:	35927
Ion	Ratio	Lower	Upper
99	100		
42	18.6	15.7	23.5
71	34.4	27.6	41.4




bis(2-Chloroethyl)ether
 Concen: 1.69 ng
 RT: 7.128 min Scan# 540
 Delta R.T. -0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45

Tgt Ion:	93	Resp:	33011
Ion	Ratio	Lower	Upper
93	100		
63	73.2	58.2	87.4
95	32.9	26.7	40.1





Abundance Scan 858 (10.480 min): BN014307.D\data.ms (-857#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.013 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	28630
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	8.8	3.5	5.3#
68	7.4	3.4	5.2#



Abundance Scan 857 (10.467 min): BN014309.D\data.ms (-842) (-)





Abundance Scan 729 (8.845 min): BN014307.D\data.ms (-725#8)





Nitrobenzene-d5
Concen: 1.57 ng
RT: 8.845 min Scan# 729
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	82	Resp:	28569
Ion	Ratio	Lower	Upper
82	100		
128	43.3	35.3	52.9
54	39.7	33.4	50.2



Abundance Scan 729 (8.845 min): BN014309.D\data.ms (-713) (-)









Naphthalene
Concen: 1.63 ng
RT: 10.517 min Scan# 861
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	128	Resp:	116110
Ion Ratio	Lower	Upper	
128	100		
129	10.9	9.0	13.4
127	13.2	10.7	16.1



Hexachlorobutadiene
Concen: 1.59 ng
RT: 10.822 min Scan# 885
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Tgt Ion:	225	Resp:	22682
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.2	50.9	76.3





Abundance Scan 1065 (12.066 min): BN014307.D\data.ms (-1032) (-)

2-Methylnaphthalene-d10
Concen: 1.59 ng
RT: 12.066 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:152 Resp: 71172
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4






Abundance Scan 1065 (12.066 min): BN014309.D\data.ms (-1020) (-)




Abundance Scan 1082 (12.142 min): BN014307.D\data.ms (-1071) (-)

2-Methylnaphthalene
Concen: 1.60 ng
RT: 12.142 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:142 Resp: 76491
Ion Ratio Lower Upper
142 100
141 87.5 70.0 105.0
115 35.1 28.8 43.2




Abundance Scan 1082 (12.142 min): BN014309.D\data.ms (-1060) (-)





Abundance Scan 1308 (14.334 min): BN014307.D\data.ms (-1308) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.333 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	164	Resp:	14761
Ion Ratio	Lower	Upper	
164	100		
162	98.5	78.6	117.8
160	44.9	36.2	54.4




Abundance Scan 1308 (14.333 min): BN014309.D\data.ms (-1292) (-)





Abundance Scan 1425 (15.818 min): BN014307.D\data.ms (-1425) (-)


2,4,6-Tribromophenol
Concen: 1.49 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45




Tgt Ion:	330	Resp:	8579
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	34.3	28.8	43.2



Abundance Scan 1425 (15.818 min): BN014309.D\data.ms (-1417) (-)









Scan 1199 (12.951 min): BN014307.D\data.ms (-1191) (-)
 2-Fluorobiphenyl
 Concen: 1.71 ng
 RT: 12.950 min Scan# 1199
 Delta R.T. -0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	172	Resp:	92600
Ion Ratio	Lower	Upper	
172	100		
171	33.9	27.9	41.9
170	22.8	18.8	28.2



Scan 1285 (14.042 min): BN014307.D\data.ms (-1278) (-)
 Acenaphthylene
 Concen: 1.62 ng
 RT: 14.042 min Scan# 1285
 Delta R.T. -0.000 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45


Tgt Ion:	152	Resp:	92113
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.9	23.9
153	13.0	10.6	15.8





Abundance Scan 1312 (14.384 min): BN014307.D\data.ms (-1304) (-)

Acenaphthene
Concen: 1.63 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	154	Resp:	67469
Ion	Ratio	Lower	Upper
154	100		
153	109.8	89.4	134.2
152	54.7	44.7	67.1






Abundance Scan 1312 (14.384 min): BN014309.D\data.ms (-1304) (-)




Abundance Scan 1390 (15.374 min): BN014307.D\data.ms (-1304) (-)

Fluorene
Concen: 1.59 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	166	Resp:	83906
Ion	Ratio	Lower	Upper
166	100		
165	97.6	78.2	117.2
167	13.4	10.9	16.3




Abundance Scan 1390 (15.374 min): BN014309.D\data.ms (-1374) (-)





Abundance Scan 1527 (17.068 min): BN014307.D\data.ms (-1511) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.079 min Scan# 1528
Delta R.T. 0.011 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	188	Resp:	29968
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	6.3	10.2	15.4#




Abundance Scan 1528 (17.079 min): BN014309.D\data.ms (-1511) (-)





Abundance Scan 1396 (15.450 min): BN014307.D\data.ms (-1322) (-)


4,6-Dinitro-2-methylphenol
Concen: 1.54 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45



Tgt Ion:	198	Resp:	5486
Ion Ratio	Lower	Upper	
198	100		
51	46.4	82.6	123.8#
105	46.1	56.1	84.1#




Abundance Scan 1396 (15.450 min): BN014309.D\data.ms (-1388) (-)






Abundance Scan 1462 (16.277 min): BN014307.D\data.ms (-1425) (-)

4-Bromophenyl-phenylether
Concen: 1.68 ng
RT: 16.276 min Scan# 1462
Delta R.T. -0.001 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	248	Resp:	25784
Ion	Ratio	Lower	Upper
248	100		
250	101.2	81.9	122.9
141	38.5	29.3	43.9






Abundance Scan 1471 (16.386 min): BN014307.D\data.ms (-1463) (-)

Hexachlorobenzene
Concen: 1.67 ng
RT: 16.385 min Scan# 1471
Delta R.T. -0.001 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45




Tgt Ion:	284	Resp:	32706
Ion	Ratio	Lower	Upper
284	100		
142	34.0	27.0	40.6
249	30.0	23.7	35.5





Abundance Scan 1484 (16.544 min): BN014307.D\data.ms (-1428) (-)

Atrazine
Concen: 1.38 ng
RT: 16.544 min Scan# 1484
Delta R.T. -0.001 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:	200	Resp:	15860
Ion Ratio	Lower	Upper	
200	100		
173	22.3	18.5	27.7
215	49.0	40.0	60.0





Abundance Scan 1484 (16.544 min): BN014309.D\data.ms (-1443) (-)




Abundance Scan 1499 (16.727 min): BN014307.D\data.ms (-1424) (-)


Pentachlorophenol
Concen: 1.48 ng
RT: 16.726 min Scan# 1499
Delta R.T. -0.001 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	266	Resp:	8849
Ion Ratio	Lower	Upper	
266	100		
264	63.1	49.8	74.8
268	64.2	50.7	76.1



Abundance Scan 1499 (16.726 min): BN014309.D\data.ms (-1491) (-)









Phenanthrene
 Concen: 1.64 ng
 RT: 17.116 min Scan# 1531
 Delta R.T. -0.001 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Tgt Ion:	178	Resp:	135674
Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.2	12.2	18.4




Anthracene
 Concen: 1.60 ng
 RT: 17.201 min Scan# 1538
 Delta R.T. -0.001 min
 Lab File: BN014309.D
 Acq: 16 Apr 2021 14:45

Tgt Ion:	178	Resp:	109784
Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.8	22.2
179	15.3	12.2	18.4






Abundance Scan 1825 (19.109 min): BN014307.D\data.ms (-1825) (-)




Fluoranthene-d10
Concen: 1.56 ng
RT: 19.111 min Scan# 1827
Delta R.T. 0.002 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	212	Resp:	135218
Ion	Ratio	Lower	Upper
212	100		
106	11.9	9.7	14.5
104	6.7	5.5	8.3






Abundance Scan 1831 (19.139 min): BN014307.D\data.ms (-1823) (-)




Fluoranthene
Concen: 1.62 ng
RT: 19.141 min Scan# 1833
Delta R.T. 0.002 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	202	Resp:	154255
Ion	Ratio	Lower	Upper
202	100		
101	10.2	8.3	12.5
203	17.3	13.8	20.6






Abundance Scan 2125 (21.271 min): BN014307.D\data.ms (-2125) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.268 min Scan# 2126
Delta R.T. -0.003 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Tgt Ion:	240	Resp:	33657
Ion	Ratio	Lower	Upper
240	100		
120	8.1	5.4	8.0#
236	26.4	20.4	30.6






Abundance Scan 1904 (19.501 min): BN014307.D\data.ms (-1904) (-)




Pyrene
Concen: 1.42 ng
RT: 19.503 min Scan# 1906
Delta R.T. 0.002 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	202	Resp:	150026
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.6	24.8
203	17.5	14.0	21.0





Abundance Scan 1947 (19.715 min): BN014307.D\data.ms (-1920) (-)




Terphenyl-d14
Concen: 1.45 ng
RT: 19.717 min Scan# 1949
Delta R.T. 0.002 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.2	9.4
122	10.5	9.0	13.6



Abundance Scan 1949 (19.717 min): BN014309.D\data.ms (-1918) (-)





Abundance Scan 2123 (21.250 min): BN014307.D\data.ms (-2120) (-)




Benzo(a)anthracene
Concen: 1.45 ng
RT: 21.247 min Scan# 2124
Delta R.T. -0.003 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45


Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.8	21.7	32.5
229	19.1	15.7	23.5



Abundance Scan 2124 (21.247 min): BN014309.D\data.ms (-2105) (-)





Abundance Scan 2128 (21.303 min): BN014307.D\data.ms (-2128) (-)




Chrysene
Concen: 1.57 ng
RT: 21.300 min Scan# 2129
Delta R.T. -0.003 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	228	Resp:	163368
Ion	Ratio	Lower	Upper
228	100		
226	30.5	23.8	35.6
229	19.2	15.6	23.4



Abundance Scan 2129 (21.300 min): BN014309.D\data.ms (-2120) (-)





Abundance Scan 2118 (21.197 min): BN014307.D\data.ms (-2118) (-)




Bis(2-ethylhexyl)phthalate
Concen: 1.07 ng
RT: 21.194 min Scan# 2119
Delta R.T. -0.003 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	149	Resp:	36949
Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.0	31.4
279	3.9	3.3	4.9




Abundance Scan 2119 (21.194 min): BN014309.D\data.ms (-2110) (-)






Abundance Scan 3205 (25.738 min): BN014307.D\data.ms (-3185) (-)

Indeno(1,2,3-cd)pyrene
Concen: 1.56 ng
RT: 25.742 min Scan# 3208
Delta R.T. 0.004 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	276	Resp:	190614
Ion Ratio	Lower	Upper	
276	100		
138	23.6	18.3	27.5
277	24.8	19.9	29.9






Abundance Scan 2551 (23.500 min): BN014307.D\data.ms (-2536) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.500 min Scan# 2553
Delta R.T. 0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45



Tgt Ion:	264	Resp:	27079
Ion Ratio	Lower	Upper	
264	100		
260	24.5	20.0	30.0
265	43.4	37.8	56.8





Abundance Scan 2355 (22.829 min): BN014307.D\data.ms (-2355) (-)




Benzo(b)fluoranthene
Concen: 1.76 ng
RT: 22.829 min Scan# 2357
Delta R.T. 0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	252	Resp:	169393
Ion	Ratio	Lower	Upper
252	100		
253	22.2	19.0	28.6
125	9.9	9.3	13.9



Abundance Scan 2357 (22.829 min): BN014309.D\data.ms (-2298) (-)





Abundance Scan 2368 (22.873 min): BN014307.D\data.ms (-2368) (-)




Benzo(k)fluoranthene
Concen: 1.73 ng
RT: 22.874 min Scan# 2370
Delta R.T. 0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Tgt Ion:	252	Resp:	166732
Ion	Ratio	Lower	Upper
252	100		
253	22.1	19.1	28.7
125	9.5	9.1	13.7




Abundance Scan 2370 (22.874 min): BN014309.D\data.ms (-2282) (-)






Abundance Scan 2522 (23.400 min): BN014307.D\data.ms (-2539) (-)

Benzo(a)pyrene
Concen: 1.70 ng
RT: 23.401 min Scan# 2524
Delta R.T. 0.000 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6




Tgt Ion:	252	Resp:	138146
Ion Ratio	Lower	Upper	
252	100		
253	22.7	20.2	30.2
125	11.5	12.4	18.6#






Abundance Scan 3209 (25.752 min): BN014307.D\data.ms (-3190) (-)


Dibenzo(a,h)anthracene
Concen: 1.71 ng
RT: 25.756 min Scan# 3212
Delta R.T. 0.004 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45



Tgt Ion:	278	Resp:	160048
Ion Ratio	Lower	Upper	
278	100		
139	15.4	12.8	19.2
279	23.7	20.0	30.0




Abundance Scan 3405 (26.423 min): BN014307.D\data.ms (-3374) (-)




Benzo(g,h,i)perylene
Concen: 1.72 ng
RT: 26.427 min Scan# 3408
Delta R.T. 0.004 min
Lab File: BN014309.D
Acq: 16 Apr 2021 14:45

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Abundance Scan 3408 (26.427 min): BN014309.D\data.ms



Tgt Ion:	276	Resp:	162668
Ion	Ratio	Lower	Upper
276	100		
277	23.7	18.9	28.3
138	20.2	16.6	24.8



Abundance Scan 3408 (26.427 min): BN014309.D\data.ms (-3348) (-)

