

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
 Data File : BN014312.D
 Acq On : 16 Apr 2021 16:36
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN041621

Quant Time: Apr 16 17:10:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 16:29:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.465	-16.3	98	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	86	0.00
4 S	2-Fluorophenol	0.400	0.459	-14.8	94	0.00
5 S	Phenol-d6	0.400	0.449	-12.2	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.453	-13.2	93	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	95	0.00
9	Naphthalene	0.400	0.453	-13.2	95	0.00
10	Hexachlorobutadiene	0.400	0.457	-14.2	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.451	-12.7	96	0.00
12	2-Methylnaphthalene	0.400	0.452	-13.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.01
14 S	2,4,6-Tribromophenol	0.400	0.394	1.5	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.446	-11.5	97	0.00
16	Acenaphthylene	0.400	0.414	-3.5	98	0.00
17	Acenaphthene	0.400	0.438	-9.5	97	0.00
18	Fluorene	0.400	0.440	-10.0	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.412	-3.0	98	0.00
21	4-Bromophenyl-phenylether	0.400	0.426	-6.5	93	0.00
22	Hexachlorobenzene	0.400	0.454	-13.5	98	0.00
23	Atrazine	0.400	0.403	-0.8	96	0.00
24	Pentachlorophenol	0.400	0.383	4.3	96	0.00
25	Phenanthrene	0.400	0.449	-12.2	99	0.00
26	Anthracene	0.400	0.430	-7.5	99	0.00
27 SURR	Fluoranthene-d10	0.400	0.441	-10.2	99	0.00
28	Fluoranthene	0.400	0.438	-9.5	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.468	-17.0	98	0.00
31 S	Terphenyl-d14	0.400	0.467	-16.8	98	0.00
32	Benzo(a)anthracene	0.400	0.431	-7.7	93	0.00
33	Chrysene	0.400	0.462	-15.5	98	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.449	-12.2	101	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.450	-12.5	94	0.00
36 I	Perylene-d12	0.400	0.400	0.0	93	0.00
37	Benzo(b)fluoranthene	0.400	0.458	-14.5	95	0.00
38	Benzo(k)fluoranthene	0.400	0.451	-12.7	94	0.00
39 C	Benzo(a)pyrene	0.400	0.442	-10.5	93	0.00
40	Dibenzo(a,h)anthracene	0.400	0.452	-13.0	94	0.00
41	Benzo(g,h,i)perylene	0.400	0.455	-13.7	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN041621