


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
 Data File : BN014303.D
 Acq On : 16 Apr 2021 10:34
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN041621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 19 10:42:58 2021






AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.8	87000	PASS
68	69	0.00	2	1.7	1735	PASS
69	198	0.00	100	37.9	103677	PASS
70	69	0.00	2	0.6	650	PASS
127	198	10	80	48.0	131475	PASS
197	198	0.00	2	0.5	1263	PASS
198	198	100	100	100.0	273664	PASS
199	198	5	9	6.9	18816	PASS
275	198	10	60	25.0	68440	PASS
365	198	1	100	3.1	8603	PASS
441	198	0.01	100	12.9	35360	PASS
442	442	50	100	100.0	223061	PASS
443	442	15	24	19.9	44355	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
Data File : BN014303.D
Acq On : 16 Apr 2021 10:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 19 01:18:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 09:55:03 2021
Response via : Initial Calibration



TIC: BN014303.D\data.ms

(70) Pentachlorophenol (C)

16.728min (-0.018) 26163.40 ng




response 344944

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.84
264.00	61.60	68.77
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041621\
Data File : BN014303.D
Acq On : 16 Apr 2021 10:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 19 01:18:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 09:55:03 2021
Response via : Initial Calibration



TIC: BN014303.D\data.ms

(77) Benzidine

19.333min (-0.018) 2683.88 ng

response 2096085

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.80
183.00	13.20	11.26
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_N
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
4/16/2021	BNA_N	<u>BN014303.D</u>
Compound Name	Response	Retention Time
DDT	1010602	20.574
DDD	24692	20.133
DDE	1127	19.627
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
25819	1036421	2.49