

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042621\
 Data File : BN014444.D
 Acq On : 26 Apr 2021 14:57
 Operator : CG/JU
 Sample : M2163-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT-059-RW8-IDWSO-202104-1

Quant Time: Apr 26 15:32:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 26 12:30:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	8811	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	33463	0.40	ng	0.00
13) Acenaphthene-d10	14.308	164	18157	0.40	ng	0.00
19) Phenanthrene-d10	17.055	188	40067	0.40	ng	# 0.00
29) Chrysene-d12	21.261	240	38142	0.40	ng	# 0.00
36) Perylene-d12	23.479	264	33249	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	5130	0.28	ng	0.00
5) Phenol-d6	6.855	99	7396	0.33	ng	0.00
8) Nitrobenzene-d5	8.819	82	5680	0.20	ng	0.00
11) 2-Methylnaphthalene-d10	12.049	152	16350	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	836	0.14	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	23981	0.33	ng	0.01
27) Fluoranthene-d10	19.099	212	32440	0.31	ng	0.00
31) Terphenyl-d14	19.700	244	29109	0.36	ng	0.00
Target Compounds						
25) Phenanthrene	17.104	178	9387	0.08	ng	Qvalue 99
28) Fluoranthene	19.129	202	17947	0.15	ng	99
30) Pyrene	19.491	202	15942	0.14	ng	97
32) Benzo(a)anthracene	21.239	228	7237	0.07	ng	93
33) Chrysene	21.292	228	9905	0.08	ng	94
34) Bis(2-ethylhexyl)phtha...	21.176	149	11474	0.32	ng	99
37) Benzo(b)fluoranthene	22.815	252	11753	0.10	ng	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042621\
Data File : BN014444.D
Acq On : 26 Apr 2021 14:57
Operator : CG/JU
Sample : M2163-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT-059-RW8-IDWSO-202104-1

Quant Time: Apr 26 15:32:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 26 12:30:05 2021
Response via : Initial Calibration

