

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042621\
 Data File : BN014438.D
 Acq On : 26 Apr 2021 11:25
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 26 12:30:28 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 26 12:30:05 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	78	0.00
2	1,4-Dioxane	0.400	0.438	-9.5	86	0.00
3	n-Nitrosodimethylamine	0.400	0.498	-24.5	84	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	73	0.00
5 S	Phenol-d6	0.400	0.416	-4.0	81	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	79	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.231	42.3#	46	0.00
9	Naphthalene	0.400	0.410	-2.5	81	0.00
10	Hexachlorobutadiene	0.400	0.399	0.3	77	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	73	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	72	0.00
15 S	2-Fluorobiphenyl	0.400	0.414	-3.5	77	0.00
16	Acenaphthylene	0.400	0.362	9.5	68	0.00
17	Acenaphthene	0.400	0.391	2.3	73	0.00
18	Fluorene	0.400	0.425	-6.2	80	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	79	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.250	37.5#	53	0.00
21	4-Bromophenyl-phenylether	0.400	0.355	11.3	72	0.00
22	Hexachlorobenzene	0.400	0.404	-1.0	82	0.00
23	Atrazine	0.400	0.227	43.3#	49	0.00
24	Pentachlorophenol	0.400	0.367	8.3	84	0.00
25	Phenanthrene	0.400	0.388	3.0	79	0.01
26	Anthracene	0.400	0.354	11.5	75	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	80	0.00
28	Fluoranthene	0.400	0.402	-0.5	83	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.01
30	Pyrene	0.400	0.365	8.8	80	0.00
31 S	Terphenyl-d14	0.400	0.344	14.0	73	0.00
32	Benzo(a)anthracene	0.400	0.382	4.5	85	0.00
33	Chrysene	0.400	0.371	7.3	78	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.891	-122.7#	214	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.338	15.5	77	0.01
36 I	Perylene-d12	0.400	0.400	0.0	87	0.00
37	Benzo(b)fluoranthene	0.400	0.374	6.5	83	0.00
38	Benzo(k)fluoranthene	0.400	0.389	2.8	88	0.00
39 C	Benzo(a)pyrene	0.400	0.381	4.8	92	0.00
40	Dibenzo(a,h)anthracene	0.400	0.320	20.0	73	0.00
41	Benzo(g,h,i)perylene	0.400	0.324	19.0	71	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4