

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050123\
 Data File : BN025210.D
 Acq On : 01 May 2023 16:16
 Operator : CG/JU
 Sample : PB152474BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK474

Quant Time: May 02 04:02:32 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 27 15:27:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.913	152	6418	0.400	ng/ul	0.00
4) Naphthalene-d8	10.721	136	21777	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.561	164	13382	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.309	188	27350	0.400	ng/ul	0.00
17) Chrysene-d12	21.506	240	16355	0.400	ng/ul	0.00
23) Perylene-d12	23.941	264	16556	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	43382	6.528	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.316	152	10973	0.392	ng/ul	0.00
18) Fluoranthene-d10	19.340	212	22531	0.522	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050123\
Data File : BN025210.D
Acq On : 01 May 2023 16:16
Operator : CG/JU
Sample : PB152474BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK474

Quant Time: May 02 04:02:32 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 27 15:27:43 2023
Response via : Initial Calibration

