

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050123\
 Data File : BN025212.D
 Acq On : 01 May 2023 17:36
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4287

Quant Time: May 02 04:02:47 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 27 15:27:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.917	152	7176	0.400	ng/ul	0.00
4) Naphthalene-d8	10.721	136	23968	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.561	164	13641	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.309	188	27690	0.400	ng/ul	0.00
17) Chrysene-d12	21.509	240	16356	0.400	ng/ul	0.00
23) Perylene-d12	23.958	264	14466	0.400	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	2827	0.380	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.316	152	12740	0.414	ng/ul	0.00
18) Fluoranthene-d10	19.340	212	21907	0.508	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.331	88	3095	0.387	ng/ul	92
5) Naphthalene	10.770	128	24275	0.395	ng/ul	100
7) 2-Methylnaphthalene	12.387	142	15310	0.407	ng/ul	100
8) 1-Methylnaphthalene	12.607	142	16180	0.404	ng/ul	100
10) Acenaphthylene	14.283	152	20365	0.401	ng/ul	100
11) Acenaphthene	14.621	153	17289	0.404	ng/ul	100
12) Fluorene	15.611	166	18976	0.396	ng/ul	99
14) Pentachlorophenol	16.971	266	1681	0.309	ng/ul	98
15) Phenanthrene	17.351	178	29723	0.380	ng/ul	100
16) Anthracene	17.444	178	24626	0.385	ng/ul	100
19) Fluoranthene	19.373	202	29517	0.499	ng/ul	100
20) Pyrene	19.735	202	29348	0.483	ng/ul	100
21) Benzo(a)anthracene	21.494	228	20320	0.412	ng/ul	99
22) Chrysene	21.546	228	21985	0.409	ng/ul	99
24) Benzo(b)fluoranthene	23.207	252	21485	0.366	ng/ul	98
25) Benzo(k)fluoranthene	23.253	252	20494	0.363	ng/ul	97
26) Benzo(a)pyrene	23.847	252	17885	0.368	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	26.497	276	22358	0.400	ng/ul#	100
28) Dibenzo(a,h)anthracene	26.524	278	17275	0.395	ng/ul	98
29) Benzo(g,h,i)perylene	27.269	276	18854	0.396	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050123\
Data File : BN025212.D
Acq On : 01 May 2023 17:36
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4287

Quant Time: May 02 04:02:47 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 27 15:27:43 2023
Response via : Initial Calibration

