

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
 Data File : BN019663.D
 Acq On : 04 May 2022 14:01
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: May 04 14:49:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 14:30:32 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	6237	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	16571	0.400	ng	# 0.01
13) Acenaphthene-d10	14.484	164	9277	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	17822	0.400	ng	0.00
29) Chrysene-d12	21.404	240	15261	0.400	ng	# 0.00
35) Perylene-d12	23.747	264	11397	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	44066	2.862	ng	0.00
5) Phenol-d6	7.024	99	55196	2.996	ng	0.00
8) Nitrobenzene-d5	9.003	82	42390	3.598	ng	0.00
11) 2-Methylnaphthalene-d10	12.238	152	86332	3.357	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	11859	3.995	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	123869	3.119	ng	0.01
27) Fluoranthene-d10	19.249	212	162877	3.210	ng	0.00
31) Terphenyl-d14	19.847	244	111904	3.335	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.362	88	21464	2.944	ng	99
3) n-Nitrosodimethylamine	3.673	42	23779	4.178	ng	# 85
6) bis(2-Chloroethyl)ether	7.284	93	54451	3.468	ng	99
9) Naphthalene	10.701	128	148596	3.136	ng	99
10) Hexachlorobutadiene	11.000	225	28267	3.152	ng	# 100
12) 2-Methylnaphthalene	12.314	142	99262	3.332	ng	99
16) Acenaphthylene	14.196	152	150526	3.599	ng	99
17) Acenaphthene	14.538	154	99627	3.330	ng	96
18) Fluorene	15.521	166	120944	3.303	ng	99
20) 4,6-Dinitro-2-methylph...	15.596	198	12284	5.377	ng	# 75
21) 4-Bromophenyl-phenylether	16.415	248	38118	3.493	ng	# 78
22) Hexachlorobenzene	16.538	284	41002	3.216	ng	98
23) Atrazine	16.682	200	25408	3.930	ng	# 93
24) Pentachlorophenol	16.867	266	17263	4.011	ng	95
25) Phenanthrene	17.256	178	189175	3.174	ng	100
26) Anthracene	17.349	178	171885	3.483	ng	99
28) Fluoranthene	19.278	202	203938	3.165	ng	99
30) Pyrene	19.641	202	201371	3.219	ng	100
32) Benzo(a)anthracene	21.387	228	179555	3.262	ng	99
33) Chrysene	21.440	228	185335	3.081	ng	98
34) Bis(2-ethylhexyl)phtha...	21.333	149	92808	3.686	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.156	276	169890	3.103	ng	99
37) Benzo(b)fluoranthene	23.036	252	155778	3.348	ng	# 88
38) Benzo(k)fluoranthene	23.083	252	163405	3.485	ng	# 86
39) Benzo(a)pyrene	23.644	252	122503	3.379	ng	# 83
40) Dibenzo(a,h)anthracene	26.170	278	138534	3.109	ng	94
41) Benzo(g,h,i)perylene	26.890	276	138005	2.924	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
Data File : BN019663.D
Acq On : 04 May 2022 14:01
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: May 04 14:49:56 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 14:30:32 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.869 min Scan# 6
Delta R.T. 0.007 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 6237
Ion Ratio Lower Upper
152 100
150 153.9 122.6 183.8
115 59.6 49.0 73.6



#2
1,4-Dioxane
Concen: 2.944 ng
RT: 3.362 min Scan# 39
Delta R.T. -0.007 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion: 88 Resp: 21464
Ion Ratio Lower Upper
88 100
43 33.9 25.2 37.8
58 84.5 67.6 101.4









#3
 n-Nitrosodimethylamine
 Concen: 4.178 ng
 RT: 3.673 min Scan# 81
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion: 42 Resp: 23779
 Ion Ratio Lower Upper
 42 100
 74 109.5 101.8 152.6
 44 6.5 8.4 12.6#



#4
 2-Fluorophenol
 Concen: 2.862 ng
 RT: 5.449 min Scan# 328
 Delta R.T. 0.007 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Tgt Ion: 112 Resp: 44066
 Ion Ratio Lower Upper
 112 100
 64 60.4 47.3 70.9
 63 31.8 24.6 36.8









#5
Phenol-d6
Concen: 2.996 ng
RT: 7.024 min Scan# 546
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 55196
Ion Ratio Lower Upper
99 100
42 22.3 15.8 23.8
71 33.6 26.5 39.7



#6
bis(2-Chloroethyl)ether
Concen: 3.468 ng
RT: 7.284 min Scan# 582
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion: 93 Resp: 54451
Ion Ratio Lower Upper
93 100
63 81.9 65.8 98.8
95 32.9 27.4 41.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.658 min Scan# 943
Delta R.T. 0.011 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	16571
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.1	13.7
54	7.3	7.8	11.6#
68	5.3	5.4	8.2#



#8
Nitrobenzene-d5
Concen: 3.598 ng
RT: 9.003 min Scan# 789
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:	82	Resp:	42390
Ion Ratio	Lower	Upper	
82	100		
128	37.3	32.4	48.6
54	61.8	44.4	66.6









#9
Naphthalene
Concen: 3.136 ng
RT: 10.701 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 148596
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.1 10.8 16.2



#10
Hexachlorobutadiene
Concen: 3.152 ng
RT: 11.000 min Scan# 976
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:225 Resp: 28267
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.8 76.2





#11

2-Methylnaphthalene-d10

Concen: 3.357 ng

RT: 12.238 min Scan# 1216

Delta R.T. 0.000 min

Lab File: BN019663.D

Acq: 04 May 2022 14:01

Instrument :

BNA_N

ClientSampleId :





SSTDICC3.2

Tgt Ion:152 Resp: 86332

Ion Ratio Lower Upper

152 100

151 20.5 16.2 24.4



#12

2-Methylnaphthalene

Concen: 3.332 ng

RT: 12.314 min Scan# 1236

Delta R.T. 0.000 min

Lab File: BN019663.D

Acq: 04 May 2022 14:01




Tgt Ion:142 Resp: 99262


Ion Ratio Lower Upper

142 100

141 87.9 71.0 106.6

115 35.8 29.6 44.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.484 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 9277
Ion Ratio Lower Upper
164 100
162 100.4 81.8 122.6
160 46.2 38.3 57.5



#14
2,4,6-Tribromophenol
Concen: 3.995 ng
RT: 15.964 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:330 Resp: 11859
Ion Ratio Lower Upper
330 100
332 97.3 76.2 114.4
141 50.2 39.7 59.5









#15
2-Fluorobiphenyl
Concen: 3.119 ng
RT: 13.116 min Scan# 1344
Delta R.T. 0.011 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 123869
Ion Ratio Lower Upper
172 100
171 34.0 28.0 42.0
170 22.2 19.0 28.4



#16
Acenaphthylene
Concen: 3.599 ng
RT: 14.196 min Scan# 1445
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:152 Resp: 150526
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.4
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 3.330 ng
RT: 14.538 min Scan# 1477
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 99627
Ion Ratio Lower Upper
154 100
153 106.7 89.3 133.9
152 52.9 44.1 66.1



#18
Fluorene
Concen: 3.303 ng
RT: 15.521 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:166 Resp: 120944
Ion Ratio Lower Upper
166 100
165 97.5 79.1 118.7
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.215 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 17822
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.7 8.6 12.8



#20
4,6-Dinitro-2-methylphenol
Concen: 5.377 ng
RT: 15.596 min Scan# 1576
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:198 Resp: 12284
Ion Ratio Lower Upper
198 100
182 9.3 16.9 25.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.493 ng
RT: 16.415 min Scan# 1655
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 38118
Ion Ratio Lower Upper
248 100
250 98.6 75.5 113.3
141 61.4 79.9 119.9#



#22
Hexachlorobenzene
Concen: 3.216 ng
RT: 16.538 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:284 Resp: 41002
Ion Ratio Lower Upper
284 100
142 42.9 33.2 49.8
249 32.5 26.4 39.6









#23
 Atrazine
 Concen: 3.930 ng
 RT: 16.682 min Scan# 1681
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:	200	Resp:	25408
Ion Ratio	Lower	Upper	
200	100		
173	26.4	28.5	42.7#
215	46.0	37.0	55.4



#24
 Pentachlorophenol
 Concen: 4.011 ng
 RT: 16.867 min Scan# 1699
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Tgt Ion:	266	Resp:	17263
Ion Ratio	Lower	Upper	
266	100		
264	62.7	48.0	72.0
268	63.3	54.2	81.2









#25
 Phenanthrene
 Concen: 3.174 ng
 RT: 17.256 min Scan# 1737
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:178 Resp: 189175




Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.2
179	15.2	12.1	18.1




#26
 Anthracene
 Concen: 3.483 ng
 RT: 17.349 min Scan# 1746
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Tgt Ion:178 Resp: 171885

Ion	Ratio	Lower	Upper
178	100		
176	18.4	14.3	21.5
179	15.1	12.2	18.4









#27
Fluoranthene-d10
Concen: 3.210 ng
RT: 19.249 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 162877
Ion Ratio Lower Upper
212 100
106 15.9 13.5 20.3
104 9.0 7.6 11.4



#28
Fluoranthene
Concen: 3.165 ng
RT: 19.278 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:202 Resp: 203938
Ion Ratio Lower Upper
202 100
101 12.5 10.6 15.8
203 17.2 13.6 20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.404 min Scan# 2441
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 15261
Ion Ratio Lower Upper
240 100
120 12.9 17.4 26.2#
236 27.7 23.0 34.4



#30
Pyrene
Concen: 3.219 ng
RT: 19.641 min Scan# 2198
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:202 Resp: 201371
Ion Ratio Lower Upper
202 100
200 20.8 16.7 25.1
203 17.6 14.2 21.4









#31
 Terphenyl-d14
 Concen: 3.335 ng
 RT: 19.847 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:244 Resp: 111904
 Ion Ratio Lower Upper
 244 100
 212 8.6 7.6 11.4
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 3.262 ng
 RT: 21.387 min Scan# 2439
 Delta R.T. 0.000 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Tgt Ion:228 Resp: 179555
 Ion Ratio Lower Upper
 228 100
 226 26.8 21.6 32.4
 229 19.4 16.5 24.7









#33
Chrysene
Concen: 3.081 ng
RT: 21.440 min Scan# 2445
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 185335
Ion Ratio Lower Upper
228 100
226 29.4 24.6 36.8
229 19.5 15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.686 ng
RT: 21.333 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:149 Resp: 92808
Ion Ratio Lower Upper
149 100
167 26.9 21.7 32.5
279 2.7 2.8 4.2#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.747 min Scan# 3011
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:264 Resp: 11397
Ion Ratio Lower Upper
264 100
260 26.1 21.0 31.6
265 48.6 67.9 101.9#



#36
Indeno(1,2,3-cd)pyrene
Concen: 3.103 ng
RT: 26.156 min Scan# 3835
Delta R.T. -0.003 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01

Tgt Ion:276 Resp: 169890
Ion Ratio Lower Upper
276 100
138 32.7 26.6 40.0
277 24.9 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 3.348 ng
RT: 23.036 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 155778
Ion Ratio Lower Upper
252 100
253 21.9 21.2 31.8
125 12.7 15.9 23.9#



#38
Benzo(k)fluoranthene
Concen: 3.485 ng
RT: 23.083 min Scan# 2784
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01


Tgt Ion:252 Resp: 163405
Ion Ratio Lower Upper
252 100
253 21.6 21.7 32.5#
125 12.7 17.0 25.4#








#39
Benzo(a)pyrene
Concen: 3.379 ng
RT: 23.644 min Scan# 2976
Delta R.T. 0.000 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01


Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2





Tgt Ion:252 Resp: 122503
Ion Ratio Lower Upper
252 100
253 22.0 23.1 34.7#
125 14.3 20.4 30.6#




#40
Dibenzo(a,h)anthracene
Concen: 3.109 ng
RT: 26.170 min Scan# 3840
Delta R.T. -0.006 min
Lab File: BN019663.D
Acq: 04 May 2022 14:01



Tgt Ion:278 Resp: 138534
Ion Ratio Lower Upper
278 100
139 21.8 19.4 29.2
279 23.5 21.8 32.8





#41
 Benzo(g,h,i)perylene
 Concen: 2.924 ng
 RT: 26.890 min Scan# 4086
 Delta R.T. -0.003 min
 Lab File: BN019663.D
 Acq: 04 May 2022 14:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:	276	Resp:	138005
Ion	Ratio	Lower	Upper
276	100		
277	23.6	20.2	30.2
138	28.0	23.5	35.3

