

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
 Data File : BN019671.D
 Acq On : 04 May 2022 18:47
 Operator : CG/JU
 Sample : N2640-04DL 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MWBR3D-B-480-500-042922-FDDL

Quant Time: May 05 04:57:45 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:20:22 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 05/05/2022
 Supervised By : Jagrut Upadhyay 05/05/2022

Supervised By : Jagrut
 Upadhyay
 05/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	5189	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	14816	0.400	ng	0.01
13) Acenaphthene-d10	14.484	164	8654	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	17659	0.400	ng	# 0.00
29) Chrysene-d12	21.404	240	13054	0.400	ng	# 0.00
35) Perylene-d12	23.749	264	9758	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	242	0.020	ng	0.00
5) Phenol-d6	7.031	99	174	0.012	ng	0.00
8) Nitrobenzene-d5	9.014	82	509m	0.042	ng	0.01
11) 2-Methylnaphthalene-d10	12.242	152	1110	0.046	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	98	0.031	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	1789	0.048	ng	0.01
27) Fluoranthene-d10	19.248	212	2201	0.045	ng	0.00
31) Terphenyl-d14	19.847	244	1488	0.049	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.362	88	10505	1.748	ng	96
34) Bis(2-ethylhexyl)phtha...	21.333	149	4669	0.191	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
Data File : BN019671.D
Acq On : 04 May 2022 18:47
Operator : CG/JU
Sample : N2640-04DL 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MWBR3D-B-480-500-042922-FDDL

Quant Time: May 05 04:57:45 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 15:20:22 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/05/2022
Supervised By : Jagrut Upadhyay 05/05/2022

Supervised By : Jagrut
Upadhyay
05/05/2022

