

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
 Data File : BN019677.D
 Acq On : 04 May 2022 23:32
 Operator : CG/JU
 Sample : PB144586BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB144586BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/05/2022
 Supervised By : Jagrut Upadhyay 05/05/2022

Quant Time: May 05 04:58:20 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:20:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	5975	0.400	ng	0.00
7) Naphthalene-d8	10.658	136	17200	0.400	ng	# 0.01
13) Acenaphthene-d10	14.484	164	8990	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	18592	0.400	ng	0.00
29) Chrysene-d12	21.404	240	11868	0.400	ng	# 0.00
35) Perylene-d12	23.746	264	9041	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.449	112	4882	0.350	ng	0.00
5) Phenol-d6	7.031	99	5501	0.322	ng	0.00
8) Nitrobenzene-d5	9.014	82	4392m	0.315	ng	0.01
11) 2-Methylnaphthalene-d10	12.242	152	9950	0.355	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	721	0.217	ng	0.00
15) 2-Fluorobiphenyl	13.116	172	13759	0.355	ng	0.01
27) Fluoranthene-d10	19.248	212	17304	0.336	ng	0.00
31) Terphenyl-d14	19.847	244	12331	0.444	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
Data File : BN019677.D
Acq On : 04 May 2022 23:32
Operator : CG/JU
Sample : PB144586BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB144586BL

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/05/2022
Supervised By : Jagrut Upadhyay 05/05/2022

Quant Time: May 05 04:58:20 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 15:20:22 2022
Response via : Initial Calibration

