

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
 Data File : BN019665.D
 Acq On : 04 May 2022 15:12
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN050422

Quant Time: May 04 18:33:47 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:20:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.400	0.0	101	0.00
3	n-Nitrosodimethylamine	0.400	0.387	3.3	98	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	99	0.00
5 S	Phenol-d6	0.400	0.375	6.3	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.01
8 S	Nitrobenzene-d5	0.400	0.374	6.5	101	0.01
9	Naphthalene	0.400	0.397	0.8	104	0.00
10	Hexachlorobutadiene	0.400	0.401	-0.3	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	105	0.00
12	2-Methylnaphthalene	0.400	0.397	0.8	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.333	16.8	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	107	0.01
16	Acenaphthylene	0.400	0.379	5.3	104	0.00
17	Acenaphthene	0.400	0.397	0.8	107	0.00
18	Fluorene	0.400	0.393	1.8	104	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.368	8.0	102	0.00
21	4-Bromophenyl-phenylether	0.400	0.383	4.3	103	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	105	0.00
23	Atrazine	0.400	0.343	14.2	100	0.00
24	Pentachlorophenol	0.400	0.347	13.3	96	0.01
25	Phenanthrene	0.400	0.396	1.0	105	0.00
26	Anthracene	0.400	0.376	6.0	104	0.00
27 SURR	Fluoranthene-d10	0.400	0.381	4.8	102	0.00
28	Fluoranthene	0.400	0.380	5.0	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
30	Pyrene	0.400	0.408	-2.0	100	0.00
31 S	Terphenyl-d14	0.400	0.410	-2.5	102	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	96	0.00
33	Chrysene	0.400	0.404	-1.0	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	96	0.00
35 I	Perylene-d12	0.400	0.400	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	92	0.00
37	Benzo(b)fluoranthene	0.400	0.399	0.3	98	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	93	0.00
39 C	Benzo(a)pyrene	0.400	0.386	3.5	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.385	3.8	93	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN050422