

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050422\
 Data File : BN019674.D
 Acq On : 04 May 2022 20:34
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 05 04:58:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN050422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 15:20:22 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.386	3.5	96	0.00
3	n-Nitrosodimethylamine	0.400	0.380	5.0	95	0.00
4 S	2-Fluorophenol	0.400	0.362	9.5	92	0.00
5 S	Phenol-d6	0.400	0.356	11.0	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.01
8 S	Nitrobenzene-d5	0.400	0.363	9.3	99	0.00
9	Naphthalene	0.400	0.399	0.3	107	0.00
10	Hexachlorobutadiene	0.400	0.398	0.5	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.397	0.8	108	0.00
12	2-Methylnaphthalene	0.400	0.403	-0.8	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	111	0.01
16	Acenaphthylene	0.400	0.376	6.0	108	0.00
17	Acenaphthene	0.400	0.396	1.0	112	0.00
18	Fluorene	0.400	0.398	0.5	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	108	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.366	8.5	104	0.00
21	4-Bromophenyl-phenylether	0.400	0.397	0.8	110	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	109	0.00
23	Atrazine	0.400	0.322	19.5	96	0.00
24	Pentachlorophenol	0.400	0.316	21.0	90	0.00
25	Phenanthrene	0.400	0.397	0.8	108	0.00
26	Anthracene	0.400	0.372	7.0	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	102	0.00
28	Fluoranthene	0.400	0.371	7.3	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.415	-3.7	100	0.00
31 S	Terphenyl-d14	0.400	0.417	-4.2	102	0.00
32	Benzo(a)anthracene	0.400	0.374	6.5	94	0.00
33	Chrysene	0.400	0.408	-2.0	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.348	13.0	93	0.00
35 I	Perylene-d12	0.400	0.400	0.0	91	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	89	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	93	0.00
38	Benzo(k)fluoranthene	0.400	0.407	-1.7	92	0.00
39 C	Benzo(a)pyrene	0.400	0.371	7.3	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.387	3.3	90	0.00
41	Benzo(g,h,i)perylene	0.400	0.376	6.0	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4