

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051722\
 Data File : BN019815.D
 Acq On : 17 May 2022 11:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 03:53:04 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 18 03:50:55 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.475	0.458	3.6	74	0.00
3	n-Nitrosodimethylamine	0.520	0.482	7.3	72	0.00
4 S	2-Fluorophenol	1.001	1.104	-10.3	85	0.00
5 S	Phenol-d6	1.254	1.418	-13.1	89	0.00
6	bis(2-Chloroethyl)ether	1.221	1.248	-2.2	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
8 S	Nitrobenzene-d5	0.302	0.295	2.3	83	0.00
9	Naphthalene	1.215	1.175	3.3	79	0.00
10	Hexachlorobutadiene	0.221	0.215	2.7	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.696	0.683	1.9	80	0.00
12	2-Methylnaphthalene	0.827	0.807	2.4	80	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.158	0.154	2.5	88	0.00
15 S	2-Fluorobiphenyl	1.746	1.682	3.7	79	0.00
16	Acenaphthylene	1.876	1.748	6.8	83	0.00
17	Acenaphthene	1.388	1.284	7.5	77	0.00
18	Fluorene	1.754	1.665	5.1	79	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
20	4,6-Dinitro-2-methylphenol	0.057	0.011	80.7#	15#	0.00
21	4-Bromophenyl-phenylether	0.244	0.236	3.3	78	0.00
22	Hexachlorobenzene	0.271	0.266	1.8	79	0.00
23	Atrazine	0.159	0.152	4.4	83	0.00
24	Pentachlorophenol	0.105	0.051	51.4#	46#	0.00
25	Phenanthrene	1.385	1.339	3.3	77	0.00
26	Anthracene	1.171	1.089	7.0	77	0.00
27 SURR	Fluoranthene-d10	1.137	1.171	-3.0	84	0.00
28	Fluoranthene	1.484	1.507	-1.5	82	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
30	Pyrene	1.697	1.619	4.6	85	0.00
31 S	Terphenyl-d14	0.943	0.890	5.6	83	0.00
32	Benzo(a)anthracene	1.458	1.426	2.2	89	0.00
33	Chrysene	1.635	1.609	1.6	87	0.00
34	Bis(2-ethylhexyl)phthalate	0.607	0.594	2.1	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	99	0.00
36	Indeno(1,2,3-cd)pyrene	1.873	1.788	4.5	102	0.00
37	Benzo(b)fluoranthene	1.746	1.607	8.0	95	0.00
38	Benzo(k)fluoranthene	1.853	1.728	6.7	97	0.00
39 C	Benzo(a)pyrene	1.374	1.294	5.8	100	0.00
40	Dibenzo(a,h)anthracene	1.510	1.449	4.0	103	0.00
41	Benzo(g,h,i)perylene	1.553	1.438	7.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4