

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052422\
 Data File : BN019953.D
 Acq On : 25 May 2022 16:08
 Operator : CG/JU
 Sample : BL145043BS
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK043

Quant Time: May 26 01:15:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN052322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 23 14:58:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.850	152	3992	0.400	ng/ul	0.00
4) Naphthalene-d8	10.638	136	13893	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.473	164	8496	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.211	188	17477	0.400	ng/ul	0.00
17) Chrysene-d12	21.400	240	10644	0.400	ng/ul	0.00
23) Perylene-d12	23.738	264	9253	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.315	96	22387	6.015	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.227	152	6815	0.301	ng/ul	0.00
18) Fluoranthene-d10	19.242	212	13641	0.358	ng/ul	0.00
Target Compounds						
					Qvalue	
20) Pyrene	19.604	202	1779	0.033	ng/ul#	1
26) Benzo(a)pyrene	23.589	252	2835	0.073	ng/ul#	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052422\
Data File : BN019953.D
Acq On : 25 May 2022 16:08
Operator : CG/JU
Sample : BL145043BS
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK043

Quant Time: May 26 01:15:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN052322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 23 14:58:28 2022
Response via : Initial Calibration









#20
Pyrene
Concen: 0.033 ng/ul
RT: 19.604 min Scan# 30
Delta R.T. -0.033 min
Lab File: BN019953.D
Acq: 25 May 2022 16:08

Instrument :
BNA_N
ClientSampleId :
SBLK043

Tgt Ion:202 Resp: 1779
Ion Ratio Lower Upper
202 100
101 170.7 11.4 17.0#
100 1284.3 9.0 13.4#



#26
Benzo(a)pyrene
Concen: 0.073 ng/ul
RT: 23.589 min Scan# 4902
Delta R.T. -0.050 min
Lab File: BN019953.D
Acq: 25 May 2022 16:08

Tgt Ion:252 Resp: 2835
Ion Ratio Lower Upper
252 100
253 39.3 23.4 35.0#
125 33.9 28.3 42.5

