

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052721\
 Data File : BN014787.D
 Acq On : 27 May 2021 12:02
 Operator : CG/JU
 Sample : M2411-06
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-RW8-GW-798-800

Quant Time: May 27 12:35:04 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	748	0.40	ng	0.00
7) Naphthalene-d8	10.505	136	2639	0.40	ng	# 0.00
13) Acenaphthene-d10	14.359	164	1879	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	4354	0.40	ng	0.00
29) Chrysene-d12	21.314	240	5576	0.40	ng	0.00
35) Perylene-d12	23.575	264	5436	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.325	112	492	0.25	ng	0.00
5) Phenol-d6	6.895	99	475	0.17	ng	0.00
8) Nitrobenzene-d5	8.870	82	741	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.098	152	1471	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	281	0.26	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	2407	0.35	ng	-0.01
27) Fluoranthene-d10	19.154	212	4838	0.28	ng	0.00
31) Terphenyl-d14	19.754	244	5572	0.42	ng	0.00
Target Compounds						Qvalue
25) Phenanthrene	17.153	178	302	0.02	ng	# 82
34) Bis(2-ethylhexyl)phtha...	21.229	149	2492	0.52	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052721\
Data File : BN014787.D
Acq On : 27 May 2021 12:02
Operator : CG/JU
Sample : M2411-06
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-RW8-GW-798-800

Quant Time: May 27 12:35:04 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 01:39:07 2021
Response via : Initial Calibration

