

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052821\
 Data File : BN014801.D
 Acq On : 28 May 2021 10:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 6/1/2021 9:16:20 AM

Quant Time: May 28 12:33:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 27 01:39:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	790	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	2677	0.40	ng	-0.01
13) Acenaphthene-d10	14.359	164	1850	0.40	ng	0.00
19) Phenanthrene-d10	17.116	188	4246	0.40	ng	0.00
29) Chrysene-d12	21.311	240	5039	0.40	ng	#-0.01
35) Perylene-d12	23.579	264	5155	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	933	0.44	ng	0.00
5) Phenol-d6	6.895	99	1183	0.41	ng	0.00
8) Nitrobenzene-d5	8.870	82	803	0.33	ng	0.00
11) 2-Methylnaphthalene-d10	12.093	152	1903	0.35	ng	0.00
14) 2,4,6-Tribromophenol	15.856	330	363	0.34	ng	-0.01
15) 2-Fluorobiphenyl	12.976	172	2698	0.40	ng	-0.01
27) Fluoranthene-d10	19.151	212	5417	0.32	ng	0.00
31) Terphenyl-d14	19.757	244	4784	0.40	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.271	88	426	0.46	ng	# 92
3) n-Nitrosodimethylamine	3.641	42	24	0.26	ng	# 42
6) bis(2-Chloroethyl)ether	7.145	93	873	0.36	ng	99
9) Naphthalene	10.543	128	2962	0.38	ng	94
10) Hexachlorobutadiene	10.835	225	810	0.40	ng	# 98
12) 2-Methylnaphthalene	12.169	142	2019	0.36	ng	99
16) Acenaphthylene	14.080	152	2673	0.36	ng	99
17) Acenaphthene	14.422	154	2005	0.37	ng	99
18) Fluorene	15.412	166	2712	0.35	ng	97
20) 4,6-Dinitro-2-methylph...	15.501	198	119	0.34	ng	95
21) 4-Bromophenyl-phenylether	16.313	248	1081	0.37	ng	93
22) Hexachlorobenzene	16.422	284	1290	0.40	ng	99
23) Atrazine	16.580	200	563	0.30	ng	99
24) Pentachlorophenol	16.775	266	307	0.45	ng	91
25) Phenanthrene	17.152	178	4559	0.36	ng	99
26) Anthracene	17.250	178	3777	0.35	ng	96
28) Fluoranthene	19.181	202	5832	0.33	ng	99
30) Pyrene	19.543	202	5866	0.41	ng	99
32) Benzo(a)anthracene	21.300	228	5225	0.34	ng	98
33) Chrysene	21.353	228	6819	0.40	ng	99
34) Bis(2-ethylhexyl)phtha...	21.237	149	1455	0.33	ng	100
36) Indeno(1,2,3-cd)pyrene	25.862	276	8108	0.38	ng	98
37) Benzo(b)fluoranthene	22.898	252	7228m	0.38	ng	
38) Benzo(k)fluoranthene	22.942	252	7541	0.38	ng	100
39) Benzo(a)pyrene	23.480	252	6488	0.39	ng	# 79
40) Dibenzo(a,h)anthracene	25.876	278	6539	0.37	ng	97
41) Benzo(g,h,i)perylene	26.553	276	7299	0.38	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052821\
Data File : BN014801.D
Acq On : 28 May 2021 10:54
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM


Quant Time: May 28 12:33:47 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN052621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 27 01:39:07 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion: 152 Resp: 790
Ion Ratio Lower Upper
152 100
150 149.6 123.4 185.2
115 80.9 73.2 109.8

Manual Integrations
APPROVED



mohammad
6/1/2021 9:16:20 AM




#2
1,4-Dioxane
Concen: 0.46 ng
RT: 3.271 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54



Tgt Ion: 88 Resp: 426
Ion Ratio Lower Upper
88 100
58 53.5 44.7 67.1
43 17.4 21.0 31.6#








#3
n-Nitrosodimethylamine
Concen: 0.26 ng
RT: 3.641 min Scan# 107
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	42	Resp:	24
Ion Ratio	Lower	Upper	
42	100		
74	0.0	0.0	0.0
44	41.7	80.0	120.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#4
2-Fluorophenol
Concen: 0.44 ng
RT: 5.317 min Scan# 315
Delta R.T. -0.008 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	112	Resp:	933
Ion Ratio	Lower	Upper	
112	100		
64	56.4	43.6	65.4
63	29.6	25.0	37.6









#5
Phenol-d6
Concen: 0.41 ng
RT: 6.895 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	99	Resp:	1183
Ion	Ratio	Lower	Upper
99	100		
42	11.4	10.2	15.2
71	46.2	32.4	48.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.36 ng
RT: 7.145 min Scan# 542
Delta R.T. -0.008 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	93	Resp:	873
Ion	Ratio	Lower	Upper
93	100		
63	60.1	47.5	71.3
95	34.4	28.5	42.7









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.013 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	136	Resp:	2677
Ion	Ratio	Lower	Upper
136	100		
137	13.4	13.2	19.8
54	10.5	9.0	13.4
68	9.7	8.8	13.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#8
Nitrobenzene-d5
Concen: 0.33 ng
RT: 8.870 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	82	Resp:	803
Ion	Ratio	Lower	Upper
82	100		
128	41.5	34.2	51.2
54	40.2	38.4	57.6









#9
Naphthalene
Concen: 0.38 ng
RT: 10.543 min Scan# 863
Delta R.T. -0.013 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	128	Resp:	2962
Ion Ratio	Lower	Upper	
128	100		
129	13.3	13.0	19.4
127	16.7	15.0	22.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM




#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.835 min Scan# 886
Delta R.T. -0.013 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	225	Resp:	810
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.1	50.8	76.2









#11
2-Methylnaphthalene-d10
Concen: 0.35 ng
RT: 12.093 min Scan# 1071
Delta R.T. -0.009 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54





Tgt Ion:152 Resp: 1903
Ion Ratio Lower Upper
152 100
151 20.6 16.8 25.2



#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 12.169 min Scan# 1088
Delta R.T. -0.004 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Tgt Ion:142 Resp: 2019
Ion Ratio Lower Upper
142 100
141 90.9 72.9 109.3
115 53.4 41.8 62.6



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.359 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:164 Resp: 1850
Ion Ratio Lower Upper
164 100
162 103.5 83.2 124.8
160 48.2 41.8 62.6

Manual Integrations
APPROVED



mohammad
6/1/2021 9:16:20 AM




#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.856 min Scan# 1428
Delta R.T. -0.013 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Tgt Ion:330 Resp: 363
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 28.1 21.8 32.6








#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.976 min Scan# 1201
Delta R.T. -0.013 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Tgt Ion:172 Resp: 2698

Ion	Ratio	Lower	Upper
172	100		
171	39.2	30.9	46.3
170	26.8	21.4	32.0





#16
Acenaphthylene
Concen: 0.36 ng
RT: 14.080 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54



Tgt Ion:152 Resp: 2673


Ion	Ratio	Lower	Upper
152	100		
151	21.1	17.1	25.7
153	13.5	9.9	14.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM







#17
Acenaphthene
Concen: 0.37 ng
RT: 14.422 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	154	Resp:	2005
Ion Ratio	Lower	Upper	
154	100		
153	116.4	91.8	137.8
152	58.2	46.0	69.0

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM




#18
Fluorene
Concen: 0.35 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	166	Resp:	2712
Ion Ratio	Lower	Upper	
166	100		
165	97.3	79.9	119.9
167	14.2	10.5	15.7









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.116 min Scan# 1531
Delta R.T. -0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54





Tgt Ion:	188	Resp:	4246
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.0	8.2	12.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.34 ng
RT: 15.501 min Scan# 1400
Delta R.T. -0.012 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Tgt Ion:	198	Resp:	119
Ion Ratio	Lower	Upper	
198	100		
51	114.1	96.4	144.6
105	85.9	72.3	108.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM






#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.313 min Scan# 1465
Delta R.T. 0.001 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4





Tgt Ion:248 Resp: 1081
Ion Ratio Lower Upper
248 100
250 101.4 77.5 116.3
141 54.6 50.2 75.4




#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.422 min Scan# 1474
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54



Tgt Ion:284 Resp: 1290
Ion Ratio Lower Upper
284 100
142 25.4 20.3 30.5
249 32.3 26.4 39.6









#23
Atrazine
Concen: 0.30 ng
RT: 16.580 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:	200	Resp:	563
Ion Ratio	Lower	Upper	
200	100		
173	36.5	29.1	43.7
215	57.5	45.1	67.7

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#24
Pentachlorophenol
Concen: 0.45 ng
RT: 16.775 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54


Tgt Ion:	266	Resp:	307
Ion Ratio	Lower	Upper	
266	100		
264	57.7	44.2	66.2
268	66.1	44.2	66.2





#25
Phenanthrene
Concen: 0.36 ng
RT: 17.152 min Scan# 1534
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:178 Resp: 4559
Ion Ratio Lower Upper
178 100
176 19.4 16.0 24.0
179 15.4 12.6 19.0

Manual Integrations
APPROVED



mohammad
6/1/2021 9:16:20 AM




#26
Anthracene
Concen: 0.35 ng
RT: 17.250 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54



Tgt Ion:178 Resp: 3777
Ion Ratio Lower Upper
178 100
176 19.0 14.2 21.2
179 15.0 13.7 20.5









#27
Fluoranthene-d10
Concen: 0.32 ng
RT: 19.151 min Scan# 1835
Delta R.T. -0.005 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion	Ratio	Lower	Upper
212	100		
106	9.5	7.4	11.2
104	5.6	4.1	6.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM




#28
Fluoranthene
Concen: 0.33 ng
RT: 19.181 min Scan# 1841
Delta R.T. -0.005 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion	Ratio	Lower	Upper
202	100		
101	8.7	7.0	10.6
203	17.4	13.5	20.3









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.311 min Scan# 2130
Delta R.T. -0.011 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54





Tgt Ion:240 Resp: 5039
Ion Ratio Lower Upper
240 100
120 9.9 5.2 7.8#
236 28.7 21.6 32.4



#30
Pyrene
Concen: 0.41 ng
RT: 19.543 min Scan# 1914
Delta R.T. -0.010 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Tgt Ion:202 Resp: 5866
Ion Ratio Lower Upper
202 100
200 21.2 16.6 24.8
203 17.4 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM



#31
Terphenyl-d14
Concen: 0.40 ng
RT: 19.757 min Scan# 1957
Delta R.T. -0.005 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:244 Resp: 4784
Ion Ratio Lower Upper
244 100
212 8.7 6.8 10.2
122 8.6 6.6 9.8




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.300 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:228 Resp: 5225
Ion Ratio Lower Upper
228 100
226 27.9 22.9 34.3
229 21.3 16.2 24.2





#33

Chrysene

Concen: 0.40 ng

RT: 21.353 min Scan# 2134

Delta R.T. 0.000 min

Lab File: BN014801.D

Acq: 28 May 2021 10:54

Tgt Ion:	228	Resp:	6819
Ion Ratio	Lower	Upper	
228	100		
226	30.1	24.3	36.5
229	20.7	16.4	24.6

Instrument :





BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.33 ng




RT: 21.237 min Scan# 2123


Delta R.T. -0.000 min

Lab File: BN014801.D

Acq: 28 May 2021 10:54

Tgt Ion:	149	Resp:	1455
Ion Ratio	Lower	Upper	
149	100		
167	35.5	28.3	42.5
279	6.9	5.5	8.3





#35

Perylene-d12

Concen: 0.40 ng

RT: 23.579 min Scan# 2576

Delta R.T. -0.007 min

Lab File: BN014801.D

Acq: 28 May 2021 10:54

Tgt Ion:	264	Resp:	5155
Ion Ratio	Lower	Upper	
264	100		
260	26.8	21.4	32.2
265	48.4	40.9	61.3

Instrument :





BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.38 ng




RT: 25.862 min Scan# 3243


Delta R.T. -0.007 min

Lab File: BN014801.D

Acq: 28 May 2021 10:54

Tgt Ion:	276	Resp:	8108
Ion Ratio	Lower	Upper	
276	100		
138	10.6	10.2	15.4
277	25.7	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.38 ng m
RT: 22.898 min Scan# 2377
Delta R.T. -0.003 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:252 Resp: 7228
Ion Ratio Lower Upper
252 100
253 28.7 21.5 32.3
125 15.5 6.6 10.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM



#38
Benzo(k)fluoranthene
Concen: 0.38 ng
RT: 22.942 min Scan# 2390
Delta R.T. -0.007 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion:252 Resp: 7541
Ion Ratio Lower Upper
252 100
253 28.4 22.7 34.1
125 8.3 6.4 9.6









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.480 min Scan# 2547
Delta R.T. -0.007 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion	Ratio	Lower	Upper
252	100		
253	36.3	25.1	37.7
125	31.7	9.4	14.0#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/1/2021 9:16:20 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.876 min Scan# 3247
Delta R.T. -0.010 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54

Tgt Ion	Ratio	Lower	Upper
278	100		
139	12.0	8.6	13.0
279	26.9	22.8	34.2







#41
Benzo(g,h,i)perylene
Concen: 0.38 ng
RT: 26.553 min Scan# 3445
Delta R.T. -0.014 min
Lab File: BN014801.D
Acq: 28 May 2021 10:54



Tgt Ion:	276	Resp:	7299
Ion Ratio	Lower	Upper	
276	100		
277	25.8	20.9	31.3
138	12.7	10.3	15.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/1/2021 9:16:20 AM