

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037114.D
 Acq On : 28 May 2025 18:23
 Operator : RC/JU
 Sample : Q2082-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-182-GW-260-262

Quant Time: May 28 18:56:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	1698	0.400	ng	0.00
7) Naphthalene-d8	10.394	136	4547	0.400	ng	-0.01
13) Acenaphthene-d10	14.256	164	2504	0.400	ng	-0.01
19) Phenanthrene-d10	17.009	188	4885	0.400	ng	# 0.00
29) Chrysene-d12	21.198	240	3172	0.400	ng	0.00
35) Perylene-d12	23.398	264	2643	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.206	112	829	0.186	ng	0.00
5) Phenol-d6	6.788	99	720	0.129	ng	0.00
8) Nitrobenzene-d5	8.760	82	1238	0.250	ng	-0.01
11) 2-Methylnaphthalene-d10	11.986	152	1631	0.255	ng	-0.01
14) 2,4,6-Tribromophenol	15.755	330	389	0.354	ng	-0.01
15) 2-Fluorobiphenyl	12.878	172	3942	0.344	ng	-0.01
27) Fluoranthene-d10	19.040	212	3792	0.283	ng	0.00
31) Terphenyl-d14	19.644	244	3379	0.498	ng	-0.01
Target Compounds						
9) Naphthalene	10.436	128	337	0.025	ng	# 63
25) Phenanthrene	17.046	178	393	0.025	ng	# 94
30) Pyrene	19.435	202	348	0.026	ng	97
34) Bis(2-ethylhexyl)phtha...	21.126	149	2004	0.273	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037114.D
 Acq On : 28 May 2025 18:23
 Operator : RC/JU
 Sample : Q2082-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-182-GW-260-262

Quant Time: May 28 18:56:36 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

