

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037124.D
 Acq On : 29 May 2025 00:23
 Operator : RC/JU
 Sample : PB168100BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB168100BS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/29/2025
 Supervised By :Jagrut Upadhyay 05/29/2025

Quant Time: May 29 01:34:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	2482	0.400	ng	0.00
7) Naphthalene-d8	10.383	136	6249	0.400	ng	#-0.02
13) Acenaphthene-d10	14.256	164	2970	0.400	ng	-0.01
19) Phenanthrene-d10	17.009	188	4819	0.400	ng	# 0.00
29) Chrysene-d12	21.198	240	3074	0.400	ng	0.00
35) Perylene-d12	23.395	264	3112	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.206	112	2242	0.345	ng	0.00
5) Phenol-d6	6.788	99	2593	0.319	ng	0.00
8) Nitrobenzene-d5	8.760	82	2259	0.332	ng	-0.01
11) 2-Methylnaphthalene-d10	11.986	152	3317m	0.377	ng	-0.01
14) 2,4,6-Tribromophenol	15.755	330	326	0.250	ng	-0.01
15) 2-Fluorobiphenyl	12.873	172	4707	0.346	ng	-0.02
27) Fluoranthene-d10	19.040	212	3641	0.276	ng	0.00
31) Terphenyl-d14	19.649	244	2422	0.368	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.133	88	861	0.283	ng	# 36
3) n-Nitrosodimethylamine	3.444	42	2136	0.326	ng	# 90
6) bis(2-Chloroethyl)ether	7.041	93	2451	0.327	ng	99
9) Naphthalene	10.436	128	6160	0.334	ng	98
10) Hexachlorobutadiene	10.725	225	1373	0.354	ng	# 99
12) 2-Methylnaphthalene	12.062	142	3449	0.290	ng	98
16) Acenaphthylene	13.967	152	5388	0.373	ng	100
17) Acenaphthene	14.320	154	3199	0.339	ng	100
18) Fluorene	15.304	166	3917	0.316	ng	100
20) 4,6-Dinitro-2-methylph...	15.400	198	334	0.363	ng	91
21) 4-Bromophenyl-phenylether	16.202	248	1107	0.364	ng	96
22) Hexachlorobenzene	16.314	284	1238	0.380	ng	99
23) Atrazine	16.475	200	861	0.324	ng	95
24) Pentachlorophenol	16.661	266	517	0.288	ng	97
25) Phenanthrene	17.046	178	5320	0.338	ng	99
26) Anthracene	17.133	178	4823	0.337	ng	99
28) Fluoranthene	19.068	202	4934	0.262	ng	98
30) Pyrene	19.430	202	4842	0.368	ng	99
32) Benzo(a)anthracene	21.180	228	3881	0.335	ng	98
33) Chrysene	21.233	228	4404	0.360	ng	98
34) Bis(2-ethylhexyl)phtha...	21.126	149	2227	0.313	ng	99
36) Indeno(1,2,3-cd)pyrene	25.599	276	4982	0.392	ng	99
37) Benzo(b)fluoranthene	22.737	252	4085	0.316	ng	95
38) Benzo(k)fluoranthene	22.781	252	4383m	0.344	ng	
39) Benzo(a)pyrene	23.298	252	3942	0.360	ng	93
40) Dibenzo(a,h)anthracene	25.614	278	3842	0.388	ng	97
41) Benzo(g,h,i)perylene	26.272	276	4293	0.399	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037124.D
 Acq On : 29 May 2025 00:23
 Operator : RC/JU
 Sample : PB168100BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB168100BS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/29/2025
 Supervised By :Jagrut Upadhyay 05/29/2025

Quant Time: May 29 01:34:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

