

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052825\
 Data File : BN037110.D
 Acq On : 28 May 2025 15:54
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 28 17:04:58 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN051425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 14 11:26:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	-0.01
2	1,4-Dioxane	0.400	0.390	2.5	99	0.00
3	n-Nitrosodimethylamine	0.400	0.408	-2.0	116	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	88	-0.01
5 S	Phenol-d6	0.400	0.341	14.8	86	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.365	8.8	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	-0.02
8 S	Nitrobenzene-d5	0.400	0.377	5.8	110	-0.02
9	Naphthalene	0.400	0.375	6.3	104	-0.01
10	Hexachlorobutadiene	0.400	0.411	-2.7	112	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	112	-0.01
12	2-Methylnaphthalene	0.400	0.381	4.8	105	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.372	7.0	99	-0.01
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	101	-0.02
16	Acenaphthylene	0.400	0.374	6.5	103	-0.02
17	Acenaphthene	0.400	0.378	5.5	104	-0.01
18	Fluorene	0.400	0.375	6.3	103	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.395	1.3	106	-0.01
21	4-Bromophenyl-phenylether	0.400	0.401	-0.3	104	-0.01
22	Hexachlorobenzene	0.400	0.419	-4.7	103	-0.01
23	Atrazine	0.400	0.380	5.0	101	-0.01
24	Pentachlorophenol	0.400	0.333	16.8	90	-0.01
25	Phenanthrene	0.400	0.378	5.5	98	-0.01
26	Anthracene	0.400	0.365	8.8	95	-0.01
27 SURR	Fluoranthene-d10	0.400	0.365	8.8	95	0.00
28	Fluoranthene	0.400	0.337	15.8	90	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	76	0.00
30	Pyrene	0.400	0.454	-13.5	86	0.00
31 S	Terphenyl-d14	0.400	0.462	-15.5	87	0.00
32	Benzo(a)anthracene	0.400	0.371	7.3	72	0.00
33	Chrysene	0.400	0.381	4.8	72	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.380	5.0	74	0.00
35 I	Perylene-d12	0.400	0.400	0.0	65	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.393	1.8	64	-0.03
37	Benzo(b)fluoranthene	0.400	0.370	7.5	63	-0.01
38	Benzo(k)fluoranthene	0.400	0.391	2.3	64	-0.01
39 C	Benzo(a)pyrene	0.400	0.366	8.5	61	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.397	0.8	65	-0.02
41	Benzo(g,h,i)perylene	0.400	0.416	-4.0	66	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4