




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN053023\
Data File : BN025726.D
Acq On : 30 May 2023 21:21
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 31 07:40:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 31 07:38:46 2023
Response via : Initial Calibration



TIC: BN025726.D\data.ms

(70) Pentachlorophenol (C)




17.034min (+ 0.000) 13510.83 ng

response 257652

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.89
264.00	61.60	64.66
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN053023\
 Data File : BN025726.D
 Acq On : 30 May 2023 21:21
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 31 07:40:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 31 07:38:46 2023
 Response via : Initial Calibration



TIC: BN025726.D\data.ms

(77) Benzidine

19.622min (-0.006) 0.00 ng

response 2329028

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.60
183.00	13.20	11.10
0.00	0.00	0.00