

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN053023\
 Data File : BN025717.D
 Acq On : 30 May 2023 13:24
 Operator : CG/JU
 Sample : 02912-02MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GM38-RW3-MW1-20230523MS

Quant Time: May 31 05:38:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 29 00:47:45 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 05/31/2023
 Supervised By :Jagrut Upadhyay 05/31/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.019	152	6495	0.400	ng	0.00
7) Naphthalene-d8	10.835	136	18190	0.400	ng	0.00
13) Acenaphthene-d10	14.651	164	9860	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	19436	0.400	ng	# 0.00
29) Chrysene-d12	21.578	240	9266	0.400	ng	# 0.00
35) Perylene-d12	24.054	264	8496	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	2097	0.126	ng	0.00
5) Phenol-d6	7.182	99	1575	0.079	ng	0.00
8) Nitrobenzene-d5	9.180	82	5117	0.336	ng	-0.01
11) 2-Methylnaphthalene-d10	12.415	152	9618	0.371	ng	0.00
14) 2,4,6-Tribromophenol	16.140	330	828	0.415	ng	-0.01
15) 2-Fluorobiphenyl	13.272	172	15360	0.377	ng	-0.01
27) Fluoranthene-d10	19.416	212	14547	0.330	ng	0.00
31) Terphenyl-d14	20.006	244	10335	0.509	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.390	88	19414	2.581	ng	100
3) n-Nitrosodimethylamine	3.765	42	637	0.079	ng	# 85
6) bis(2-Chloroethyl)ether	7.434	93	6129	0.298	ng	# 85
9) Naphthalene	10.878	128	16740	0.336	ng	99
10) Hexachlorobutadiene	11.166	225	2681	0.311	ng	# 97
12) 2-Methylnaphthalene	12.487	142	10843	0.317	ng	99
16) Acenaphthylene	14.373	152	15163	0.326	ng	99
17) Acenaphthene	14.705	154	10458	0.341	ng	99
18) Fluorene	15.693	166	13760	0.339	ng	100
20) 4,6-Dinitro-2-methylph...	15.780	198	1110	0.522	ng	# 61
21) 4-Bromophenyl-phenylether	16.574	248	3896	0.353	ng	91
22) Hexachlorobenzene	16.698	284	3596	0.357	ng	97
23) Atrazine	16.847	200	3407	0.387	ng	# 92
24) Pentachlorophenol	17.046	266	1553	1.229	ng	# 83
25) Phenanthrene	17.430	178	21910	0.370	ng	100
26) Anthracene	17.530	178	18346	0.343	ng	100
28) Fluoranthene	19.449	202	20834	0.323	ng	99
30) Pyrene	19.811	202	20874	0.463	ng	99
32) Benzo(a)anthracene	21.560	228	12909	0.374	ng	99
33) Chrysene	21.614	228	14436	0.385	ng	100
34) Bis(2-ethylhexyl)phtha...	21.471	149	8174	0.389	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.636	276	14733	0.452	ng	# 62
37) Benzo(b)fluoranthene	23.291	252	12809	0.396	ng	97
38) Benzo(k)fluoranthene	23.341	252	12504	0.371	ng	96
39) Benzo(a)pyrene	23.940	252	10640	0.344	ng	94
40) Dibenzo(a,h)anthracene	26.656	278	11901m	0.461	ng	
41) Benzo(g,h,i)perylene	27.437	276	13584	0.461	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN053023\
Data File : BN025717.D
Acq On : 30 May 2023 13:24
Operator : CG/JU
Sample : 02912-02MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :


GM38-RW3-MW1-20230523MS


Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023


Quant Time: May 31 05:38:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 29 00:47:45 2023
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.019 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion: 152 Resp: 649
Ion Ratio Lower Upper
152 100
150 152.6 121.0 181.6
115 58.7 48.5 72.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#2
1,4-Dioxane
Concen: 2.581 ng
RT: 3.390 min Scan# 29
Delta R.T. -0.007 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion: 88 Resp: 19414
Ion Ratio Lower Upper
88 100
43 33.2 26.5 39.7
58 71.4 56.8 85.2





#3
n-Nitrosodimethylamine
Concen: 0.079 ng
RT: 3.765 min Scan# 81
Delta R.T. 0.022 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion: 42 Resp: 63
Ion Ratio Lower Upper
42 100
74 117.6 82.2 123.2
44 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#4
2-Fluorophenol
Concen: 0.126 ng
RT: 5.564 min Scan# 330
Delta R.T. 0.007 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:112 Resp: 2097
Ion Ratio Lower Upper
112 100
64 60.3 45.7 68.5
63 33.8 26.2 39.2





#5
Phenol-d6
Concen: 0.079 ng
RT: 7.182 min Scan# 554
Delta R.T. 0.007 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion: 99 Resp: 1579
Ion Ratio Lower Upper
99 100
42 24.4 17.1 25.7
71 44.6 29.3 43.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#6
bis(2-Chloroethyl)ether
Concen: 0.298 ng
RT: 7.434 min Scan# 589
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion: 93 Resp: 6129
Ion Ratio Lower Upper
93 100
63 76.1 49.1 73.7#
95 31.9 22.6 33.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.835 min Scan# 948
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument : BNA_N
ClientSampleId : GM38-RW3-MW1-20230523MS



Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.1	13.7
54	7.2	5.4	8.0
68	5.4	4.3	6.5




Manual Integrations APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#8
Nitrobenzene-d5
Concen: 0.336 ng
RT: 9.180 min Scan# 792
Delta R.T. -0.011 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Tgt Ion	Ratio	Lower	Upper
82	100		
128	35.7	32.5	48.7
54	55.0	40.6	60.8





#9

Naphthalene

Concen: 0.336 ng

RT: 10.878 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN025717.D

Acq: 30 May 2023 13:24

Instrument :

BNA_N

ClientSampleId :





GM38-RW3-MW1-20230523MS

Tgt Ion:	128	Resp:	16740
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.3	13.9
127	13.4	11.0	16.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#10

Hexachlorobutadiene

Concen: 0.311 ng




RT: 11.166 min Scan# 978


Delta R.T. -0.011 min

Lab File: BN025717.D

Acq: 30 May 2023 13:24


Tgt Ion:	225	Resp:	2681
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.4	48.4	72.6





#11
2-Methylnaphthalene-d10
Concen: 0.371 ng
RT: 12.415 min Scan# 1250
Delta R.T. -0.008 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS




Tgt Ion:152 Resp: 9618
Ion Ratio Lower Upper
152 100
151 18.3 16.1 24.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023




#12
2-Methylnaphthalene
Concen: 0.317 ng
RT: 12.487 min Scan# 1269
Delta R.T. -0.008 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24



Tgt Ion:142 Resp: 10843
Ion Ratio Lower Upper
142 100
141 89.1 70.7 106.1
115 40.6 31.5 47.3






#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.651 min Scan# 1560
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Instrument :

BNA_N

ClientSampleId :

GM38-RW3-MW1-20230523MS






Tgt Ion:164 Resp: 9860
Ion Ratio Lower Upper
164 100
162 101.0 86.4 129.6
160 48.3 41.9 62.9


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023




#14
2,4,6-Tribromophenol
Concen: 0.415 ng
RT: 16.140 min Scan# 1691
Delta R.T. -0.012 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Tgt Ion:330 Resp: 828
Ion Ratio Lower Upper
330 100
332 95.5 80.0 120.0
141 62.2 45.8 68.6





#15
2-Fluorobiphenyl
Concen: 0.377 ng
RT: 13.272 min Scan# 1432
Delta R.T. -0.011 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS




Tgt Ion:172 Resp: 15360
Ion Ratio Lower Upper
172 100
171 35.1 28.6 42.8
170 24.4 19.7 29.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023




#16
Acenaphthylene
Concen: 0.326 ng
RT: 14.373 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Tgt Ion:152 Resp: 15163
Ion Ratio Lower Upper
152 100
151 19.4 15.8 23.6
153 12.9 10.6 15.8





#17
Acenaphthene
Concen: 0.341 ng
RT: 14.705 min Scan# 1565
Delta R.T. -0.010 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:154 Resp: 10458
Ion Ratio Lower Upper
154 100
153 115.3 91.9 137.9
152 57.4 45.4 68.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#18
Fluorene
Concen: 0.339 ng
RT: 15.693 min Scan# 1655
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Tgt Ion:166 Resp: 13760
Ion Ratio Lower Upper
166 100
165 99.2 79.3 118.9
167 13.7 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.393 min Scan# 1792
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Instrument : BNA_N
ClientSampleId : GM38-RW3-MW1-20230523MS

Tgt Ion:188 Resp: 19430
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.8 12.2 18.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#20
4,6-Dinitro-2-methylphenol
Concen: 0.522 ng
RT: 15.780 min Scan# 1662
Delta R.T. -0.012 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:198 Resp: 1110
Ion Ratio Lower Upper
198 100
51 79.6 110.8 166.2#
105 59.7 61.9 92.9#





#21
4-Bromophenyl-phenylether
Concen: 0.353 ng
RT: 16.574 min Scan# 1726
Delta R.T. -0.013 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:248 Resp: 3890
Ion Ratio Lower Upper
248 100
250 93.8 77.5 116.3
141 94.5 64.0 96.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#22
Hexachlorobenzene
Concen: 0.357 ng
RT: 16.698 min Scan# 1736
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:284 Resp: 3596
Ion Ratio Lower Upper
284 100
142 55.0 42.0 63.0
249 35.0 28.6 43.0





#23
Atrazine
Concen: 0.387 ng
RT: 16.847 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:200 Resp: 340
Ion Ratio Lower Upper
200 100
173 24.9 27.9 41.9
215 44.9 35.1 52.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#24
Pentachlorophenol
Concen: 1.229 ng
RT: 17.046 min Scan# 1764
Delta R.T. -0.025 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Tgt Ion:266 Resp: 1553
Ion Ratio Lower Upper
266 100
264 60.9 60.0 90.0
268 66.6 0.0 0.0#









#25
Phenanthrene
Concen: 0.370 ng
RT: 17.430 min Scan# 1796
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS

Tgt Ion:178 Resp: 21910
Ion Ratio Lower Upper
178 100
176 19.4 15.4 23.2
179 15.5 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#26
Anthracene
Concen: 0.343 ng
RT: 17.530 min Scan# 1803
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:178 Resp: 18346
Ion Ratio Lower Upper
178 100
176 18.4 14.8 22.2
179 15.3 12.2 18.2





#27
Fluoranthene-d10
Concen: 0.330 ng
RT: 19.416 min Scan# 2077
Delta R.T. -0.005 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS




Tgt Ion:212 Resp: 1454
Ion Ratio Lower Upper
212 100
106 16.0 12.3 18.5
104 9.5 7.2 10.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023




#28
Fluoranthene
Concen: 0.323 ng
RT: 19.449 min Scan# 2084
Delta R.T. -0.005 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Tgt Ion:202 Resp: 20834
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 17.1 13.3 19.9





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.578 min Scan# 2483
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS




Tgt Ion:240 Resp: 9260
Ion Ratio Lower Upper
240 100
120 14.4 15.4 23.0
236 29.7 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023




#30
Pyrene
Concen: 0.463 ng
RT: 19.811 min Scan# 2162
Delta R.T. -0.005 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Tgt Ion:202 Resp: 20874
Ion Ratio Lower Upper
202 100
200 21.1 17.1 25.7
203 17.6 14.5 21.7





#31
Terphenyl-d14
Concen: 0.509 ng
RT: 20.006 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:244 Resp: 1033
Ion Ratio Lower Upper
244 100
212 10.7 8.5 12.7
122 20.0 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#32
Benzo(a)anthracene
Concen: 0.374 ng
RT: 21.560 min Scan# 2481
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:228 Resp: 12909
Ion Ratio Lower Upper
228 100
226 28.3 22.9 34.3
229 20.3 15.9 23.9





#33
Chrysene
Concen: 0.385 ng
RT: 21.614 min Scan# 2487
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS




Tgt Ion: 228 Resp: 14430
Ion Ratio Lower Upper
228 100
226 30.7 24.5 36.7
229 20.4 16.2 24.4

Manual Integrations
APPROVED



Reviewed By : Christian Giraldo 05/31/2023
Supervised By : Jagrut Upadhyay 05/31/2023




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.389 ng
RT: 21.471 min Scan# 2471
Delta R.T. -0.000 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24



Tgt Ion: 149 Resp: 8174
Ion Ratio Lower Upper
149 100
167 26.6 20.4 30.6
279 3.0 3.3 4.9





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.054 min Scan# 31

Delta R.T. -0.015 min

Lab File: BN025717.D


Acq: 30 May 2023 13:24

Instrument :

BNA_N

ClientSampleId :

GM38-RW3-MW1-20230523MS



Tgt Ion:264 Resp: 8490

Ion Ratio Lower Upper

264 100

260 26.8 22.3 33.5



265 67.9 54.3 81.5

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/31/2023

Supervised By :Jagrut Upadhyay 05/31/2023



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.452 ng

RT: 26.636 min Scan# 4021

Delta R.T. -0.044 min

Lab File: BN025717.D

Acq: 30 May 2023 13:24




Tgt Ion:276 Resp: 14733


Ion Ratio Lower Upper

276 100

138 29.7 10.5 15.7#

277 24.4 9.2 13.8#





#37

Benzo(b)fluoranthene

Concen: 0.396 ng

RT: 23.291 min Scan# 2877

Delta R.T. -0.012 min

Lab File: BN025717.D

Acq: 30 May 2023 13:24

Instrument :

BNA_N

ClientSampleId :





GM38-RW3-MW1-20230523MS

Tgt Ion:252	Resp:	12809
Ion Ratio	Lower	Upper
252	100	
253	34.3	26.2 39.4
125	17.2	12.9 19.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#38

Benzo(k)fluoranthene

Concen: 0.371 ng




RT: 23.341 min Scan# 2894


Delta R.T. -0.015 min

Lab File: BN025717.D

Acq: 30 May 2023 13:24


Tgt Ion:252	Resp:	12504
Ion Ratio	Lower	Upper
252	100	
253	35.1	26.1 39.1
125	16.8	12.5 18.7





#39
Benzo(a)pyrene
Concen: 0.344 ng
RT: 23.940 min Scan# 3099
Delta R.T. -0.015 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24




Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:252 Resp: 10640
Ion Ratio Lower Upper
252 100
253 41.4 30.2 45.2
125 21.1 14.8 22.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023



#40
Dibenzo(a,h)anthracene
Concen: 0.461 ng m
RT: 26.656 min Scan# 4028
Delta R.T. -0.047 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24


Tgt Ion:278 Resp: 11901
Ion Ratio Lower Upper
278 100
139 22.1 18.6 28.0
279 35.6 27.8 41.8





#41
Benzo(g,h,i)perylene
Concen: 0.461 ng
RT: 27.437 min Scan# 41
Delta R.T. -0.032 min
Lab File: BN025717.D
Acq: 30 May 2023 13:24

Instrument :
BNA_N
ClientSampleId :
GM38-RW3-MW1-20230523MS



Tgt Ion:276 Resp: 13584
Ion Ratio Lower Upper
276 100
277 25.5 21.4 32.0
138 29.2 22.8 34.2

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 05/31/2023
Supervised By :Jagrut Upadhyay 05/31/2023

