

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060722\
 Data File : BN020057.D
 Acq On : 08 Jun 2022 09:07
 Operator : CG/JU
 Sample : PB145379BS
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS379

Quant Time: Jun 08 09:41:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
 Supervised By :mohammad ahmed 06/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	144592	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	673818	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	438090	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	896946	20.000	ng/u1	0.00
79) Chrysene-d12	21.386	240	824882	20.000	ng/u1	0.00
88) Perylene-d12	23.716	264	797444	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	21122	6.327	ng/uL	0.00
4) Pyridine-d5	3.687	84	297562	32.611	ng/u1	0.00
7) Phenol-d5	6.987	99	432626	36.159	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.152	67	260641	34.867	ng/u1	0.00
11) 2-Chlorophenol-d4	7.358	132	354168	36.431	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	373473	36.079	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	183984	35.890	ng/u1	0.00
24) 2-Nitrophenol-d4	9.693	143	202020	37.133	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	345659	36.395	ng/u1	0.00
31) 4-Chloroaniline-d4	10.746	131	476756	30.998	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1116585	35.029	ng/u1	0.00
49) Acenaphthylene-d8	14.146	160	1296624	35.076	ng/u1	0.00
54) 4-Nitrophenol-d4	14.646	143	231514	36.245	ng/u1	0.00
60) Fluorene-d10	15.440	176	927178	35.150	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	194198	37.377	ng/u1	0.00
73) Anthracene-d10	17.292	188	1415102	36.150	ng/u1	0.00
81) Pyrene-d10	19.586	212	1590955	35.577	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.569	264	1479699	37.924	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	41985	11.913	ng/uL#	91
5) Pyridine	3.705	79	292052	30.531	ng/u1	92
6) Benzaldehyde	6.958	77	249570	44.269	ng/u1	94
8) Phenol	7.017	94	429859	33.740	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.246	93	328192	32.913	ng/u1	98
12) 2-Chlorophenol	7.387	128	342380	33.694	ng/u1	98
13) 2-Methylphenol	8.264	108	329763	33.668	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.358	45	508985	32.551	ng/u1	97
16) Acetophenone	8.640	105	520156	33.043	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.634	70	266948	33.196	ng/u1	92
18) 4-Methylphenol	8.593	108	367394	33.986	ng/u1	96
19) Hexachloroethane	8.899	117	133432	32.634	ng/u1	87
22) Nitrobenzene	9.016	77	390513	32.855	ng/u1	93
23) Isophorone	9.540	82	771390	32.870	ng/u1	99
25) 2-Nitrophenol	9.728	139	207661	34.543	ng/u1	98
26) 2,4-Dimethylphenol	9.793	107	405919	32.592	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.028	93	463274	32.258	ng/u1	98
29) 2,4-Dichlorophenol	10.263	162	329961	33.586	ng/u1	98
30) Naphthalene	10.663	128	1127861	32.297	ng/u1	99
32) 4-Chloroaniline	10.769	127	480940	31.263	ng/u1	98
33) Hexachlorobutadiene	10.957	225	181309	32.145	ng/u1	98
34) Caprolactam	11.522	113	127964m	33.355	ng/u1	
35) 4-Chloro-3-methylphenol	11.899	107	402725	34.717	ng/u1	97
36) 2-Methylnaphthalene	12.275	142	782124	33.128	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060722\
 Data File : BN020057.D
 Acq On : 08 Jun 2022 09:07
 Operator : CG/JU
 Sample : PB145379BS
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS379


Quant Time: Jun 08 09:41:04 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 06 15:14:59 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
 Supervised By :mohammad ahmed 06/09/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.493	142	796398	33.056	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.646	216	367610	32.901	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	197753	29.095	ng/ul	98
41) 2,4,6-Trichlorophenol	12.887	196	261518	33.945	ng/ul	98
42) 2,4,5-Trichlorophenol	12.957	196	287033	33.640	ng/ul	98
43) 1,1'-Biphenyl	13.287	154	1051599	32.417	ng/ul	98
44) 2-Chloronaphthalene	13.328	162	797908	32.423	ng/ul	97
45) 2-Nitroaniline	13.528	65	274903	35.906	ng/ul	95
47) Dimethylphthalate	13.910	163	1047268	32.636	ng/ul	99
48) 2,6-Dinitrotoluene	14.022	165	227001	36.181	ng/ul	97
50) Acenaphthylene	14.169	152	1346756	32.906	ng/ul	99
51) 3-Nitroaniline	14.351	138	243204	36.667	ng/ul	95
52) Acenaphthene	14.516	153	865509	32.375	ng/ul	99
53) 2,4-Dinitrophenol	14.557	184	125701	31.293	ng/ul	96
55) 4-Nitrophenol	14.663	109	185426	34.008	ng/ul	97
56) Dibenzofuran	14.851	168	1207545	32.061	ng/ul	99
57) 2,4-Dinitrotoluene	14.810	165	332783	35.823	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.075	232	223483	33.366	ng/ul	89
59) Diethylphthalate	15.275	149	1100944	33.324	ng/ul	97
61) Fluorene	15.498	166	935009	31.903	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.493	204	429838	31.847	ng/ul	92
63) 4-Nitroaniline	15.510	138	250482	40.414	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.581	198	180973	34.356	ng/ul	99
67) N-Nitrosodiphenylamine	15.704	169	859978	33.444	ng/ul	99
68) 4-Bromophenyl-phenylether	16.387	248	264941	33.919	ng/ul	92
69) Hexachlorobenzene	16.504	284	304553	33.919	ng/ul#	89
70) Atrazine	16.657	200	244340	27.234	ng/ul	96
71) Pentachlorophenol	16.845	266	180613	32.297	ng/ul	98
72) Phenanthrene	17.234	178	1585644	33.498	ng/ul	99
74) Anthracene	17.328	178	1575511	33.101	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.251	216	388600	32.354	ng/ul	98
76) Pentachlorobenzene	14.775	250	350977	33.063	ng/ul	94
77) Carbazole	17.592	167	1530755	35.352	ng/ul	97
78) Di-n-butylphthalate	18.163	149	1919762	35.754	ng/ul	99
80) Fluoranthene	19.251	202	1827034	33.828	ng/ul	97
82) Pyrene	19.616	202	1850297	33.393	ng/ul	96
83) Butylbenzylphthalate	20.522	149	900305	36.104	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.298	252	578568	36.723	ng/ul#	98
85) Benzo(a)anthracene	21.369	228	1731749	33.246	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.304	149	1259798	35.806	ng/ul#	98
87) Chrysene	21.422	228	1687613	33.117	ng/ul	98
89) Di-n-octyl phthalate	22.210	149	2324194	35.809	ng/ul	100
90) Benzo(b)fluoranthene	23.010	252	1793693	35.839	ng/ul	98
91) Benzo(k)fluoranthene	23.057	252	1665954	34.967	ng/ul	98
93) Benzo(a)pyrene	23.616	252	1687782	34.838	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.110	276	1648903	33.081	ng/ul	94
95) Dibenzo(a,h)anthracene	26.127	278	1424298	33.455	ng/ul	97
96) Benzo(g,h,i)perylene	26.845	276	1399155	33.240	ng/ul#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.913 ng/uL
RT: 3.317 min Scan# 39
Delta R.T. 0.001 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion: 88 Resp: 41985
Ion Ratio Lower Upper
88 100
43 36.5 22.4 33.6
58 81.6 61.0 91.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#5
Pyridine
Concen: 30.531 ng/uL
RT: 3.705 min Scan# 105
Delta R.T. -0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion: 79 Resp: 292052
Ion Ratio Lower Upper
79 100
52 89.8 65.0 97.4
51 41.7 31.6 47.4





#6
Benzaldehyde
Concen: 44.269 ng/ul
RT: 6.958 min Scan# 658
Delta R.T. -0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion: 77 Resp: 249570
Ion Ratio Lower Upper
77 100
105 99.3 74.7 112.1
106 93.6 70.2 105.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#8
Phenol
Concen: 33.740 ng/ul
RT: 7.017 min Scan# 668
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion: 94 Resp: 429859
Ion Ratio Lower Upper
94 100
65 30.6 23.6 35.4
66 41.1 31.2 46.8





#10
Bis(2-Chloroethyl)ether
Concen: 32.913 ng/ul
RT: 7.246 min Scan# 707
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion: 93 Resp: 328192
Ion Ratio Lower Upper
93 100
63 80.2 63.4 95.0
95 31.4 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#12
2-Chlorophenol
Concen: 33.694 ng/ul
RT: 7.387 min Scan# 731
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion:128 Resp: 342380
Ion Ratio Lower Upper
128 100
64 47.0 38.6 57.8
130 32.8 25.0 37.4





#13
2-Methylphenol
Concen: 33.668 ng/ul
RT: 8.264 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion: 108 Resp: 32976
Ion Ratio Lower Upper
108 100
107 91.0 71.5 107.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.551 ng/ul
RT: 8.358 min Scan# 896
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion: 45 Resp: 508985
Ion Ratio Lower Upper
45 100
77 12.4 10.9 16.3
79 9.6 8.6 13.0





#16
Acetophenone
Concen: 33.043 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion: 105 Resp: 520150
Ion Ratio Lower Upper
105 100
77 77.1 63.5 95.3
51 30.2 23.5 35.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#17
N-Nitroso-di-n-propylamine
Concen: 33.196 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion: 70 Resp: 266948
Ion Ratio Lower Upper
70 100
42 67.0 47.8 71.6
101 10.3 7.8 11.6
130 22.9 15.8 23.8





#18
4-Methylphenol
Concen: 33.986 ng/ul
RT: 8.593 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion:108 Resp: 367394
Ion Ratio Lower Upper
108 100
107 113.9 88.0 132.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#19
Hexachloroethane
Concen: 32.634 ng/ul
RT: 8.899 min Scan# 988
Delta R.T. -0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion:117 Resp: 133432
Ion Ratio Lower Upper
117 100
201 87.0 83.2 124.8
199 56.9 50.2 75.4





#22
Nitrobenzene
Concen: 32.855 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion: 77 Resp: 390511
Ion Ratio Lower Upper
77 100
123 48.8 34.5 51.7
65 15.0 11.1 16.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#23
Isophorone
Concen: 32.870 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07



Tgt Ion: 82 Resp: 771390
Ion Ratio Lower Upper
82 100
95 7.7 5.8 8.8
138 17.9 13.8 20.8





#25

2-Nitrophenol

Concen: 34.543 ng/ul

RT: 9.728 min Scan# 1129

Delta R.T. 0.000 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07

Instrument :

BNA_N

ClientSampleId :

SLCS379

Tgt Ion:139 Resp: 20766

Ion Ratio Lower Upper

139 100





65 52.9 43.7 65.5

109 33.7 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#26

2,4-Dimethylphenol

Concen: 32.592 ng/ul

RT: 9.793 min Scan# 1140

Delta R.T. 0.000 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07




Tgt Ion:107 Resp: 405919


Ion Ratio Lower Upper

107 100

121 50.4 38.3 57.5


122 86.2 66.3 99.5





#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.258 ng/ul
 RT: 10.028 min Scan# 1179
 Delta R.T. 0.000 min
 Lab File: BN020057.D
 Acq: 08 Jun 2022 09:07




Instrument :
 BNA_N
 ClientSampleId :
 SLCS379




Tgt Ion: 93 Resp: 463274
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.4 38.0
 123 11.6 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
 Supervised By :mohammad ahmed 06/09/2022




#29
 2,4-Dichlorophenol
 Concen: 33.586 ng/ul
 RT: 10.263 min Scan# 1220
 Delta R.T. 0.000 min
 Lab File: BN020057.D
 Acq: 08 Jun 2022 09:07




Tgt Ion:162 Resp: 329961
 Ion Ratio Lower Upper
 162 100
 164 68.0 52.9 79.3
 98 38.0 29.8 44.8





#30
Naphthalene
Concen: 32.297 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:128 Resp: 112786
Ion Ratio Lower Upper
128 100
129 10.7 8.7 13.1
127 13.0 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#32
4-Chloroaniline
Concen: 31.263 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:127 Resp: 480940
Ion Ratio Lower Upper
127 100
129 31.6 26.1 39.1





#33
Hexachlorobutadiene
Concen: 32.145 ng/ul
RT: 10.957 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:225 Resp: 181309
Ion Ratio Lower Upper
225 100
223 62.9 51.8 77.6
227 64.9 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#34
Caprolactam
Concen: 33.355 ng/ul m
RT: 11.522 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:113 Resp: 127964
Ion Ratio Lower Upper
113 100
55 188.7 145.5 218.3
56 135.0 106.5 159.7





#35
4-Chloro-3-methylphenol
Concen: 34.717 ng/ul
RT: 11.899 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:107 Resp: 402725
Ion Ratio Lower Upper
107 100
144 25.9 20.6 31.0
142 81.1 62.2 93.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#36
2-Methylnaphthalene
Concen: 33.128 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07



Tgt Ion:142 Resp: 782124
Ion Ratio Lower Upper
142 100
141 88.9 70.6 105.8





#37

1-Methylnaphthalene

Concen: 33.056 ng/ul

RT: 12.493 min Scan# 1599

Delta R.T. 0.000 min

Lab File: BN020057.D


Acq: 08 Jun 2022 09:07

Instrument :

BNA_N

ClientSampleId :

SLCS379



Tgt Ion:142 Resp: 796398

Ion Ratio Lower Upper

142 100

141 91.9 70.9 106.3




116 3.9 3.1 4.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022

Supervised By :mohammad ahmed 06/09/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 32.901 ng/ul

RT: 12.646 min Scan# 1625

Delta R.T. -0.006 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07

Tgt Ion:216 Resp: 367610




Ion Ratio Lower Upper


216 100

214 78.2 63.4 95.0

179 23.4 17.8 26.8


108 21.2 15.1 22.7





#40
Hexachlorocyclopentadiene
Concen: 29.095 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:237 Resp: 19775
Ion Ratio Lower Upper
237 100
235 62.8 48.8 73.2
272 12.6 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#41
2,4,6-Trichlorophenol
Concen: 33.945 ng/ul
RT: 12.887 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:196 Resp: 261518
Ion Ratio Lower Upper
196 100
198 98.1 79.8 119.8
200 30.4 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 33.640 ng/ul
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:196 Resp: 28703
Ion Ratio Lower Upper
196 100
198 95.1 77.6 116.4
200 29.9 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#43
1,1'-Biphenyl
Concen: 32.417 ng/ul
RT: 13.287 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:154 Resp: 1051599
Ion Ratio Lower Upper
154 100
153 40.6 31.5 47.3
76 12.7 10.5 15.7





#44
2-Chloronaphthalene
Concen: 32.423 ng/ul
RT: 13.328 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion:162 Resp: 797908
Ion Ratio Lower Upper
162 100
164 31.7 25.8 38.6
127 42.1 31.4 47.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#45
2-Nitroaniline
Concen: 35.906 ng/ul
RT: 13.528 min Scan# 1775
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion: 65 Resp: 274903
Ion Ratio Lower Upper
65 100
92 68.4 54.1 81.1
138 103.1 77.0 115.4





#47
Dimethylphthalate
Concen: 32.636 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:163 Resp: 104726
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.1 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#48
2,6-Dinitrotoluene
Concen: 36.181 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:165 Resp: 227001
Ion Ratio Lower Upper
165 100
89 60.3 50.6 75.8
121 21.2 17.9 26.9





#50
Acenaphthylene
Concen: 32.906 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:152 Resp: 1346750
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.0 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#51
3-Nitroaniline
Concen: 36.667 ng/ul
RT: 14.351 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:138 Resp: 243204
Ion Ratio Lower Upper
138 100
108 11.4 8.6 12.8
92 122.2 92.9 139.3





#52
Acenaphthene
Concen: 32.375 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:153 Resp: 865509
Ion Ratio Lower Upper
153 100
152 47.1 38.4 57.6
154 86.5 69.7 104.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#53
2,4-Dinitrophenol
Concen: 31.293 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:184 Resp: 125701
Ion Ratio Lower Upper
184 100
63 75.4 55.9 83.9
154 66.3 53.1 79.7





#55
4-Nitrophenol
Concen: 34.008 ng/ul
RT: 14.663 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:109 Resp: 185420
Ion Ratio Lower Upper
109 100
139 125.0 100.1 150.1
65 124.4 104.6 157.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#56
Dibenzofuran
Concen: 32.061 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:168 Resp: 1207545
Ion Ratio Lower Upper
168 100
139 39.5 31.1 46.7





#57
2,4-Dinitrotoluene
Concen: 35.823 ng/ul
RT: 14.810 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:165 Resp: 332783
Ion Ratio Lower Upper
165 100
63 48.0 36.3 54.5
182 2.7 2.3 3.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 33.366 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:232 Resp: 223483
Ion Ratio Lower Upper
232 100
131 57.4 38.5 57.7
130 3.0 2.2 3.2
166 36.9 25.8 38.8





#59
Diethylphthalate
Concen: 33.324 ng/ul
RT: 15.275 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:149 Resp: 1100944
Ion Ratio Lower Upper
149 100
177 20.4 17.8 26.8
150 12.4 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#61
Fluorene
Concen: 31.903 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:166 Resp: 935009
Ion Ratio Lower Upper
166 100
165 98.9 77.7 116.5
167 13.3 11.1 16.7





#62
4-Chlorophenyl-phenylether
Concen: 31.847 ng/ul
RT: 15.493 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:204 Resp: 429838
Ion Ratio Lower Upper
204 100
206 33.0 26.7 40.1
141 71.3 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#63
4-Nitroaniline
Concen: 40.414 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:138 Resp: 250482
Ion Ratio Lower Upper
138 100
92 52.1 41.0 61.6
108 86.0 65.3 97.9





#66
4,6-Dinitro-2-methylphenol
Concen: 34.356 ng/ul
RT: 15.581 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:198 Resp: 180978
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.4
77 25.7 20.1 30.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#67
N-Nitrosodiphenylamine
Concen: 33.444 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:169 Resp: 859978
Ion Ratio Lower Upper
169 100
168 66.0 52.4 78.6
167 35.9 27.5 41.3





#68
4-Bromophenyl-phenylether
Concen: 33.919 ng/ul
RT: 16.387 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion:248 Resp: 26494
Ion Ratio Lower Upper
248 100
250 95.5 77.8 116.8
141 91.2 62.6 93.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#69
Hexachlorobenzene
Concen: 33.919 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07


Tgt Ion:284 Resp: 304553
Ion Ratio Lower Upper
284 100
142 45.6 27.8 41.8#
249 33.9 25.7 38.5





#70
Atrazine
Concen: 27.234 ng/ul
RT: 16.657 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:200 Resp: 244340
Ion Ratio Lower Upper
200 100
173 28.4 20.5 30.7
215 46.0 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#71
Pentachlorophenol
Concen: 32.297 ng/ul
RT: 16.845 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:266 Resp: 180613
Ion Ratio Lower Upper
266 100
264 64.8 50.2 75.2
268 62.1 49.3 73.9





#72
Phenanthrene
Concen: 33.498 ng/ul
RT: 17.234 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:178 Resp: 1585644
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 19.2 15.2 22.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#74
Anthracene
Concen: 33.101 ng/ul
RT: 17.328 min Scan# 2421
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:178 Resp: 1575511
Ion Ratio Lower Upper
178 100
179 15.2 13.0 19.4
176 18.6 15.5 23.3





#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.354 ng/uL
RT: 13.251 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:216 Resp: 388600
Ion Ratio Lower Upper
216 100
214 77.8 63.2 94.8
218 49.5 37.9 56.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#76
Pentachlorobenzene
Concen: 33.063 ng/uL
RT: 14.775 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:250 Resp: 350977
Ion Ratio Lower Upper
250 100
248 60.7 53.0 79.4
252 62.6 53.4 80.2





#77
 Carbazole
 Concen: 35.352 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BN020057.D
 Acq: 08 Jun 2022 09:07

Instrument :
 BNA_N
 ClientSampleId :
 SLCS379






Tgt Ion:167 Resp: 153075


Ion	Ratio	Lower	Upper
167	100		
166	21.5	16.1	24.1
139	13.6	10.2	15.2

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
 Supervised By :mohammad ahmed 06/09/2022






#78
 Di-n-butylphthalate
 Concen: 35.754 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BN020057.D
 Acq: 08 Jun 2022 09:07



Tgt Ion:149 Resp: 1919762


Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.1	10.7
104	5.4	4.1	6.1





#80
Fluoranthene
Concen: 33.828 ng/u1
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:202 Resp: 1827034
Ion Ratio Lower Upper
202 100
101 12.2 8.9 13.3
100 9.7 6.7 10.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#82
Pyrene
Concen: 33.393 ng/u1
RT: 19.616 min Scan# 2810
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:202 Resp: 1850297
Ion Ratio Lower Upper
202 100
101 13.7 9.7 14.5
100 11.7 8.0 12.0





#83
Butylbenzylphthalate
Concen: 36.104 ng/ul
RT: 20.522 min Scan# 2964
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:149 Resp: 900309
Ion Ratio Lower Upper
149 100
91 68.6 58.8 88.2
206 19.9 18.1 27.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#84
3,3'-Dichlorobenzidine
Concen: 36.723 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:252 Resp: 578568
Ion Ratio Lower Upper
252 100
254 64.0 50.8 76.2
126 13.8 8.2 12.4





#85
Benzo(a)anthracene
Concen: 33.246 ng/ul
RT: 21.369 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:228 Resp: 1731749
Ion Ratio Lower Upper
228 100
229 20.1 16.3 24.5
226 27.1 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 35.806 ng/ul
RT: 21.304 min Scan# 3097
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Tgt Ion:149 Resp: 1259798
Ion Ratio Lower Upper
149 100
167 27.1 22.5 33.7
279 3.4 3.5 5.3





#87
Chrysene
Concen: 33.117 ng/ul
RT: 21.422 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379



Tgt Ion:228 Resp: 1687611
Ion Ratio Lower Upper
228 100
226 29.2 24.2 36.2
229 18.8 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#89
Di-n-octyl phthalate
Concen: 35.809 ng/ul
RT: 22.210 min Scan# 3251
Delta R.T. 0.000 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07

Tgt Ion:149 Resp: 2324194





#90

Benzo(b)fluoranthene

Concen: 35.839 ng/u1

RT: 23.010 min Scan# 3386

Delta R.T. 0.006 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07

Instrument :

BNA_N

ClientSampleId :

SLCS379

Tgt Ion:252 Resp: 179369

Ion Ratio Lower Upper

252 100





253 20.6 16.5 24.7

125 11.7 7.8 11.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#91

Benzo(k)fluoranthene

Concen: 34.967 ng/u1

RT: 23.057 min Scan# 3395

Delta R.T. 0.006 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07




Tgt Ion:252 Resp: 1665954


Ion Ratio Lower Upper

252 100

253 21.8 18.2 27.4


125 11.4 8.2 12.4





#93
Benzo(a)pyrene
Concen: 34.838 ng/ul
RT: 23.616 min Scan# 3488
Delta R.T. 0.006 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07




Instrument :
BNA_N
ClientSampleId :
SLCS379




Tgt Ion:252 Resp: 168778
Ion Ratio Lower Upper
252 100
253 23.7 18.7 28.1
125 13.7 9.2 13.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 33.081 ng/ul
RT: 26.110 min Scan# 3914
Delta R.T. 0.012 min
Lab File: BN020057.D
Acq: 08 Jun 2022 09:07



Tgt Ion:276 Resp: 1648903
Ion Ratio Lower Upper
276 100
138 29.2 19.5 29.3
277 25.5 21.4 32.0





#95

Dibenzo(a,h)anthracene

Concen: 33.455 ng/ul

RT: 26.127 min Scan# 3915

Delta R.T. 0.006 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07

Instrument :

BNA_N

ClientSampleId :

SLCS379

Tgt Ion:278 Resp: 1424298

Ion Ratio Lower Upper

278 100





139 19.3 13.8 20.8

279 23.6 18.3 27.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 06/08/2022
Supervised By :mohammad ahmed 06/09/2022



#96

Benzo(g,h,i)perylene

Concen: 33.240 ng/ul

RT: 26.845 min Scan# 4039

Delta R.T. 0.000 min

Lab File: BN020057.D

Acq: 08 Jun 2022 09:07

Tgt Ion:276 Resp: 1399155

Ion Ratio Lower Upper

276 100

138 25.6 17.0 25.6#

277 22.4 17.8 26.6

