

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
 Data File : BN020249.D
 Acq On : 17 Jun 2022 21:40
 Operator : CG/JU
 Sample : PB145573BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK573

Quant Time: Jun 17 22:58:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 22:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	118917	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	618370	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	430465	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	898586	20.000	ng/u1	0.00
79) Chrysene-d12	21.345	240	799523	20.000	ng/u1	0.00
88) Perylene-d12	23.657	264	650570	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	20549	7.484	ng/uL	0.00
4) Pyridine-d5	3.658	84	296675	39.534	ng/u1	0.00
7) Phenol-d5	6.981	99	423371	43.025	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	250564	40.756	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	339272	42.434	ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	379085	44.529	ng/u1	0.00
21) Nitrobenzene-d5	8.946	128	175776	37.364	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	202851	40.630	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	347257	39.842	ng/u1	0.00
31) 4-Chloroaniline-d4	10.716	131	548270	38.845	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	1205039	38.473	ng/u1	0.00
49) Acenaphthylene-d8	14.110	160	1412220	38.879	ng/u1	0.00
54) 4-Nitrophenol-d4	14.657	143	205740	32.780	ng/u1	0.00
60) Fluorene-d10	15.416	176	1028095	39.667	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.534	200	192255	36.936	ng/u1	0.00
73) Anthracene-d10	17.263	188	1599649	40.790	ng/u1	0.00
81) Pyrene-d10	19.557	212	1797481	41.470	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.510	264	1340765	42.121	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
Data File : BN020249.D
Acq On : 17 Jun 2022 21:40
Operator : CG/JU
Sample : PB145573BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK573

Quant Time: Jun 17 22:58:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 22:53:31 2022
Response via : Initial Calibration

