

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
 Data File : BN020232.D
 Acq On : 17 Jun 2022 10:41
 Operator : CG/JU
 Sample : PB145580BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK580

Quant Time: Jun 17 22:56:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 22:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	139529	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	714693	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	492496	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	1063286	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	945420	20.000	ng/u1	0.00
88) Perylene-d12	23.663	264	773062	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.252	96	19887	6.173	ng/uL	0.00
4) Pyridine-d5	3.658	84	293401	33.322	ng/u1	0.00
7) Phenol-d5	6.981	99	410696	35.572	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	244257	33.861	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	335486	35.762	ng/u1	0.00
15) 4-Methylphenol-d8	8.510	113	372954	37.337	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	167767	30.855	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	188474	32.662	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.216	165	345195	34.267	ng/u1	0.00
31) 4-Chloroaniline-d4	10.716	131	537097	32.924	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	1201574	33.531	ng/u1	0.00
49) Acenaphthylene-d8	14.110	160	1383149	33.283	ng/u1	0.00
54) 4-Nitrophenol-d4	14.657	143	192027	26.742	ng/u1	0.00
60) Fluorene-d10	15.410	176	1005430	33.906	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.534	200	181608	29.486	ng/u1	0.00
73) Anthracene-d10	17.263	188	1595583	34.384	ng/u1	0.00
81) Pyrene-d10	19.557	212	1754087	34.224	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.515	264	1311155	34.664	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
Data File : BN020232.D
Acq On : 17 Jun 2022 10:41
Operator : CG/JU
Sample : PB145580BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK580

Quant Time: Jun 17 22:56:00 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 22:53:31 2022
Response via : Initial Calibration

