

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
 Data File : BN020255.D
 Acq On : 18 Jun 2022 10:18
 Operator : CG/JU
 Sample : PB145573BL
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK573

Quant Time: Jun 20 02:28:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 22:53:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	133329	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.569	136	647879	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.416	164	429600	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.157	188	898188	20.000	ng/u1	0.00
79) Chrysene-d12	21.351	240	816562	20.000	ng/u1	0.00
88) Perylene-d12	23.668	264	668586	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.246	96	21468	6.974	ng/uL	0.00
4) Pyridine-d5	3.652	84	300976	35.772	ng/u1	-0.01
7) Phenol-d5	6.969	99	394929	35.797	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.110	67	233227	33.836	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.316	132	312515	34.862	ng/u1	-0.01
15) 4-Methylphenol-d8	8.498	113	347146	36.369	ng/u1	-0.01
21) Nitrobenzene-d5	8.928	128	160119	32.486	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.651	143	176483	33.738	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.204	165	315163	34.513	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.704	131	498517	33.711	ng/u1	0.00
46) Dimethylphthalate-d6	13.827	166	1044977	33.430	ng/u1	0.00
49) Acenaphthylene-d8	14.104	160	1218308	33.608	ng/u1	0.00
54) 4-Nitrophenol-d4	14.651	143	173453	27.692	ng/u1	0.00
60) Fluorene-d10	15.404	176	876530	33.887	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.527	200	157136	30.202	ng/u1	0.00
73) Anthracene-d10	17.257	188	1372145	35.004	ng/u1	0.00
81) Pyrene-d10	19.557	212	1559095	35.220	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.521	264	1187698	36.307	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061722\
Data File : BN020255.D
Acq On : 18 Jun 2022 10:18
Operator : CG/JU
Sample : PB145573BL
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK573

Quant Time: Jun 20 02:28:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN060622.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 17 22:53:31 2022
Response via : Initial Calibration

