

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN062122\  
 Data File : BN020309.D  
 Acq On : 21 Jun 2022 16:02  
 Operator : CG/JU  
 Sample : N3313-01  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 YB7L1

Manual Integrations  
 APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022  
 Supervised By :mohammad ahmed 06/28/2022

Quant Time: Jun 22 00:36:49 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN061822.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 20 01:13:51 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.796  | 152  | 610      | 0.400 | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.579 | 136  | 2114     | 0.400 | ng/ul  | 0.00     |
| 9) Acenaphthene-d10         | 14.418 | 164  | 1580     | 0.400 | ng/ul  | # 0.00   |
| 13) Phenanthrene-d10        | 17.162 | 188  | 4074     | 0.400 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.353 | 240  | 6215     | 0.400 | ng/ul  | 0.00     |
| 23) Perylene-d12            | 23.659 | 264  | 9896     | 0.400 | ng/ul  | # 0.00   |
|                             |        |      |          |       |        |          |
| System Monitoring Compounds |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8           | 3.256  | 96   | 3205m    | 4.601 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.173 | 152  | 1360     | 0.389 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.192 | 212  | 5591     | 0.436 | ng/ul  | 0.00     |
|                             |        |      |          |       |        |          |
| Target Compounds            |        |      |          |       |        | Qvalue   |
| 2) 1,4-Dioxane              | 3.294  | 88   | 541      | 0.807 | ng/ul# | 86       |
| -----                       |        |      |          |       |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN062122\  
Data File : BN020309.D  
Acq On : 21 Jun 2022 16:02  
Operator : CG/JU  
Sample : N3313-01  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
YB7L1

Manual Integrations  
APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022  
Supervised By :mohammad ahmed 06/28/2022

Quant Time: Jun 22 00:36:49 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN061822.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 20 01:13:51 2022  
Response via : Initial Calibration

