

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062122\
 Data File : BN020327.D
 Acq On : 22 Jun 2022 03:28
 Operator : CG/JU
 Sample : N3359-16
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4T1

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022
 Supervised By :mohammad ahmed 06/28/2022

Quant Time: Jun 22 04:05:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 20 01:13:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	1987	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.617	136	9241858	0.400	ng/ul	# 0.03
9) Acenaphthene-d10	14.460	164	10170	0.400	ng/ul	0.04
13) Phenanthrene-d10	17.166	188	12211m	0.400	ng/ul	0.00
17) Chrysene-d12	21.352	240	13973	0.400	ng/ul	# 0.00
23) Perylene-d12	23.661	264	14802	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.256	96	13970	6.157	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.371	152	290125m	0.019	ng/ul	0.19
18) Fluoranthene-d10	19.192	212	15771	0.547	ng/ul	0.00
Target Compounds						
					Qvalue	
11) Acenaphthene	14.534	153	35826	0.770	ng/ul#	42
12) Fluorene	15.492	166	3877	0.080	ng/ul#	30
14) Pentachlorophenol	16.837	266	664	0.283	ng/ul	97
15) Phenanthrene	17.209	178	9530	0.216	ng/ul#	40
16) Anthracene	17.297	178	11202	0.300	ng/ul#	36
19) Fluoranthene	19.220	202	4391	0.078	ng/ul#	1
20) Pyrene	19.583	202	3813	0.062	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062122\
Data File : BN020327.D
Acq On : 22 Jun 2022 03:28
Operator : CG/JU
Sample : N3359-16
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4T1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 06/22/2022
Supervised By :mohammad ahmed 06/28/2022

Quant Time: Jun 22 04:05:37 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 20 01:13:51 2022
Response via : Initial Calibration

