

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062322\
 Data File : BN020376.D
 Acq On : 23 Jun 2022 16:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 24 04:07:51 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 15:57:15 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	30	0.00
2	1,4-Dioxane	0.400	0.511	-27.7#	38	0.00
3	n-Nitrosodimethylamine	0.400	0.770	-92.5#	86	0.00
4 S	2-Fluorophenol	0.400	0.475	-18.7	35	0.00
5 S	Phenol-d6	0.400	0.427	-6.7	33	0.00
6	bis(2-Chloroethyl)ether	0.400	0.433	-8.2	34	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	30	0.00
8 S	Nitrobenzene-d5	0.400	0.377	5.8	32	0.00
9	Naphthalene	0.400	0.391	2.3	31	0.00
10	Hexachlorobutadiene	0.400	0.374	6.5	30	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	32	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	29	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	32	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.551	-37.8#	47	0.00
15 S	2-Fluorobiphenyl	0.400	0.347	13.3	30	0.00
16	Acenaphthylene	0.400	0.377	5.8	33	0.00
17	Acenaphthene	0.400	0.383	4.3	31	0.00
18	Fluorene	0.400	0.396	1.0	33	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	39	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.531	-32.8#	63	0.01
21	4-Bromophenyl-phenylether	0.400	0.325	18.8	34	0.00
22	Hexachlorobenzene	0.400	0.324	19.0	34	0.00
23	Atrazine	0.400	0.404	-1.0	42	-0.01
24	Pentachlorophenol	0.400	0.669	-67.3#	73	-0.01
25	Phenanthrene	0.400	0.364	9.0	38	0.01
26	Anthracene	0.400	0.363	9.3	38	0.00
27 SURR	Fluoranthene-d10	0.400	0.465	-16.3	47	0.00
28	Fluoranthene	0.400	0.451	-12.7	46	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	65	0.00
30	Pyrene	0.400	0.282	29.5#	49	0.00
31 S	Terphenyl-d14	0.400	0.288	28.0#	49	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	64	0.00
33	Chrysene	0.400	0.389	2.8	65	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.314	21.5	57	0.00
35 I	Perylene-d12	0.400	0.400	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	92	0.00
37	Benzo(b)fluoranthene	0.400	0.346	13.5	78	0.00
38	Benzo(k)fluoranthene	0.400	0.349	12.8	77	0.00
39 C	Benzo(a)pyrene	0.400	0.374	6.5	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.416	-4.0	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4