

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020476.D
 Acq On : 26 Jun 2022 21:54
 Operator : CG/JU
 Sample : N3354-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-49

Quant Time: Jun 27 02:52:12 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	3107	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	10525	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	7158	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	13442	0.400	ng	# 0.00
29) Chrysene-d12	21.306	240	11561	0.400	ng	0.00
35) Perylene-d12	23.586	264	9216	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1322	0.153	ng	0.00
5) Phenol-d6	6.951	99	1133	0.110	ng	0.00
8) Nitrobenzene-d5	8.896	82	2436	0.286	ng	0.00
11) 2-Methylnaphthalene-d10	12.117	152	6588	0.365	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	1113	0.413	ng	0.00
15) 2-Fluorobiphenyl	12.998	172	8242	0.279	ng	0.00
27) Fluoranthene-d10	19.143	212	14593	0.365	ng	0.00
31) Terphenyl-d14	19.746	244	12257	0.444	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.225	149	3029	0.130	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
Data File : BN020476.D
Acq On : 26 Jun 2022 21:54
Operator : CG/JU
Sample : N3354-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-49

Quant Time: Jun 27 02:52:12 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

