

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020485.D
 Acq On : 27 Jun 2022 05:45
 Operator : CG/JU
 Sample : N3354-12
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-56

Quant Time: Jun 27 07:50:13 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	3369	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	10823	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	7187	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	12514	0.400	ng	0.00
29) Chrysene-d12	21.306	240	9707	0.400	ng	# 0.00
35) Perylene-d12	23.586	264	7807	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1515	0.162	ng	0.00
5) Phenol-d6	6.952	99	1097	0.098	ng	0.00
8) Nitrobenzene-d5	8.897	82	2622	0.299	ng	0.00
11) 2-Methylnaphthalene-d10	12.117	152	6549	0.353	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	823	0.304	ng	0.00
15) 2-Fluorobiphenyl	12.998	172	8718	0.294	ng	0.00
27) Fluoranthene-d10	19.143	212	13118	0.353	ng	0.00
31) Terphenyl-d14	19.746	244	11820	0.510	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.584	128	1646	0.051	ng	# 7
12) 2-Methylnaphthalene	12.193	142	514	0.024	ng	# 56
16) Acenaphthylene	14.089	152	2567	0.074	ng	# 19
17) Acenaphthene	14.431	154	15806	0.673	ng	96
24) Pentachlorophenol	16.785	266	337	0.314	ng	# 72
26) Anthracene	17.246	178	3614	0.099	ng	# 80
34) Bis(2-ethylhexyl)phtha...	21.225	149	1583	0.081	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
Data File : BN020485.D
Acq On : 27 Jun 2022 05:45
Operator : CG/JU
Sample : N3354-12
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-56

Quant Time: Jun 27 07:50:13 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

