

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020494.D
 Acq On : 27 Jun 2022 11:10
 Operator : CG/JU
 Sample : N3354-23
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP061422

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 06/28/2022
 Supervised By : Jagrut Upadhyay 06/30/2022

Quant Time: Jun 28 00:33:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	3009	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	9909	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	7043	0.400	ng	# 0.00
19) Phenanthrene-d10	17.113	188	12696	0.400	ng	0.00
29) Chrysene-d12	21.297	240	11526	0.400	ng	0.00
35) Perylene-d12	23.583	264	10387	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1286	0.154	ng	0.00
5) Phenol-d6	6.952	99	1021	0.102	ng	0.00
8) Nitrobenzene-d5	8.886	82	1844	0.230	ng	-0.01
11) 2-Methylnaphthalene-d10	12.117	152	4860	0.286	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	779	0.294	ng	0.00
15) 2-Fluorobiphenyl	12.998	172	6728	0.231	ng	0.00
27) Fluoranthene-d10	19.143	212	12474	0.331	ng	0.00
31) Terphenyl-d14	19.746	244	9318	0.338	ng	0.00
Target Compounds						
12) 2-Methylnaphthalene	12.189	142	1159	0.059	ng	# 79
17) Acenaphthene	14.431	154	17981	0.781	ng	97
18) Fluorene	15.415	166	5953	0.198	ng	# 95
26) Anthracene	17.236	178	8096	0.219	ng	# 89
28) Fluoranthene	19.173	202	29158	0.624	ng	100
30) Pyrene	19.535	202	19599	0.402	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.225	149	1521	0.066	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
Data File : BN020494.D
Acq On : 27 Jun 2022 11:10
Operator : CG/JU
Sample : N3354-23
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP061422

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 06/28/2022
Supervised By :Jagrut Upadhyay 06/30/2022

Quant Time: Jun 28 00:33:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration

