

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020470.D
 Acq On : 26 Jun 2022 17:39
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622

Quant Time: Jun 27 02:10:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.524	0.506	3.4	112	0.00
3	n-Nitrosodimethylamine	0.478	0.427	10.7	106	0.00
4 S	2-Fluorophenol	1.110	1.082	2.5	111	0.00
5 S	Phenol-d6	1.327	1.259	5.1	111	0.00
6	bis(2-Chloroethyl)ether	1.142	1.122	1.8	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.01
8 S	Nitrobenzene-d5	0.324	0.296	8.6	114	0.00
9	Naphthalene	1.195	1.161	2.8	115	0.00
10	Hexachlorobutadiene	0.218	0.210	3.7	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.686	0.660	3.8	116	0.00
12	2-Methylnaphthalene	0.794	0.764	3.8	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.151	0.128	15.2	112	0.00
15 S	2-Fluorobiphenyl	1.651	1.619	1.9	114	0.00
16	Acenaphthylene	1.923	1.777	7.6	116	0.00
17	Acenaphthene	1.308	1.258	3.8	114	0.00
18	Fluorene	1.707	1.615	5.4	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4,6-Dinitro-2-methylphenol	0.051	0.037	27.5#	115	0.00
21	4-Bromophenyl-phenylether	0.233	0.215	7.7	118	0.00
22	Hexachlorobenzene	0.257	0.244	5.1	120	0.00
23	Atrazine	0.183	0.158	13.7	113	0.00
24	Pentachlorophenol	0.061	0.044	27.9#	115	0.00
25	Phenanthrene	1.349	1.261	6.5	118	0.00
26	Anthracene	1.163	1.042	10.4	116	0.00
27 SURR	Fluoranthene-d10	1.188	1.097	7.7	118	0.00
28	Fluoranthene	1.471	1.368	7.0	119	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
30	Pyrene	1.694	1.639	3.2	119	0.00
31 S	Terphenyl-d14	0.956	0.928	2.9	119	0.00
32	Benzo(a)anthracene	1.429	1.321	7.6	118	0.00
33	Chrysene	1.582	1.510	4.6	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.805	0.687	14.7	116	0.00
35 I	Perylene-d12	1.000	1.000	0.0	118	0.00
36	Indeno(1,2,3-cd)pyrene	1.782	1.603	10.0	111	0.00
37	Benzo(b)fluoranthene	1.541	1.465	4.9	117	0.00
38	Benzo(k)fluoranthene	1.641	1.539	6.2	118	0.00
39 C	Benzo(a)pyrene	1.337	1.308	2.2	121	0.00
40	Dibenzo(a,h)anthracene	1.429	1.291	9.7	113	0.00
41	Benzo(g,h,i)perylene	1.561	1.387	11.1	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN062622