

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020470.D
 Acq On : 26 Jun 2022 17:39
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622

Quant Time: Jun 27 02:10:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.386	3.5	112	0.00
3	n-Nitrosodimethylamine	0.400	0.357	10.8	106	0.00
4 S	2-Fluorophenol	0.400	0.390	2.5	111	0.00
5 S	Phenol-d6	0.400	0.380	5.0	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	114	0.01
8 S	Nitrobenzene-d5	0.400	0.366	8.5	114	0.00
9	Naphthalene	0.400	0.389	2.8	115	0.00
10	Hexachlorobutadiene	0.400	0.386	3.5	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	116	0.00
12	2-Methylnaphthalene	0.400	0.385	3.8	114	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.400	0.339	15.3	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	114	0.00
16	Acenaphthylene	0.400	0.370	7.5	116	0.00
17	Acenaphthene	0.400	0.385	3.8	114	0.00
18	Fluorene	0.400	0.379	5.3	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	118	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.415	-3.7	115	0.00
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	118	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	120	0.00
23	Atrazine	0.400	0.345	13.8	113	0.00
24	Pentachlorophenol	0.400	0.396	1.0	115	0.00
25	Phenanthrene	0.400	0.374	6.5	118	0.00
26	Anthracene	0.400	0.358	10.5	116	0.00
27 SURR	Fluoranthene-d10	0.400	0.369	7.8	118	0.00
28	Fluoranthene	0.400	0.372	7.0	119	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	120	0.00
30	Pyrene	0.400	0.387	3.3	119	0.00
31 S	Terphenyl-d14	0.400	0.388	3.0	119	0.00
32	Benzo(a)anthracene	0.400	0.370	7.5	118	0.00
33	Chrysene	0.400	0.382	4.5	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.341	14.8	116	0.00
35 I	Perylene-d12	0.400	0.400	0.0	118	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.360	10.0	111	0.00
37	Benzo(b)fluoranthene	0.400	0.380	5.0	117	0.00
38	Benzo(k)fluoranthene	0.400	0.375	6.3	118	0.00
39 C	Benzo(a)pyrene	0.400	0.391	2.3	121	0.00
40	Dibenzo(a,h)anthracene	0.400	0.361	9.8	113	0.00
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN062622