

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020479.D
 Acq On : 26 Jun 2022 23:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 02:53:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.370	7.5	119	0.00
3	n-Nitrosodimethylamine	0.400	0.371	7.3	122	0.00
4 S	2-Fluorophenol	0.400	0.408	-2.0	129	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	135	0.00
6	bis(2-Chloroethyl)ether	0.400	0.402	-0.5	131	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	134	0.00
9	Naphthalene	0.400	0.390	2.5	130	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	131	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	130	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	132	0.00
14 S	2,4,6-Tribromophenol	0.400	0.400	0.0	153	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	128	0.00
16	Acenaphthylene	0.400	0.387	3.3	140	0.00
17	Acenaphthene	0.400	0.381	4.8	131	0.00
18	Fluorene	0.400	0.385	3.8	134	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	131	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.497	-24.2	184	0.00
21	4-Bromophenyl-phenylether	0.400	0.383	4.3	136	-0.01
22	Hexachlorobenzene	0.400	0.384	4.0	136	0.00
23	Atrazine	0.400	0.373	6.8	136	0.00
24	Pentachlorophenol	0.400	0.526	-31.5#	206	0.00
25	Phenanthrene	0.400	0.375	6.3	132	0.00
26	Anthracene	0.400	0.375	6.3	136	0.00
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	137	0.00
28	Fluoranthene	0.400	0.385	3.8	137	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	139	0.00
30	Pyrene	0.400	0.387	3.3	138	0.00
31 S	Terphenyl-d14	0.400	0.381	4.8	135	0.00
32	Benzo(a)anthracene	0.400	0.393	1.8	145	0.00
33	Chrysene	0.400	0.378	5.5	136	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.430	-7.5	168	0.00
35 I	Perylene-d12	0.400	0.400	0.0	128	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.339	15.3	114	-0.02
37	Benzo(b)fluoranthene	0.400	0.396	1.0	133	-0.01
38	Benzo(k)fluoranthene	0.400	0.389	2.8	133	-0.01
39 C	Benzo(a)pyrene	0.400	0.388	3.0	130	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.325	18.8	110	-0.01
41	Benzo(g,h,i)perylene	0.400	0.326	18.5	108	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4