

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020463.D
 Acq On : 26 Jun 2022 13:05
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jun 27 01:37:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 01:36:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	2596	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	7929	0.400	ng	0.01
13) Acenaphthene-d10	14.367	164	4702	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	9869	0.400	ng	# 0.00
29) Chrysene-d12	21.306	240	8823	0.400	ng	0.00
35) Perylene-d12	23.597	264	8556	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	845	0.111	ng	0.00
5) Phenol-d6	6.952	99	907	0.093	ng	0.00
8) Nitrobenzene-d5	8.897	82	615	0.111	ng	0.00
11) 2-Methylnaphthalene-d10	12.121	152	1374	0.093	ng	0.00
14) 2,4,6-Tribromophenol	15.872	330	188	0.084	ng	0.01
15) 2-Fluorobiphenyl	12.998	172	1988	0.118	ng	0.00
27) Fluoranthene-d10	19.152	212	3043	0.088	ng	0.00
31) Terphenyl-d14	19.750	244	2251	0.128	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.232	88	398	0.134	ng	97
3) n-Nitrosodimethylamine	3.572	42	289	0.107	ng	# 79
6) bis(2-Chloroethyl)ether	7.175	93	738	0.097	ng	99
9) Naphthalene	10.573	128	2411	0.102	ng	# 90
10) Hexachlorobutadiene	10.872	225	461	0.116	ng	# 98
12) 2-Methylnaphthalene	12.197	142	1589	0.092	ng	97
16) Acenaphthylene	14.089	152	2153	0.098	ng	100
17) Acenaphthene	14.431	154	1572	0.102	ng	96
18) Fluorene	15.425	166	2104	0.095	ng	100
21) 4-Bromophenyl-phenylether	16.313	248	587	0.108	ng	94
22) Hexachlorobenzene	16.425	284	671	0.109	ng	96
23) Atrazine	16.579	200	471	0.116	ng	# 85
24) Pentachlorophenol	16.784	266	135	0.050	ng	97
25) Phenanthrene	17.154	178	3518	0.104	ng	98
26) Anthracene	17.246	178	2806	0.096	ng	99
28) Fluoranthene	19.181	202	3709	0.083	ng	98
30) Pyrene	19.544	202	4001	0.123	ng	99
32) Benzo(a)anthracene	21.297	228	3216	0.102	ng	96
33) Chrysene	21.351	228	3617	0.100	ng	94
34) Bis(2-ethylhexyl)phtha...	21.234	149	2296	0.158	ng	# 98
36) Indeno(1,2,3-cd)pyrene	25.939	276	3648	0.087	ng	95
37) Benzo(b)fluoranthene	22.910	252	3209	0.095	ng	# 68
38) Benzo(k)fluoranthene	22.951	252	3368	0.097	ng	# 70
39) Benzo(a)pyrene	23.495	252	2789	0.095	ng	# 55
40) Dibenzo(a,h)anthracene	25.951	278	2864	0.086	ng	# 80
41) Benzo(g,h,i)perylene	26.656	276	3264	0.088	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
Data File : BN020463.D
Acq On : 26 Jun 2022 13:05
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Quant Time: Jun 27 01:37:18 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 01:36:24 2022
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.739 min Scan# 64
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05


Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1





Tgt Ion:152 Resp: 2596
 Ion Ratio Lower Upper
 152 100
 150 155.6 127.9 191.9
 115 67.0 51.4 77.2




#2
 1,4-Dioxane
 Concen: 0.134 ng
 RT: 3.232 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05



Tgt Ion: 88 Resp: 398
 Ion Ratio Lower Upper
 88 100
 43 39.2 30.2 45.2
 58 88.2 73.0 109.6









#3
 n-Nitrosodimethylamine
 Concen: 0.107 ng
 RT: 3.572 min Scan# 61
 Delta R.T. 0.014 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion: 42 Resp: 289
 Ion Ratio Lower Upper
 42 100
 74 124.9 84.0 126.0
 44 0.0 9.8 14.8#



#4
 2-Fluorophenol
 Concen: 0.111 ng
 RT: 5.348 min Scan# 314
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05

Tgt Ion: 112 Resp: 845
 Ion Ratio Lower Upper
 112 100
 64 61.2 49.1 73.7
 63 33.3 26.3 39.5









#5
Phenol-d6
Concen: 0.093 ng
RT: 6.952 min Scan# 51
Delta R.T. 0.007 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 907
Ion Ratio Lower Upper
99 100
42 23.6 19.3 28.9
71 34.2 27.5 41.3



#6
bis(2-Chloroethyl)ether
Concen: 0.097 ng
RT: 7.175 min Scan# 567
Delta R.T. 0.007 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion: 93 Resp: 738
Ion Ratio Lower Upper
93 100
63 84.6 67.4 101.0
95 35.2 26.5 39.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.530 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	7929
Ion Ratio	Lower	Upper	
136	100		
137	12.0	9.0	13.6
54	8.6	6.2	9.2
68	5.9	4.4	6.6



#8
Nitrobenzene-d5
Concen: 0.111 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:	82	Resp:	615
Ion Ratio	Lower	Upper	
82	100		
128	48.2	35.0	52.6
54	70.1	46.2	69.4









#9
Naphthalene
Concen: 0.102 ng
RT: 10.573 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 2411
Ion Ratio Lower Upper
128 100
129 15.1 8.9 13.3#
127 17.3 10.8 16.2#



#10
Hexachlorobutadiene
Concen: 0.116 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:225 Resp: 461
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 65.5 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.093 ng
RT: 12.121 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 1374
Ion Ratio Lower Upper
152 100
151 19.1 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.092 ng
RT: 12.197 min Scan# 1205
Delta R.T. 0.004 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:142 Resp: 1589
Ion Ratio Lower Upper
142 100
141 88.9 71.6 107.4
115 43.3 29.8 44.8





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.367 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN020463.D

Acq: 26 Jun 2022 13:05

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:164 Resp: 4702

Ion Ratio Lower Upper

164 100

162 103.4 82.1 123.1

160 48.4 38.6 58.0



#14

2,4,6-Tribromophenol

Concen: 0.084 ng

RT: 15.872 min Scan# 1602

Delta R.T. 0.010 min

Lab File: BN020463.D

Acq: 26 Jun 2022 13:05




Tgt Ion:330 Resp: 188


Ion Ratio Lower Upper

330 100

332 322.9 77.6 116.4#

141 50.5 41.0 61.4









#15
2-Fluorobiphenyl
Concen: 0.118 ng
RT: 12.998 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	172	Resp:	1988
Ion Ratio	Lower	Upper	
172	100		
171	37.5	27.2	40.8
170	26.0	18.2	27.4



#16
Acenaphthylene
Concen: 0.098 ng
RT: 14.089 min Scan# 1435
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:	152	Resp:	2153
Ion Ratio	Lower	Upper	
152	100		
151	20.3	16.2	24.4
153	13.3	10.7	16.1









#17
Acenaphthene
Concen: 0.102 ng
RT: 14.431 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 1572
Ion Ratio Lower Upper
154 100
153 114.8 87.5 131.3
152 56.3 43.8 65.6



#18
Fluorene
Concen: 0.095 ng
RT: 15.425 min Scan# 1560
Delta R.T. 0.011 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:166 Resp: 2104
Ion Ratio Lower Upper
166 100
165 98.1 78.6 118.0
167 13.1 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.113 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 9869
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.0 7.4 11.2#



#21
4-Bromophenyl-phenylether
Concen: 0.108 ng
RT: 16.313 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:248 Resp: 587
Ion Ratio Lower Upper
248 100
250 99.8 75.3 112.9
141 79.0 67.3 100.9









#22
Hexachlorobenzene
Concen: 0.109 ng
RT: 16.425 min Scan# 1656
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:284 Resp: 671
Ion Ratio Lower Upper
284 100
142 45.0 35.2 52.8
249 37.3 26.6 40.0



#23
Atrazine
Concen: 0.116 ng
RT: 16.579 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:200 Resp: 471
Ion Ratio Lower Upper
200 100
173 41.7 23.6 35.4#
215 52.8 36.6 54.8









#24
 Pentachlorophenol
 Concen: 0.050 ng
 RT: 16.784 min Scan# 1691
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	266	Resp:	135
Ion	Ratio	Lower	Upper
266	100		
264	62.2	51.3	76.9
268	68.9	52.6	79.0



#25
 Phenanthrene
 Concen: 0.104 ng
 RT: 17.154 min Scan# 1727
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05

Tgt Ion:	178	Resp:	3518
Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.3	22.9
179	16.5	12.2	18.4









#26
 Anthracene
 Concen: 0.096 ng
 RT: 17.246 min Scan# 1736
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:178 Resp: 2806
 Ion Ratio Lower Upper
 178 100
 176 19.1 14.9 22.3
 179 14.7 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.088 ng
 RT: 19.152 min Scan# 2082
 Delta R.T. 0.004 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05

Tgt Ion:212 Resp: 3043
 Ion Ratio Lower Upper
 212 100
 106 16.4 12.6 19.0
 104 10.0 7.2 10.8









#28
Fluoranthene
Concen: 0.083 ng
RT: 19.181 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:202 Resp: 3709
Ion Ratio Lower Upper
202 100
101 13.1 9.6 14.4
203 17.2 13.5 20.3



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.306 min Scan# 2430
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:240 Resp: 8823
Ion Ratio Lower Upper
240 100
120 19.6 14.6 21.8
236 28.3 22.6 33.8









#30
Pyrene
Concen: 0.123 ng
RT: 19.544 min Scan# 2175
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:202 Resp: 4001
Ion Ratio Lower Upper
202 100
200 20.7 16.6 25.0
203 18.4 14.2 21.2



#31
Terphenyl-d14
Concen: 0.128 ng
RT: 19.750 min Scan# 2224
Delta R.T. -0.004 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:244 Resp: 2251
Ion Ratio Lower Upper
244 100
212 12.0 7.1 10.7#
122 0.0 0.0 0.0









#32
Benzo(a)anthracene
Concen: 0.102 ng
RT: 21.297 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:228 Resp: 3216
Ion Ratio Lower Upper
228 100
226 28.3 21.2 31.8
229 21.6 15.6 23.4



#33
Chrysene
Concen: 0.100 ng
RT: 21.351 min Scan# 2435
Delta R.T. 0.009 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:228 Resp: 3617
Ion Ratio Lower Upper
228 100
226 32.1 23.4 35.0
229 22.3 15.7 23.5









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.158 ng
 RT: 21.234 min Scan# 2422
 Delta R.T. -0.000 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:149 Resp: 2296
 Ion Ratio Lower Upper
 149 100
 167 26.1 21.7 32.5
 279 4.4 2.2 3.2#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.597 min Scan# 2960
 Delta R.T. -0.003 min
 Lab File: BN020463.D
 Acq: 26 Jun 2022 13:05

Tgt Ion:264 Resp: 8556
 Ion Ratio Lower Upper
 264 100
 260 26.4 20.9 31.3
 265 45.2 28.0 42.0#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.087 ng
RT: 25.939 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:276 Resp: 3648
Ion Ratio Lower Upper
276 100
138 27.3 25.9 38.9
277 24.7 19.8 29.8



#37
Benzo(b)fluoranthene
Concen: 0.095 ng
RT: 22.910 min Scan# 2725
Delta R.T. 0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:252 Resp: 3209
Ion Ratio Lower Upper
252 100
253 36.8 18.5 27.7#
125 30.0 11.4 17.0#









#38
Benzo(k)fluoranthene
Concen: 0.097 ng
RT: 22.951 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 3368
Ion Ratio Lower Upper
252 100
253 37.7 18.0 27.0#
125 25.0 11.0 16.4#



#39
Benzo(a)pyrene
Concen: 0.095 ng
RT: 23.495 min Scan# 2925
Delta R.T. -0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:252 Resp: 2789
Ion Ratio Lower Upper
252 100
253 43.3 19.0 28.4#
125 39.6 13.8 20.8#









#40
Dibenzo(a,h)anthracene
Concen: 0.086 ng
RT: 25.951 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion:278 Resp: 2864
Ion Ratio Lower Upper
278 100
139 29.2 18.6 28.0#
279 38.5 19.7 29.5#



#41
Benzo(g,h,i)perylene
Concen: 0.088 ng
RT: 26.656 min Scan# 4006
Delta R.T. 0.003 min
Lab File: BN020463.D
Acq: 26 Jun 2022 13:05

Tgt Ion:276 Resp: 3264
Ion Ratio Lower Upper
276 100
277 29.7 19.5 29.3#
138 33.1 22.6 34.0

