

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
 Data File : BN020470.D
 Acq On : 26 Jun 2022 17:39
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622

Quant Time: Jun 27 02:10:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	3411	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	10587	0.400	ng	0.01
13) Acenaphthene-d10	14.367	164	6216	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	13343	0.400	ng	0.00
29) Chrysene-d12	21.306	240	11607	0.400	ng	0.00
35) Perylene-d12	23.595	264	10663	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	3690	0.390	ng	0.00
5) Phenol-d6	6.944	99	4293	0.380	ng	0.00
8) Nitrobenzene-d5	8.897	82	3138	0.366	ng	0.00
11) 2-Methylnaphthalene-d10	12.121	152	6988	0.385	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	794	0.339	ng	0.00
15) 2-Fluorobiphenyl	12.998	172	10062	0.392	ng	0.00
27) Fluoranthene-d10	19.147	212	14631	0.369	ng	0.00
31) Terphenyl-d14	19.750	244	10772	0.388	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.232	88	1727	0.386	ng	98
3) n-Nitrosodimethylamine	3.564	42	1456	0.357	ng	# 90
6) bis(2-Chloroethyl)ether	7.168	93	3828	0.393	ng	99
9) Naphthalene	10.573	128	12290	0.389	ng	99
10) Hexachlorobutadiene	10.872	225	2225	0.386	ng	# 100
12) 2-Methylnaphthalene	12.197	142	8088	0.385	ng	99
16) Acenaphthylene	14.089	152	11047	0.370	ng	99
17) Acenaphthene	14.431	154	7820	0.385	ng	96
18) Fluorene	15.415	166	10039	0.379	ng	99
20) 4,6-Dinitro-2-methylph...	15.511	198	496	0.415	ng	# 73
21) 4-Bromophenyl-phenylether	16.313	248	2874	0.370	ng	# 87
22) Hexachlorobenzene	16.426	284	3253	0.379	ng	99
23) Atrazine	16.579	200	2113	0.345	ng	96
24) Pentachlorophenol	16.784	266	585	0.396	ng	96
25) Phenanthrene	17.154	178	16824	0.374	ng	100
26) Anthracene	17.246	178	13904	0.358	ng	100
28) Fluoranthene	19.181	202	18257	0.372	ng	99
30) Pyrene	19.544	202	19026	0.387	ng	100
32) Benzo(a)anthracene	21.297	228	15328	0.370	ng	98
33) Chrysene	21.351	228	17526	0.382	ng	100
34) Bis(2-ethylhexyl)phtha...	21.234	149	7975	0.341	ng	99
36) Indeno(1,2,3-cd)pyrene	25.933	276	17098	0.360	ng	99
37) Benzo(b)fluoranthene	22.905	252	15620	0.380	ng	96
38) Benzo(k)fluoranthene	22.951	252	16406	0.375	ng	97
39) Benzo(a)pyrene	23.495	252	13945	0.391	ng	# 90
40) Dibenzo(a,h)anthracene	25.948	278	13768	0.361	ng	98
41) Benzo(g,h,i)perylene	26.644	276	14787	0.355	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062622\
Data File : BN020470.D
Acq On : 26 Jun 2022 17:39
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN062622


Quant Time: Jun 27 02:10:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.739 min Scan# 645
 Delta R.T. -0.000 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622






Tgt Ion:152 Resp: 3411
 Ion Ratio Lower Upper
 152 100
 150 157.2 127.9 191.9
 115 67.6 51.4 77.2



#2
 1,4-Dioxane
 Concen: 0.386 ng
 RT: 3.232 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Tgt Ion: 88 Resp: 1727
 Ion Ratio Lower Upper
 88 100
 43 34.1 30.2 45.2
 58 90.6 73.0 109.6









#3
n-Nitrosodimethylamine
Concen: 0.357 ng
RT: 3.564 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion: 42 Resp: 1456
Ion Ratio Lower Upper
42 100
74 114.4 84.0 126.0
44 6.8 9.8 14.8#



#4
2-Fluorophenol
Concen: 0.390 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion: 112 Resp: 3690
Ion Ratio Lower Upper
112 100
64 61.3 49.1 73.7
63 32.6 26.3 39.5








#5
Phenol-d6
Concen: 0.380 ng
RT: 6.944 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion: 99 Resp: 4293
Ion Ratio Lower Upper
99 100
42 21.3 19.3 28.9
71 35.7 27.5 41.3



#6
bis(2-Chloroethyl)ether
Concen: 0.393 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion: 93 Resp: 3828
Ion Ratio Lower Upper
93 100
63 82.6 67.4 101.0
95 33.6 26.5 39.7








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.530 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:	136	Resp:	10587
Ion Ratio	Lower	Upper	
136	100		
137	11.8	9.0	13.6
54	8.0	6.2	9.2
68	5.5	4.4	6.6



#8
Nitrobenzene-d5
Concen: 0.366 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:	82	Resp:	3138
Ion Ratio	Lower	Upper	
82	100		
128	42.4	35.0	52.6
54	56.9	46.2	69.4









#9
Naphthalene
Concen: 0.389 ng
RT: 10.573 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:	128	Resp:	12290
Ion	Ratio	Lower	Upper
128	100		
129	11.4	8.9	13.3
127	13.7	10.8	16.2



#10
Hexachlorobutadiene
Concen: 0.386 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:	225	Resp:	2225
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.8	50.9	76.3








#11
2-Methylnaphthalene-d10
Concen: 0.385 ng
RT: 12.121 min Scan# 1185
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622


Tgt Ion:152 Resp: 6988
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.385 ng
RT: 12.197 min Scan# 1205
Delta R.T. 0.004 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39


Tgt Ion:142 Resp: 8088
Ion Ratio Lower Upper
142 100
141 87.9 71.6 107.4
115 37.3 29.8 44.8







#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.367 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Instrument :
BNA_N
ClientSampleId :
ICVBN062622






Tgt Ion:164 Resp: 6216
Ion Ratio Lower Upper
164 100
162 105.4 82.1 123.1
160 49.1 38.6 58.0



#14
2,4,6-Tribromophenol
Concen: 0.339 ng
RT: 15.861 min Scan# 1601
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:330 Resp: 794
Ion Ratio Lower Upper
330 100
332 155.3 77.6 116.4#
141 55.5 41.0 61.4








#15
2-Fluorobiphenyl
Concen: 0.392 ng
RT: 12.998 min Scan# 1333
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:	172	Resp:	10062
Ion Ratio	Lower	Upper	
172	100		
171	34.6	27.2	40.8
170	23.0	18.2	27.4



#16
Acenaphthylene
Concen: 0.370 ng
RT: 14.089 min Scan# 1435
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:	152	Resp:	11047
Ion Ratio	Lower	Upper	
152	100		
151	19.9	16.2	24.4
153	12.9	10.7	16.1








#17
Acenaphthene
Concen: 0.385 ng
RT: 14.431 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:154 Resp: 7820
Ion Ratio Lower Upper
154 100
153 114.5 87.5 131.3
152 55.9 43.8 65.6



#18
Fluorene
Concen: 0.379 ng
RT: 15.415 min Scan# 1559
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:166 Resp: 10039
Ion Ratio Lower Upper
166 100
165 99.6 78.6 118.0
167 13.4 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.113 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:188 Resp: 13343
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.1 7.4 11.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.415 ng
RT: 15.511 min Scan# 1568
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39


Tgt Ion:198 Resp: 496
Ion Ratio Lower Upper
198 100
182 26.2 12.1 18.1#
77 0.0 0.0 0.0








#21
4-Bromophenyl-phenylether
Concen: 0.370 ng
RT: 16.313 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Instrument :
BNA_N
ClientSampleId :
ICVBN062622







Tgt Ion:248 Resp: 2874
Ion Ratio Lower Upper
248 100
250 100.9 75.3 112.9
141 66.6 67.3 100.9#



#22
Hexachlorobenzene
Concen: 0.379 ng
RT: 16.426 min Scan# 1656
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:284 Resp: 3253
Ion Ratio Lower Upper
284 100
142 43.1 35.2 52.8
249 34.2 26.6 40.0





#23

Atrazine

Concen: 0.345 ng

RT: 16.579 min Scan# 1671

Delta R.T. -0.000 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39

Instrument :

BNA_N

ClientSampleId :

ICVBN062622





Tgt Ion:200 Resp: 2113

Ion Ratio Lower Upper

200 100

173 34.1 23.6 35.4

215 45.9 36.6 54.8



#24

Pentachlorophenol

Concen: 0.396 ng

RT: 16.784 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39




Tgt Ion:266 Resp: 585

Ion Ratio Lower Upper

266 100

264 63.1 51.3 76.9

268 60.0 52.6 79.0









#25
Phenanthrene
Concen: 0.374 ng
RT: 17.154 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:178 Resp: 16824
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.3 12.2 18.4



#26
Anthracene
Concen: 0.358 ng
RT: 17.246 min Scan# 1736
Delta R.T. 0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:178 Resp: 13904
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 15.3 12.2 18.4








#27
Fluoranthene-d10
Concen: 0.369 ng
RT: 19.147 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622


Tgt Ion:212 Resp: 14631
Ion Ratio Lower Upper
212 100
106 15.2 12.6 19.0
104 8.6 7.2 10.8



#28
Fluoranthene
Concen: 0.372 ng
RT: 19.181 min Scan# 2089
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39

Tgt Ion:202 Resp: 18257
Ion Ratio Lower Upper
202 100
101 12.4 9.6 14.4
203 16.7 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.306 min Scan# 2430

Delta R.T. -0.000 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39

Instrument :

BNA_N

ClientSampleId :

ICVBN062622





Tgt Ion:240 Resp: 11607

Ion Ratio Lower Upper

240 100

120 17.0 14.6 21.8

236 28.0 22.6 33.8



#30

Pyrene

Concen: 0.387 ng

RT: 19.544 min Scan# 2175

Delta R.T. -0.000 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39




Tgt Ion:202 Resp: 19026

Ion Ratio Lower Upper

202 100

200 20.6 16.6 25.0

203 17.7 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.388 ng
 RT: 19.750 min Scan# 21
 Delta R.T. -0.004 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622


Tgt Ion:244 Resp: 10772
 Ion Ratio Lower Upper
 244 100
 212 9.3 7.1 10.7
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 0.370 ng
 RT: 21.297 min Scan# 2429
 Delta R.T. 0.000 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Tgt Ion:228 Resp: 15328
 Ion Ratio Lower Upper
 228 100
 226 27.3 21.2 31.8
 229 20.3 15.6 23.4





#33

Chrysene

Concen: 0.382 ng

RT: 21.351 min Scan# 2435

Delta R.T. 0.009 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39

Instrument :

BNA_N

ClientSampleId :

ICVBN062622





Tgt Ion:228 Resp: 17526

Ion Ratio Lower Upper

228 100

226 29.2 23.4 35.0

229 19.7 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.341 ng

RT: 21.234 min Scan# 2422

Delta R.T. -0.000 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39




Tgt Ion:149 Resp: 7975

Ion Ratio Lower Upper

149 100

167 26.6 21.7 32.5


279 3.1 2.2 3.2







#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.595 min Scan# 2959
 Delta R.T. -0.006 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622



Tgt Ion:264 Resp: 10663




Ion	Ratio	Lower	Upper
264	100		
260	26.1	20.9	31.3
265	37.1	28.0	42.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.360 ng
 RT: 25.933 min Scan# 3759
 Delta R.T. -0.003 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Tgt Ion:276 Resp: 17098

Ion	Ratio	Lower	Upper
276	100		
138	31.7	25.9	38.9
277	25.2	19.8	29.8





#37

Benzo(b)fluoranthene

Concen: 0.380 ng

RT: 22.905 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39

Instrument :

BNA_N

ClientSampleId :

ICVBN062622





Tgt Ion:252 Resp: 15620

Ion Ratio Lower Upper

252 100

253 24.7 18.5 27.7

125 16.7 11.4 17.0



#38

Benzo(k)fluoranthene

Concen: 0.375 ng

RT: 22.951 min Scan# 2739

Delta R.T. -0.003 min

Lab File: BN020470.D

Acq: 26 Jun 2022 17:39




Tgt Ion:252 Resp: 16406

Ion Ratio Lower Upper

252 100

253 24.4 18.0 27.0

125 14.3 11.0 16.4








#39
Benzo(a)pyrene
Concen: 0.391 ng
RT: 23.495 min Scan# 2925
Delta R.T. -0.003 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39




Instrument :
BNA_N
ClientSampleId :
ICVBN062622

Tgt Ion:252 Resp: 13945
Ion Ratio Lower Upper
252 100
253 26.3 19.0 28.4
125 24.9 13.8 20.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.361 ng
RT: 25.948 min Scan# 3764
Delta R.T. -0.000 min
Lab File: BN020470.D
Acq: 26 Jun 2022 17:39


Tgt Ion:278 Resp: 13768
Ion Ratio Lower Upper
278 100
139 23.0 18.6 28.0
279 26.2 19.7 29.5





#41
 Benzo(g,h,i)perylene
 Concen: 0.355 ng
 RT: 26.644 min Scan# 40
 Delta R.T. -0.009 min
 Lab File: BN020470.D
 Acq: 26 Jun 2022 17:39

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062622



Tgt Ion:	276	Resp:	14787
Ion	Ratio	Lower	Upper
276	100		
277	24.8	19.5	29.3
138	29.2	22.6	34.0

