

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN063020\
 Data File : BN011080.D
 Acq On : 30 Jun 2020 14:29
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN063020

Quant Time: Jun 30 15:24:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 14:12:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.564	0.618	-9.6	116	0.00
3	n-Nitrosodimethylamine	0.285	0.305	-7.0	122	0.00
4 S	2-Fluorophenol	0.995	1.082	-8.7	115	-0.02
5 S	Phenol-d6	1.144	1.211	-5.9	110	0.00
6	bis(2-Chloroethyl)ether	1.039	1.067	-2.7	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
8 S	Nitrobenzene-d5	0.340	0.327	3.8	113	0.00
9	Naphthalene	1.194	1.204	-0.8	119	-0.01
10	Hexachlorobutadiene	0.303	0.290	4.3	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.692	0.625	9.7	101	0.00
12	2-Methylnaphthalene	0.799	0.738	7.6	107	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.168	0.164	2.4	106	-0.01
15 S	2-Fluorobiphenyl	1.884	1.944	-3.2	111	0.00
16	Acenaphthylene	2.006	2.001	0.2	109	0.00
17	Acenaphthene	1.359	1.365	-0.4	110	0.00
18	Fluorene	1.698	1.726	-1.6	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	-0.01
20	4,6-Dinitro-2-methylphenol	0.043	0.037	14.0	103	0.00
21	4-Bromophenyl-phenylether	0.266	0.259	2.6	105	0.00
22	Hexachlorobenzene	0.313	0.322	-2.9	111	0.00
23	Atrazine	0.188	0.188	0.0	106	0.00
24	Pentachlorophenol	0.084	0.088	-4.8	116	0.00
25	Phenanthrene	1.295	1.312	-1.3	109	0.00
26	Anthracene	1.072	1.071	0.1	107	-0.01
27 SURR	Fluoranthene-d10	1.210	1.215	-0.4	109	0.00
28	Fluoranthene	1.457	1.472	-1.0	110	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
30	Pyrene	1.492	1.570	-5.2	109	0.00
31 S	Terphenyl-d14	1.041	1.107	-6.3	109	0.00
32	Benzo(a)anthracene	1.337	1.326	0.8	97	0.00
33	Chrysene	1.534	1.599	-4.2	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.540	0.534	1.1	98	0.00
35	Indeno(1,2,3-cd)pyrene	1.604	1.876	-17.0	119	0.00
36 I	Perylene-d12	1.000	1.000	0.0	104	0.00
37	Benzo(b)fluoranthene	1.603	1.705	-6.4	107	0.00
38	Benzo(k)fluoranthene	1.716	1.785	-4.0	97	0.00
39 C	Benzo(a)pyrene	1.382	1.440	-4.2	108	0.00
40	Dibenzo(a,h)anthracene	1.416	1.576	-11.3	121	0.00
41	Benzo(g,h,i)perylene	1.560	1.697	-8.8	117	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN063020\
Data File : BN011080.D
Acq On : 30 Jun 2020 14:29
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN063020

Quant Time: Jun 30 15:24:38 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 14:12:13 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0