

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
 Data File : BN015265.D
 Acq On : 01 Jul 2021 18:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 18:57:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.731	152	17223	0.400	ng	# 0.00
7) Naphthalene-d8	10.492	136	74562	0.400	ng	# 0.00
13) Acenaphthene-d10	14.345	164	37160	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	68783	0.400	ng	0.00
29) Chrysene-d12	21.291	240	53594	0.400	ng	#-0.01
35) Perylene-d12	23.539	264	44324	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.343	112	21323	0.486	ng	-0.02
5) Phenol-d6	6.901	99	26579	0.481	ng	0.00
8) Nitrobenzene-d5	8.870	82	23917	0.573	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	47795	0.387	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	4255	0.485	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	61323	0.406	ng	0.00
27) Fluoranthene-d10	19.133	212	71277	0.351	ng	0.00
31) Terphenyl-d14	19.733	244	50665	0.404	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.351	88	3092	0.383	ng	# 54
3) n-Nitrosodimethylamine	3.692	42	6738	0.445	ng	# 67
6) bis(2-Chloroethyl)ether	7.169	93	22669	0.417	ng	# 83
9) Naphthalene	10.543	128	84813	0.397	ng	97
10) Hexachlorobutadiene	10.834	225	16880	0.419	ng	# 99
12) 2-Methylnaphthalene	12.159	142	54245	0.395	ng	# 90
16) Acenaphthylene	14.066	152	64765	0.393	ng	98
17) Acenaphthene	14.409	154	44462	0.385	ng	100
18) Fluorene	15.398	166	53628	0.382	ng	99
20) 4,6-Dinitro-2-methylph...	15.474	198	281	0.409	ng	# 81
21) 4-Bromophenyl-phenylether	16.288	248	17386	0.399	ng	# 87
22) Hexachlorobenzene	16.397	284	20845	0.421	ng	# 92
23) Atrazine	16.555	200	9678	0.430	ng	94
24) Pentachlorophenol	16.750	266	2616	0.550	ng	98
25) Phenanthrene	17.127	178	87047	0.393	ng	99
26) Anthracene	17.225	178	67156	0.354	ng	99
28) Fluoranthene	19.157	202	91647	0.385	ng	# 97
30) Pyrene	19.525	202	85516	0.419	ng	99
32) Benzo(a)anthracene	21.281	228	58968	0.321	ng	96
33) Chrysene	21.323	228	79138	0.395	ng	98
34) Bis(2-ethylhexyl)phtha...	21.217	149	18200	0.425	ng	# 85
36) Indeno(1,2,3-cd)pyrene	25.802	276	51638	0.325	ng	# 90
37) Benzo(b)fluoranthene	22.869	252	65026	0.374	ng	# 94
38) Benzo(k)fluoranthene	22.910	252	78267	0.422	ng	# 94
39) Benzo(a)pyrene	23.444	252	62880	0.396	ng	# 86
40) Dibenzo(a,h)anthracene	25.816	278	37300	0.299	ng	# 91
41) Benzo(g,h,i)perylene	26.487	276	44730	0.325	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
Data File : BN015265.D
Acq On : 01 Jul 2021 18:07
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jul 01 18:57:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.731 min Scan# 827
 Delta R.T. -0.001 min
 Lab File: BN015265.D
 Acq: 01 Jul 2021 18:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 152 Resp: 17223
 Ion Ratio Lower Upper
 152 100
 150 156.5 119.0 178.6
 115 61.3 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.383 ng
 RT: 3.351 min Scan# 352
 Delta R.T. -0.001 min
 Lab File: BN015265.D
 Acq: 01 Jul 2021 18:07

Tgt Ion: 88 Resp: 3092
 Ion Ratio Lower Upper
 88 100
 43 30.7 12.9 19.3#
 58 81.5 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.445 ng
RT: 3.692 min Scan# 389
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 6738
Ion Ratio Lower Upper
42 100
74 138.3 149.5 224.3#
44 6.9 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.486 ng
RT: 5.343 min Scan# 568
Delta R.T. -0.019 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion: 112 Resp: 21323
Ion Ratio Lower Upper
112 100
64 58.0 39.0 58.6
63 30.0 20.6 30.8









#5
Phenol-d6
Concen: 0.481 ng
RT: 6.901 min Scan# 737
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 26579
Ion Ratio Lower Upper
99 100
42 18.7 8.9 13.3#
71 31.9 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.417 ng
RT: 7.169 min Scan# 766
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion: 93 Resp: 22669
Ion Ratio Lower Upper
93 100
63 76.7 46.6 70.0#
95 32.9 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.492 min Scan# 1062
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:136	Resp:	74562
Ion Ratio	Lower	Upper
136	100	
137	11.1	9.8 14.8
54	9.1	5.4 8.2#
68	7.1	4.9 7.3



#8
Nitrobenzene-d5
Concen: 0.573 ng
RT: 8.870 min Scan# 934
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion: 82	Resp:	23917
Ion Ratio	Lower	Upper
82	100	
128	42.8	31.6 47.4
54	47.6	29.0 43.4#









#9
Naphthalene
Concen: 0.397 ng
RT: 10.543 min Scan# 1066
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 84813
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.1 11.8 17.6



#10
Hexachlorobutadiene
Concen: 0.419 ng
RT: 10.834 min Scan# 1089
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:225 Resp: 16880
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.387 ng
RT: 12.083 min Scan# 1270
Delta R.T. -0.005 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 47795
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.395 ng
RT: 12.159 min Scan# 1287
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:142 Resp: 54245
Ion Ratio Lower Upper
142 100
141 88.3 72.3 108.5
115 34.0 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.345 min Scan# 1510
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 37160
Ion Ratio Lower Upper
164 100
162 100.8 83.8 125.6
160 46.2 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.485 ng
RT: 15.842 min Scan# 1628
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:330 Resp: 4255
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.0 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.406 ng
RT: 12.962 min Scan# 1401
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 61323
Ion Ratio Lower Upper
172 100
171 35.0 30.6 45.8
170 23.8 20.2 30.4



#16
Acenaphthylene
Concen: 0.393 ng
RT: 14.066 min Scan# 1488
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:152 Resp: 64765
Ion Ratio Lower Upper
152 100
151 19.5 16.7 25.1
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.385 ng
RT: 14.409 min Scan# 1515
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 44462
Ion Ratio Lower Upper
154 100
153 113.6 90.9 136.3
152 57.4 45.8 68.8



#18
Fluorene
Concen: 0.382 ng
RT: 15.398 min Scan# 1593
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:166 Resp: 53628
Ion Ratio Lower Upper
166 100
165 98.3 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.091 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 68783
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.0 9.0 13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.409 ng
RT: 15.474 min Scan# 1599
Delta R.T. -0.014 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:198 Resp: 281
Ion Ratio Lower Upper
198 100
182 31.1 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.399 ng
RT: 16.288 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 17386
Ion Ratio Lower Upper
248 100
250 94.4 86.2 129.4
141 54.3 36.2 54.2#



#22
Hexachlorobenzene
Concen: 0.421 ng
RT: 16.397 min Scan# 1673
Delta R.T. -0.013 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:284 Resp: 20845
Ion Ratio Lower Upper
284 100
142 35.2 22.2 33.2#
249 31.3 26.6 39.8





#23

Atrazine

Concen: 0.430 ng

RT: 16.555 min Scan# 1686





Delta R.T. -0.001 min

Lab File: BN015265.D

Acq: 01 Jul 2021 18:07

Tgt Ion:	200	Resp:	9678
Ion Ratio	Lower	Upper	
200	100		
173	27.2	21.7	32.5
215	47.8	43.1	64.7

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



#24

Pentachlorophenol

Concen: 0.550 ng




RT: 16.750 min Scan# 1702


Delta R.T. -0.001 min

Lab File: BN015265.D

Acq: 01 Jul 2021 18:07

Tgt Ion:	266	Resp:	2616
Ion Ratio	Lower	Upper	
266	100		
264	62.0	51.0	76.4
268	64.9	50.6	76.0









#25
Phenanthrene
Concen: 0.393 ng
RT: 17.127 min Scan# 1733
Delta R.T. -0.001 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 87047
Ion Ratio Lower Upper
178 100
176 18.9 15.6 23.4
179 15.4 12.1 18.1



#26
Anthracene
Concen: 0.354 ng
RT: 17.225 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:178 Resp: 67156
Ion Ratio Lower Upper
178 100
176 18.3 15.1 22.7
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.351 ng
RT: 19.133 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 71277
Ion Ratio Lower Upper
212 100
106 14.1 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.385 ng
RT: 19.157 min Scan# 2036
Delta R.T. -0.006 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:202 Resp: 91647
Ion Ratio Lower Upper
202 100
101 11.6 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.291 min Scan# 2328
Delta R.T. -0.012 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 53594
Ion Ratio Lower Upper
240 100
120 10.6 4.2 6.4#
236 26.4 21.4 32.2



#30
Pyrene
Concen: 0.419 ng
RT: 19.525 min Scan# 2110
Delta R.T. -0.005 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:202 Resp: 85516
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.3 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.404 ng
 RT: 19.733 min Scan# 2152
 Delta R.T. -0.006 min
 Lab File: BN015265.D
 Acq: 01 Jul 2021 18:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 50665
 Ion Ratio Lower Upper
 244 100
 212 7.9 6.3 9.5
 122 13.7 6.2 9.2#



#32
 Benzo(a)anthracene
 Concen: 0.321 ng
 RT: 21.281 min Scan# 2327
 Delta R.T. -0.000 min
 Lab File: BN015265.D
 Acq: 01 Jul 2021 18:07

Tgt Ion:228 Resp: 58968
 Ion Ratio Lower Upper
 228 100
 226 25.8 23.0 34.6
 229 19.8 15.7 23.5









#33
Chrysene
Concen: 0.395 ng
RT: 21.323 min Scan# 2331
Delta R.T. -0.012 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228	Resp:	79138
Ion Ratio	Lower	Upper
228	100	
226	30.6	25.0 37.6
229	19.0	15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.425 ng
RT: 21.217 min Scan# 2321
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07


Tgt Ion:149	Resp:	18200
Ion Ratio	Lower	Upper
149	100	
167	26.6	28.9 43.3#
279	3.2	4.9 7.3#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.539 min Scan# 2764
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:264 Resp: 44324




Ion	Ratio	Lower	Upper
264	100		
260	25.4	20.6	31.0
265	24.1	21.4	32.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.325 ng
RT: 25.802 min Scan# 3425
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:276 Resp: 51638

Ion	Ratio	Lower	Upper
276	100		
138	23.1	10.0	15.0#
277	24.7	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.374 ng
RT: 22.869 min Scan# 2568
Delta R.T. -0.008 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 65026
Ion Ratio Lower Upper
252 100
253 22.6 18.5 27.7
125 14.8 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.422 ng
RT: 22.910 min Scan# 2580
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:252 Resp: 78267
Ion Ratio Lower Upper
252 100
253 22.2 19.1 28.7
125 11.1 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.396 ng
RT: 23.444 min Scan# 2736
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 62880
Ion Ratio Lower Upper
252 100
253 23.7 19.4 29.2
125 26.3 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.299 ng
RT: 25.816 min Scan# 3429
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Tgt Ion:278 Resp: 37300
Ion Ratio Lower Upper
278 100
139 18.9 9.0 13.4#
279 24.0 21.0 31.4





#41
Benzo(g,h,i)perylene
Concen: 0.325 ng
RT: 26.487 min Scan# 3625
Delta R.T. -0.011 min
Lab File: BN015265.D
Acq: 01 Jul 2021 18:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:276 Resp: 44730
Ion Ratio Lower Upper
276 100
277 23.8 19.7 29.5
138 24.4 9.8 14.6#

