

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
 Data File : BN015267.D
 Acq On : 01 Jul 2021 19:19
 Operator : CG/JU
 Sample : M2783-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D4-20210618

Manual Integrations
 APPROVED

mohammad
 7/2/2021 10:38:48 AM

Quant Time: Jul 02 00:59:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	20412	0.40	ng	# 0.00
7) Naphthalene-d8	10.495	136	92110	0.40	ng	0.00
13) Acenaphthene-d10	14.345	164	45689	0.40	ng	0.00
19) Phenanthrene-d10	17.090	188	86478	0.40	ng	0.00
29) Chrysene-d12	21.291	240	66979	0.40	ng	#-0.01
35) Perylene-d12	23.542	264	50273	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	8039	0.15	ng	0.00
5) Phenol-d6	6.904	99	5380	0.08	ng	0.00
8) Nitrobenzene-d5	8.873	82	19933	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	50888	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	4384	0.44	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	75316	0.41	ng	0.00
27) Fluoranthene-d10	19.127	212	96199	0.38	ng	0.00
31) Terphenyl-d14	19.733	244	83523	0.53	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.354	88	75161m	7.85	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.163	93	1598	0.02	ng	# 86
24) Pentachlorophenol	16.750	266	110	0.25	ng	96
34) Bis(2-ethylhexyl)phtha...	21.217	149	11094	0.28	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
Data File : BN015267.D
Acq On : 01 Jul 2021 19:19
Operator : CG/JU
Sample : M2783-18
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D4-20210618

Manual Integrations
APPROVED

mohammad
7/2/2021 10:38:48 AM

Quant Time: Jul 02 00:59:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

