

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
 Data File : BN015285.D
 Acq On : 02 Jul 2021 07:20
 Operator : CG/JU
 Sample : PB137445BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137445BL

Quant Time: Jul 02 08:29:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.735	152	24848	0.400	ng	# 0.00
7) Naphthalene-d8	10.495	136	110901	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	54206	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	97752	0.400	ng	0.00
29) Chrysene-d12	21.291	240	63185	0.400	ng	#-0.01
35) Perylene-d12	23.539	264	51033	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.374	112	23210	0.367	ng	0.01
5) Phenol-d6	6.914	99	25545	0.320	ng	0.01
8) Nitrobenzene-d5	8.861	82	28091	0.452	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	57745	0.314	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	3769	0.381	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	80247	0.364	ng	0.00
27) Fluoranthene-d10	19.127	212	85265	0.295	ng	0.00
31) Terphenyl-d14	19.733	244	98928	0.669	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
Data File : BN015285.D
Acq On : 02 Jul 2021 07:20
Operator : CG/JU
Sample : PB137445BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB137445BL

Quant Time: Jul 02 08:29:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

